

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062722\
 Data File : BF129023.D
 Acq On : 27 Jun 2022 11:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 29 16:23:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 16:19:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	70	-0.02
2	1,4-Dioxane	40.000	54.039	-35.1#	106	-0.03
3	Pyridine	40.000	37.969	5.1	74	-0.05
4	n-Nitrosodimethylamine	40.000	37.124	7.2	72	-0.04
5 S	2-Fluorophenol	80.000	73.006	8.7	74	-0.02
6	Aniline	40.000	38.128	4.7	72	-0.02
7 S	Phenol-d6	80.000	76.946	3.8	73	-0.02
8	2-Chlorophenol	40.000	38.391	4.0	72	-0.02
9	Benzaldehyde	40.000	34.218	14.5	74	-0.02
10 C	Phenol	40.000	38.480	3.8	73	-0.02
11	bis(2-Chloroethyl)ether	40.000	38.088	4.8	71	-0.02
12	1,3-Dichlorobenzene	40.000	37.842	5.4	71	-0.02
13 C	1,4-Dichlorobenzene	40.000	38.186	4.5	72	-0.02
14	1,2-Dichlorobenzene	40.000	39.216	2.0	74	-0.02
15	Benzyl Alcohol	40.000	38.164	4.6	72	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	36.348	9.1	68	-0.02
17	2-Methylphenol	40.000	37.842	5.4	71	-0.02
18	Hexachloroethane	40.000	39.603	1.0	73	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	38.173	4.6	72	-0.01
20	3+4-Methylphenols	40.000	38.642	3.4	72	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	70	-0.02
22	Acetophenone	40.000	36.231	9.4	73	-0.02
23 S	Nitrobenzene-d5	80.000	75.954	5.1	73	-0.02
24	Nitrobenzene	40.000	37.818	5.5	72	-0.02
25	Isophorone	40.000	37.100	7.2	71	-0.02
26 C	2-Nitrophenol	40.000	37.822	5.4	71	-0.01
27	2,4-Dimethylphenol	40.000	37.256	6.9	71	-0.02
28	bis(2-Chloroethoxy)methane	40.000	37.252	6.9	71	-0.02
29 C	2,4-Dichlorophenol	40.000	37.296	6.8	70	-0.02
30	1,2,4-Trichlorobenzene	40.000	37.715	5.7	72	-0.02
31	Naphthalene	40.000	37.526	6.2	71	-0.02
32	Benzoic acid	40.000	29.847	25.4#	56	-0.02
33	4-Chloroaniline	40.000	34.052	14.9	65	-0.02
34 C	Hexachlorobutadiene	40.000	39.503	1.2	73	-0.02
35	Caprolactam	40.000	40.965	-2.4	102	-0.02
36 C	4-Chloro-3-methylphenol	40.000	41.617	-4.0	102	-0.02
37	2-Methylnaphthalene	40.000	42.304	-5.8	101	-0.02
38	1-Methylnaphthalene	40.000	42.730	-6.8	100	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	36.125	9.7	101	-0.02
41 P	Hexachlorocyclopentadiene	40.000	39.968	0.1	107	-0.02
42 S	2,4,6-Tribromophenol	80.000	76.145	4.8	99	-0.02
43 C	2,4,6-Trichlorophenol	40.000	36.876	7.8	98	-0.02
44	2,4,5-Trichlorophenol	40.000	37.746	5.6	100	-0.02
45 S	2-Fluorobiphenyl	80.000	74.302	7.1	100	-0.02
46	1,1'-Biphenyl	40.000	37.544	6.1	101	-0.02
47	2-Chloronaphthalene	40.000	37.405	6.5	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062722\
 Data File : BF129023.D
 Acq On : 27 Jun 2022 11:09
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 29 16:23:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 16:19:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.321	4.2	101	-0.02
49	Acenaphthylene	40.000	37.486	6.3	101	-0.02
50	Dimethylphthalate	40.000	37.838	5.4	102	-0.01
51	2,6-Dinitrotoluene	40.000	37.956	5.1	101	-0.02
52 C	Acenaphthene	40.000	36.867	7.8	100	-0.02
53	3-Nitroaniline	40.000	38.448	3.9	103	-0.02
54 P	2,4-Dinitrophenol	40.000	32.801	18.0	88	-0.02
55	Dibenzofuran	40.000	37.421	6.4	102	-0.02
56 P	4-Nitrophenol	40.000	33.770	15.6	87	-0.02
57	2,4-Dinitrotoluene	40.000	37.646	5.9	100	-0.02
58	Fluorene	40.000	37.650	5.9	102	-0.02
59	2,3,4,6-Tetrachlorophenol	40.000	40.213	-0.5	98	-0.02
60	Diethylphthalate	40.000	37.732	5.7	102	-0.02
61	4-Chlorophenyl-phenylether	40.000	37.569	6.1	102	-0.02
62	4-Nitroaniline	40.000	37.797	5.5	100	-0.02
63	Azobenzene	40.000	37.669	5.8	101	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	-0.02
65	4,6-Dinitro-2-methylphenol	40.000	39.495	1.3	97	-0.01
66 c	n-Nitrosodiphenylamine	40.000	37.280	6.8	100	-0.02
67	4-Bromophenyl-phenylether	40.000	37.359	6.6	100	-0.02
68	Hexachlorobenzene	40.000	38.666	3.3	102	-0.02
69	Atrazine	40.000	39.419	1.5	102	-0.02
70 C	Pentachlorophenol	40.000	34.681	13.3	94	-0.02
71	Phenanthrene	40.000	37.578	6.1	101	-0.02
72	Anthracene	40.000	37.288	6.8	101	-0.02
73	Carbazole	40.000	37.582	6.0	102	-0.02
74	Di-n-butylphthalate	40.000	37.610	6.0	100	-0.02
75 C	Fluoranthene	40.000	39.377	1.6	101	-0.02
76 I	Chrysene-d12	20.000	20.000	0.0	99	-0.02
77	Benzydine	40.000	-20.000	150.0#	124	-0.02
78	Pyrene	40.000	37.772	5.6	100	-0.02
79 S	Terphenyl-d14	80.000	75.243	5.9	100	-0.02
80	Butylbenzylphthalate	40.000	38.145	4.6	99	-0.02
81	Benzo(a)anthracene	40.000	37.463	6.3	99	-0.02
82	3,3'-Dichlorobenzidine	40.000	35.827	10.4	99	-0.02
83	Chrysene	40.000	37.549	6.1	99	-0.02
84	Bis(2-ethylhexyl)phthalate	40.000	37.484	6.3	98	-0.02
85 c	Di-n-octyl phthalate	40.000	37.865	5.3	100	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	101	-0.02
87	Indeno(1,2,3-cd)pyrene	40.000	37.272	6.8	107	-0.04
88	Benzo(b)fluoranthene	40.000	40.649	-1.6	110	-0.02
89	Benzo(k)fluoranthene	40.000	35.354	11.6	91	-0.02
90 C	Benzo(a)pyrene	40.000	38.278	4.3	103	-0.03
91	Dibenzo(a,h)anthracene	40.000	37.981	5.0	107	-0.04
92	Benzo(g,h,i)perylene	40.000	37.527	6.2	108	-0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF062722\
Data File : BF129023.D
Acq On : 27 Jun 2022 11:09
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 29 16:23:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 16:19:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0