

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062819\
 Data File : BF115295.D
 Acq On : 28 Jun 2019 21:29
 Operator : HP/JU
 Sample : K3528-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-06-COMP

Manual Integrations
 APPROVED

mohammad
 7/1/2019 3:05:55 PM

Quant Time: Jun 29 06:34:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:26:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	273782	20.00	ng	0.00
21) Naphthalene-d8	7.88	136	1055137	20.00	ng	0.00
39) Acenaphthene-d10	9.62	164	536691	20.00	ng	0.00
64) Phenanthrene-d10	11.10	188	1030829	20.00	ng	0.00
76) Chrysene-d12	13.71	240	759323	20.00	ng	0.00
87) Perylene-d12	15.07	264	760770	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	1522503	85.82	ng	0.00
7) Phenol-d6	6.24	99	1939342	90.06	ng	0.00
23) Nitrobenzene-d5	7.16	82	1181660	68.89	ng	0.00
42) 2,4,6-Tribromophenol	10.41	330	562086	99.32	ng	0.00
45) 2-Fluorobiphenyl	8.95	172	2185830	69.18	ng	0.00
79) Terphenyl-d14	12.68	244	2291729	64.17	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.35	163	352483	8.927	ng	98
75) Fluoranthene	12.30	202	206915	3.737	ng	95
78) Pyrene	12.52	202	188737	3.234	ng	98
83) Chrysene	13.74	228	127276	2.550	ng	99
84) Bis(2-ethylhexyl)phthalate	13.72	149	138988	4.119	ng	99
88) Benzo(b)fluoranthene	14.70	252	172556m	3.635	ng	
90) Benzo(a)pyrene	15.01	252	106787	2.496	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062819\
Data File : BF115295.D
Acq On : 28 Jun 2019 21:29
Operator : HP/JU
Sample : K3528-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-06-COMP

Manual Integrations
APPROVED

mohammad
7/1/2019 3:05:55 PM

Quant Time: Jun 29 06:34:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:26:43 2019
Response via : Initial Calibration

