

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062819\
 Data File : BF115306.D
 Acq On : 29 Jun 2019 2:21
 Operator : HP/JU
 Sample : K3531-01
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-GF-02-043-A

Quant Time: Jun 29 07:06:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:26:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	241415	20.00	ng	0.00
21) Naphthalene-d8	7.88	136	926506	20.00	ng	0.00
39) Acenaphthene-d10	9.62	164	446434	20.00	ng	0.00
64) Phenanthrene-d10	11.09	188	829421	20.00	ng	0.00
76) Chrysene-d12	13.72	240	607986	20.00	ng	0.00
87) Perylene-d12	15.08	264	454128	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	1392878	89.04	ng	0.01
7) Phenol-d6	6.24	99	1767277	93.07	ng	0.00
23) Nitrobenzene-d5	7.16	82	1023478	67.95	ng	0.00
42) 2,4,6-Tribromophenol	10.41	330	470437	99.93	ng	0.00
45) 2-Fluorobiphenyl	8.96	172	1838671	69.96	ng	0.00
79) Terphenyl-d14	12.68	244	1884505	65.90	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.35	163	304950	9.284	ng	99
75) Fluoranthene	12.30	202	108288	2.430	ng	95
78) Pyrene	12.52	202	105306	2.254	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062819\
Data File : BF115306.D
Acq On : 29 Jun 2019 2:21
Operator : HP/JU
Sample : K3531-01
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-GF-02-043-A

Quant Time: Jun 29 07:06:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:26:43 2019
Response via : Initial Calibration

