

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062920\
 Data File : BF120731.D
 Acq On : 29 Jun 2020 16:56
 Operator : JU/CG
 Sample : L3129-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 1405-1406

Manual Integrations
 APPROVED

mohammad
 6/30/2020 4:14:34 PM

Quant Time: Jun 30 13:40:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 29 15:30:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	108243	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	347578	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	78527	20.00	ng	0.00
64) Phenanthrene-d10	11.39	188	109043	20.00	ng	0.08
76) Chrysene-d12	13.99	240	118241	20.00	ng	0.04
86) Perylene-d12	15.44	264	184284	20.00	ng	0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	391852	62.07	ng	0.00
7) Phenol-d6	6.44	99	282067	35.87	ng	0.01
23) Nitrobenzene-d5	7.36	82	405542	65.22	ng	0.00
42) 2,4,6-Tribromophenol	10.66	330	114670	125.60	ng	0.04
45) 2-Fluorobiphenyl	9.16	172	507981	98.97	ng	0.00
79) Terphenyl-d14	12.96	244	390274	72.91	ng	0.06
Target Compounds						
10) Phenol	6.45	94	102217	12.183	ng	85
20) 3+4-Methylphenols	7.20	107	543892m	71.528	ng	
46) 1,1'-Biphenyl	9.26	154	20335	3.468	ng	87
50) Dimethylphthalate	9.56	163	60529	10.703	ng	# 78
75) Fluoranthene	12.61	202	18187	2.935	ng	# 9
78) Pyrene	12.83	202	19647m	2.346	ng	
84) Bis(2-ethylhexyl)phthalate	13.97	149	37921	8.354	ng	# 86
87) Indeno(1,2,3-cd)pyrene	16.90	276	26979	2.407	ng	# 47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062920\
Data File : BF120731.D
Acq On : 29 Jun 2020 16:56
Operator : JU/CG
Sample : L3129-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1405-1406

Manual Integrations
APPROVED

mohammad
6/30/2020 4:14:34 PM

Quant Time: Jun 30 13:40:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 29 15:30:14 2020
Response via : Initial Calibration

