

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070722\
 Data File : BF129162.D
 Acq On : 07 Jul 2022 16:51
 Operator : CG\JU
 Sample : N3584-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CHRT-2716S

Quant Time: Jul 07 17:51:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 29 17:05:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.951	152	203207	20.000	ng	-0.01
21) Naphthalene-d8	8.234	136	663037	20.000	ng	-0.01
39) Acenaphthene-d10	10.004	164	255690	20.000	ng	0.00
64) Phenanthrene-d10	11.533	188	230435	20.000	ng	# 0.04
76) Chrysene-d12	14.180	240	470841	20.000	ng	# 0.03
86) Perylene-d12	15.680	264	593883	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.569	112	1565120	130.214	ng	-0.02
7) Phenol-d6	6.575	99	1867032	125.070	ng	-0.01
23) Nitrobenzene-d5	7.510	82	1035680	93.169	ng	-0.02
42) 2,4,6-Tribromophenol	10.810	330	392389	141.135	ng	0.01
45) 2-Fluorobiphenyl	9.310	172	1591516	99.577	ng	-0.01
79) Terphenyl-d14	13.133	244	1351179	59.593	ng	0.05
Target Compounds						Qvalue
9) Benzaldehyde	6.504	77	921	Below Cal	#	84
19) n-Nitroso-di-n-propyla...	7.510	70	134762	15.606	ng	# 73
25) Isophorone	7.510	82	1024944	54.557	ng	# 65
35) Caprolactam	8.704	113	6409	2.196	ng	# 1
49) Acenaphthylene	9.863	152	48463	2.247	ng	# 1
51) 2,6-Dinitrotoluene	9.798	165	42950	13.140	ng	# 29
53) 3-Nitroaniline	9.951	138	66558	17.075	ng	# 13
54) 2,4-Dinitrophenol	10.057	184	792	7.276	ng	# 1
56) 4-Nitrophenol	10.128	139	14893	4.676	ng	# 1
57) 2,4-Dinitrotoluene	10.210	165	14501	3.594	ng	# 35
59) 2,3,4,6-Tetrachlorophenol	10.310	232	9749	2.260	ng	# 1
61) 4-Chlorophenyl-phenyle...	10.575	204	56601	7.178	ng	# 14
62) 4-Nitroaniline	10.345	138	43195	11.175	ng	# 85
63) Azobenzene	10.698	77	37777	2.406	ng	# 24
65) 4,6-Dinitro-2-methylph...	10.628	198	8906	8.736	ng	# 1
66) n-Nitrosodiphenylamine	10.681	169	105926	15.136	ng	# 52
69) Atrazine	11.251	200	6316	3.207	ng	# 1
70) Pentachlorophenol	11.292	266	4975	3.060	ng	# 1
71) Phenanthrene	11.557	178	84083	7.637	ng	# 1
74) Di-n-butylphthalate	12.104	149	114483	9.982	ng	# 34
75) Fluoranthene	12.775	202	94113	8.173	ng	# 34
77) Benzidine	12.833	184	40895	4.878	ng	# 1
81) Benzo(a)anthracene	14.169	228	67810	2.359	ng	# 62
83) Chrysene	14.204	228	73609	2.758	ng	# 70
88) Benzo(b)fluoranthene	15.227	252	180736	5.060	ng	# 66
89) Benzo(k)fluoranthene	15.227	252	180736	5.217	ng	# 66
90) Benzo(a)pyrene	15.610	252	78860	2.708	ng	# 76
92) Benzo(g,h,i)perylene	17.686	276	65409	2.196	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF070722\
Data File : BF129162.D
Acq On : 07 Jul 2022 16:51
Operator : CG\JU
Sample : N3584-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CHRT-2716S

Quant Time: Jul 07 17:51:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF062922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 29 17:05:31 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.951 min Scan# 795
Delta R.T. -0.012 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S

Tgt Ion:152 Resp: 203207
Ion Ratio Lower Upper
152 100
150 155.9 123.7 185.5
115 59.0 49.9 74.9



#5
2-Fluorophenol
Concen: 130.214 ng
RT: 5.569 min Scan# 558
Delta R.T. -0.018 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:112 Resp: 1565120
Ion Ratio Lower Upper
112 100
64 58.6 53.3 79.9
63 28.8 28.0 42.0





#7
Phenol-d6
Concen: 125.070 ng
RT: 6.575 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51





Instrument :

BNA_F

ClientSampleId :




CHRT-2716S

Tgt Ion: 99 Resp: 1867032
Ion Ratio Lower Upper
99 100
42 15.6 17.3 25.9#
71 30.5 28.9 43.3



#19
n-Nitroso-di-n-propylamine
Concen: 15.606 ng
RT: 7.510 min Scan# 888
Delta R.T. 0.135 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion: 70 Resp: 134762
Ion Ratio Lower Upper
70 100
42 47.1 52.3 78.5#
101 0.0 7.5 11.3#
130 2.0 16.1 24.1#








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.234 min Scan# 1011
Delta R.T. -0.011 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S

Tgt Ion:136 Resp: 663037
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 8.3 7.5 11.3
68 6.7 4.7 7.1



#23
Nitrobenzene-d5
Concen: 93.169 ng
RT: 7.510 min Scan# 888
Delta R.T. -0.018 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion: 82 Resp: 1035680
Ion Ratio Lower Upper
82 100
128 45.4 34.0 51.0
54 50.1 46.9 70.3









#25
Isophorone
Concen: 54.557 ng
RT: 7.510 min Scan# 81
Delta R.T. -0.282 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion: 82 Resp: 1024944
Ion Ratio Lower Upper
82 100
95 0.1 5.7 8.5#
138 0.0 14.2 21.2#



#35
Caprolactam
Concen: 2.196 ng
RT: 8.704 min Scan# 1091
Delta R.T. -0.006 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion: 113 Resp: 6409
Ion Ratio Lower Upper
113 100
55 638.0 188.5 228.5#
56 131.5 130.6 170.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:164 Resp: 255690
Ion Ratio Lower Upper
164 100
162 95.2 79.5 119.3
160 44.0 34.2 51.4



#42
2,4,6-Tribromophenol
Concen: 141.135 ng
RT: 10.810 min Scan# 1449
Delta R.T. 0.012 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:330 Resp: 392389
Ion Ratio Lower Upper
330 100
332 97.9 76.8 115.2
141 16.2 28.0 42.0#





#45

2-Fluorobiphenyl

Concen: 99.577 ng

RT: 9.310 min Scan# 1196

Delta R.T. -0.012 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:172 Resp: 1591516

Ion Ratio Lower Upper

172 100

171 35.4 28.8 43.2

170 24.1 19.8 29.8



#49

Acenaphthylene

Concen: 2.247 ng

RT: 9.863 min Scan# 1288

Delta R.T. -0.006 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:152 Resp: 48463


Ion Ratio Lower Upper

152 100

151 112.7 16.2 24.4#

153 45.7 10.7 16.1#





#51

2,6-Dinitrotoluene

Concen: 13.140 ng

RT: 9.798 min Scan# 1

Delta R.T. 0.012 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:165 Resp: 42950

Ion Ratio Lower Upper

165 100

63 5.0 44.9 67.3#

89 3.3 43.6 65.4#



#53

3-Nitroaniline

Concen: 17.075 ng

RT: 9.951 min Scan# 1303

Delta R.T. -0.006 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:138 Resp: 66558


Ion Ratio Lower Upper

138 100

108 52.0 10.9 16.3#

92 26.0 98.2 147.2#









#54
2,4-Dinitrophenol
Concen: 7.276 ng
RT: 10.057 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:184 Resp: 792
Ion Ratio Lower Upper
184 100
63 683.0 64.5 96.7#
154 808.3 50.6 75.8#



#56
4-Nitrophenol
Concen: 4.676 ng
RT: 10.128 min Scan# 1333
Delta R.T. 0.024 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:139 Resp: 14893
Ion Ratio Lower Upper
139 100
109 813.7 50.7 90.7#
65 133.9 74.5 114.5#









#57
2,4-Dinitrotoluene
Concen: 3.594 ng
RT: 10.210 min Scan# 1347
Delta R.T. 0.024 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:165 Resp: 14501
Ion Ratio Lower Upper
165 100
63 15.1 39.8 59.6#
89 10.6 60.2 90.2#
182 10.1 2.2 3.2#



#59
2,3,4,6-Tetrachlorophenol
Concen: 2.260 ng
RT: 10.310 min Scan# 1364
Delta R.T. -0.018 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:232 Resp: 9749
Ion Ratio Lower Upper
232 100
131 276.8 36.1 54.1#
130 144.7 1.8 2.8#
166 0.0 24.2 36.2#









#61
4-Chlorophenyl-phenylether
Concen: 7.178 ng
RT: 10.575 min Scan# 1409
Delta R.T. 0.029 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:204 Resp: 56601
Ion Ratio Lower Upper
204 100
206 74.0 25.8 38.6#
141 127.8 47.0 70.4#



#62
4-Nitroaniline
Concen: 11.175 ng
RT: 10.345 min Scan# 1370
Delta R.T. -0.229 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:138 Resp: 43195
Ion Ratio Lower Upper
138 100
92 48.3 33.5 73.5
108 98.2 61.0 101.0









#63
Azobenzene
Concen: 2.406 ng
RT: 10.698 min Scan# 1431
Delta R.T. -0.006 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Instrument :
BNA_F
ClientSampleId :
CHRT-2716S

Tgt Ion: 77 Resp: 37777




Ion	Ratio	Lower	Upper
77	100		
182	10.7	5.6	45.6
105	118.0	1.6	41.6#
51	17.3	15.1	55.1




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.533 min Scan# 1572
Delta R.T. 0.035 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion: 188 Resp: 230435

Ion	Ratio	Lower	Upper
188	100		
94	36.9	7.0	10.6#
80	31.7	7.6	11.4#









#65
4,6-Dinitro-2-methylphenol
Concen: 8.736 ng
RT: 10.628 min Scan# 1418
Delta R.T. 0.029 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:198 Resp: 8906
Ion Ratio Lower Upper
198 100
51 174.0 37.6 77.6#
105 1018.3 27.0 67.0#



#66
n-Nitrosodiphenylamine
Concen: 15.136 ng
RT: 10.681 min Scan# 1427
Delta R.T. 0.018 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:169 Resp: 105926
Ion Ratio Lower Upper
169 100
168 13.8 50.1 75.1#
167 23.6 27.9 41.9#





#69

Atrazine

Concen: 3.207 ng

RT: 11.251 min Scan# 1

Delta R.T. 0.059 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:200 Resp: 6316

Ion Ratio Lower Upper

200 100

173 446.5 6.3 46.3#

215 462.8 28.3 68.3#



#70

Pentachlorophenol

Concen: 3.060 ng

RT: 11.292 min Scan# 1531

Delta R.T. -0.000 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:266 Resp: 4975


Ion Ratio Lower Upper

266 100

268 12.1 51.0 76.4#

264 189.0 49.4 74.2#









#71
Phenanthrene
Concen: 7.637 ng
RT: 11.557 min Scan# 1156
Delta R.T. 0.035 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:178 Resp: 84083
Ion Ratio Lower Upper
178 100
176 33.1 16.0 24.0#
179 147.9 12.6 19.0#



#74
Di-n-butylphthalate
Concen: 9.982 ng
RT: 12.104 min Scan# 1669
Delta R.T. 0.053 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:149 Resp: 114483
Ion Ratio Lower Upper
149 100
150 43.0 7.5 11.3#
104 13.2 4.4 6.6#





#75

Fluoranthene

Concen: 8.173 ng

RT: 12.775 min Scan# 1

Delta R.T. 0.059 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:202 Resp: 94113

Ion Ratio Lower Upper

202 100

101 14.6 0.0 31.4

203 63.2 0.0 38.0#



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.180 min Scan# 2022

Delta R.T. 0.029 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:240 Resp: 470841


Ion Ratio Lower Upper

240 100

120 14.2 8.9 13.3#

236 27.5 20.8 31.2





#77

Benzidine

Concen: 4.878 ng

RT: 12.833 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:184 Resp: 40895

Ion Ratio Lower Upper

184 100

185 106.6 11.9 17.9#

183 402.6 9.3 13.9#



#79

Terphenyl-d14

Concen: 59.593 ng

RT: 13.133 min Scan# 1844

Delta R.T. 0.047 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:244 Resp: 1351179


Ion Ratio Lower Upper

244 100

212 6.9 6.1 9.1

122 10.8 8.8 13.2









#81
Benzo(a)anthracene
Concen: 2.359 ng
RT: 14.169 min Scan# 2015
Delta R.T. 0.029 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:228 Resp: 67810
Ion Ratio Lower Upper
228 100
226 37.5 22.1 33.1#
229 49.5 15.8 23.8#



#83
Chrysene
Concen: 2.758 ng
RT: 14.204 min Scan# 2026
Delta R.T. 0.029 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:228 Resp: 73609
Ion Ratio Lower Upper
228 100
226 38.2 24.3 36.5#
229 44.6 15.9 23.9#





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.680 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51

Instrument :

BNA_F

ClientSampleId :

CHRT-2716S





Tgt Ion:264 Resp: 593883

Ion Ratio Lower Upper

264 100

260 23.1 18.4 27.6

265 22.0 17.2 25.8



#88

Benzo(b)fluoranthene

Concen: 5.060 ng

RT: 15.227 min Scan# 2200

Delta R.T. 0.006 min

Lab File: BF129162.D

Acq: 07 Jul 2022 16:51




Tgt Ion:252 Resp: 180736


Ion Ratio Lower Upper

252 100

253 31.9 17.7 26.5#

125 33.7 8.0 12.0#









#89
Benzo(k)fluoranthene
Concen: 5.217 ng
RT: 15.227 min Scan# 21
Delta R.T. -0.024 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51




Instrument :
BNA_F
ClientSampleId :
CHRT-2716S


Tgt Ion:252 Resp: 180736
Ion Ratio Lower Upper
252 100
253 31.9 17.8 26.6#
125 33.7 8.3 12.5#



#90
Benzo(a)pyrene
Concen: 2.708 ng
RT: 15.610 min Scan# 2265
Delta R.T. -0.000 min
Lab File: BF129162.D
Acq: 07 Jul 2022 16:51

Tgt Ion:252 Resp: 78860
Ion Ratio Lower Upper
252 100
253 26.6 17.0 25.6#
125 29.4 8.8 13.2#





#92
 Benzo(g,h,i)perylene
 Concen: 2.196 ng
 RT: 17.686 min Scan# 20
 Delta R.T. -0.024 min
 Lab File: BF129162.D
 Acq: 07 Jul 2022 16:51

Instrument :
 BNA_F
 ClientSampleId :
 CHRT-2716S

Tgt Ion:276 Resp: 65409
 Ion Ratio Lower Upper
 276 100
 277 28.0 19.1 28.7
 138 26.8 18.0 27.0

