

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071520\  
 Data File : BF120870.D  
 Acq On : 15 Jul 2020 13:43  
 Operator : JU/CG  
 Sample : L3259-07  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 C-1-ROLLOFF

Quant Time: Jul 15 14:57:25 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070220.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 13 14:16:14 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.83  | 152  | 180052   | 20.00 | ng    | 0.00      |
| 21) Naphthalene-d8          | 8.12  | 136  | 685690   | 20.00 | ng    | 0.00      |
| 39) Acenaphthene-d10        | 9.87  | 164  | 362462   | 20.00 | ng    | 0.00      |
| 64) Phenanthrene-d10        | 11.35 | 188  | 697140   | 20.00 | ng    | 0.00      |
| 76) Chrysene-d12            | 13.99 | 240  | 579469   | 20.00 | ng    | -0.01     |
| 86) Perylene-d12            | 15.44 | 264  | 520427   | 20.00 | ng    | -0.01     |
| System Monitoring Compounds |       |      |          |       |       |           |
| 5) 2-Fluorophenol           | 5.45  | 112  | 637797   | 54.68 | ng    | 0.00      |
| 7) Phenol-d6                | 6.46  | 99   | 846560   | 56.83 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 7.39  | 82   | 549916   | 46.47 | ng    | -0.01     |
| 42) 2,4,6-Tribromophenol    | 10.66 | 330  | 192328   | 52.82 | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 9.19  | 172  | 1061159  | 44.29 | ng    | 0.00      |
| 79) Terphenyl-d14           | 12.94 | 244  | 1110932  | 37.54 | ng    | 0.00      |
| Target Compounds            |       |      |          |       |       |           |
| 50) Dimethylphthalate       | 9.58  | 163  | 69115    | 2.614 | ng    | Qvalue 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF071520\  
Data File : BF120870.D  
Acq On : 15 Jul 2020 13:43  
Operator : JU/CG  
Sample : L3259-07  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
C-1-ROLLOFF

Quant Time: Jul 15 14:57:25 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF070220.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jul 13 14:16:14 2020  
Response via : Initial Calibration

