


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072821\
 Data File : BF124826.D
 Acq On : 28 Jul 2021 14:15
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 22 15:28:57 2021






AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.2	4746	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.7	3811	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.3	5414	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	10992	PASS
199	198	5	9	6.3	687	PASS
275	198	10	60	25.7	2825	PASS
365	198	1	100	1.6	178	PASS
441	198	0.01	100	13.5	1488	PASS
442	442	50	100	100.0	9161	PASS
443	442	15	24	21.7	1988	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072821\
Data File : BF124826.D
Acq On : 28 Jul 2021 14:15
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jul 28 15:26:23 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 22 15:28:57 2021
Response via : Initial Calibration



TIC: BF124826.D\data.ms

(77) Benzidine

12.733min (-0.000) 0.00 ng




response 45917

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.50	13.57
183.00	10.10	11.04
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072821\
 Data File : BF124826.D
 Acq On : 28 Jul 2021 14:15
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Jul 28 15:26:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 22 15:28:57 2021
 Response via : Initial Calibration



TIC: BF124826.D\data.ms

(70) Pentachlorophenol (C)

11.198min (-0.000) 0.00 ng

response 9898

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.10	62.06
264.00	69.90	61.26
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/28/2021	BNA_F	<u>BF124826.D</u>
Compound Name	Response	Retention Time
DDT	25926	13.55
DDD	254	13.25
DDE	0	12.745
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
254	26180	0.97