

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/1/2022	BNA_F	<u>BF129464.D</u>
Compound Name	Response	Retention Time
DDT	598843	13.557
DDD	8375	13.257
DDE	957	12.921
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9332	608175	1.53