

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/1/2022	BNA_F	<u>BF129475.D</u>
Compound Name	Response	Retention Time
DDT	707861	13.563
DDD	10128	13.263
DDE	1302	12.921
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
11430	719291	1.59