

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080122\
 Data File : BF129465.D
 Acq On : 01 Aug 2022 10:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 01:14:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	71	0.00
2	1,4-Dioxane	40.000	38.799	3.0	70	0.00
3	Pyridine	40.000	38.438	3.9	67	0.00
4	n-Nitrosodimethylamine	40.000	39.053	2.4	70	0.00
5 S	2-Fluorophenol	80.000	78.017	2.5	71	0.00
6	Aniline	40.000	36.882	7.8	67	0.00
7 S	Phenol-d6	80.000	78.695	1.6	72	0.00
8	2-Chlorophenol	40.000	39.714	0.7	72	0.00
9	Benzaldehyde	40.000	36.161	9.6	69	0.00
10 C	Phenol	40.000	40.190	-0.5	74	0.00
11	bis(2-Chloroethyl)ether	40.000	39.112	2.2	71	0.00
12	1,3-Dichlorobenzene	40.000	39.621	0.9	72	0.00
13 C	1,4-Dichlorobenzene	40.000	39.712	0.7	72	0.00
14	1,2-Dichlorobenzene	40.000	39.622	0.9	71	0.00
15	Benzyl Alcohol	40.000	40.794	-2.0	73	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.750	8.1	66	0.00
17	2-Methylphenol	40.000	38.446	3.9	70	0.00
18	Hexachloroethane	40.000	39.635	0.9	72	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.773	3.1	72	0.00
20	3+4-Methylphenols	40.000	37.654	5.9	72	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	71	0.00
22	Acetophenone	40.000	39.390	1.5	73	0.00
23 S	Nitrobenzene-d5	80.000	80.959	-1.2	73	0.00
24	Nitrobenzene	40.000	39.624	0.9	71	0.00
25	Isophorone	40.000	38.717	3.2	71	0.00
26 C	2-Nitrophenol	40.000	39.438	1.4	70	0.00
27	2,4-Dimethylphenol	40.000	39.735	0.7	72	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.149	2.1	72	0.00
29 C	2,4-Dichlorophenol	40.000	39.627	0.9	71	0.00
30	1,2,4-Trichlorobenzene	40.000	40.190	-0.5	73	0.00
31	Naphthalene	40.000	39.897	0.3	73	0.00
32	Benzoic acid	40.000	26.454	33.9#	46	0.00
33	4-Chloroaniline	40.000	38.157	4.6	70	0.00
34 C	Hexachlorobutadiene	40.000	41.650	-4.1	76	0.00
35	Caprolactam	40.000	39.606	1.0	71	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.264	-0.7	73	0.00
37	2-Methylnaphthalene	40.000	40.083	-0.2	73	0.00
38	1-Methylnaphthalene	40.000	40.267	-0.7	75	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	73	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.555	1.1	75	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.281	9.3	62	0.00
42 S	2,4,6-Tribromophenol	80.000	83.393	-4.2	78	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.284	1.8	72	0.00
44	2,4,5-Trichlorophenol	40.000	38.571	3.6	72	0.00
45 S	2-Fluorobiphenyl	80.000	74.426	7.0	78	0.00
46	1,1'-Biphenyl	40.000	39.329	1.7	74	0.00
47	2-Chloronaphthalene	40.000	39.114	2.2	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080122\
 Data File : BF129465.D
 Acq On : 01 Aug 2022 10:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 01:14:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 02 01:11:25 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.777	0.6	72	0.00
49	Acenaphthylene	40.000	39.622	0.9	75	0.00
50	Dimethylphthalate	40.000	40.092	-0.2	76	0.00
51	2,6-Dinitrotoluene	40.000	39.904	0.2	74	0.00
52 C	Acenaphthene	40.000	39.321	1.7	73	0.00
53	3-Nitroaniline	40.000	38.195	4.5	70	0.00
54 P	2,4-Dinitrophenol	40.000	35.354	11.6	62	0.00
55	Dibenzofuran	40.000	40.135	-0.3	75	0.00
56 P	4-Nitrophenol	40.000	35.442	11.4	64	0.00
57	2,4-Dinitrotoluene	40.000	40.652	-1.6	73	0.00
58	Fluorene	40.000	40.276	-0.7	77	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.881	2.8	71	0.00
60	Diethylphthalate	40.000	40.690	-1.7	76	0.00
61	4-Chlorophenyl-phenylether	40.000	40.540	-1.3	77	0.00
62	4-Nitroaniline	40.000	39.277	1.8	73	0.00
63	Azobenzene	40.000	39.153	2.1	73	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.960	2.6	69	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.738	0.7	76	0.00
67	4-Bromophenyl-phenylether	40.000	40.307	-0.8	77	0.00
68	Hexachlorobenzene	40.000	40.128	-0.3	76	0.00
69	Atrazine	40.000	39.524	1.2	75	0.00
70 C	Pentachlorophenol	40.000	33.661	15.8	60	0.00
71	Phenanthrene	40.000	39.116	2.2	75	0.00
72	Anthracene	40.000	39.896	0.3	78	0.00
73	Carbazole	40.000	38.944	2.6	75	0.00
74	Di-n-butylphthalate	40.000	40.341	-0.9	78	0.00
75 C	Fluoranthene	40.000	40.098	-0.2	77	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	76	0.00
77	Benzidine	40.000	33.236	16.9	60	0.00
78	Pyrene	40.000	40.768	-1.9	77	0.00
79 S	Terphenyl-d14	80.000	82.943	-3.7	80	0.00
80	Butylbenzylphthalate	40.000	40.991	-2.5	76	0.00
81	Benzo(a)anthracene	40.000	39.462	1.3	76	0.00
82	3,3'-Dichlorobenzidine	40.000	37.194	7.0	70	0.00
83	Chrysene	40.000	39.486	1.3	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.248	1.9	74	0.00
85 c	Di-n-octyl phthalate	40.000	38.281	4.3	74	0.00
86 I	Perylene-d12	20.000	20.000	0.0	68	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.066	-5.2	66	0.00
88	Benzo(b)fluoranthene	40.000	42.562	-6.4	75	0.00
89	Benzo(k)fluoranthene	40.000	39.470	1.3	66	0.00
90 C	Benzo(a)pyrene	40.000	39.393	1.5	67	0.00
91	Dibenzo(a,h)anthracene	40.000	41.965	-4.9	65	0.00
92	Benzo(g,h,i)perylene	40.000	43.908	-9.8	67	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080122\
Data File : BF129465.D
Acq On : 01 Aug 2022 10:40
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 01:14:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 02 01:11:25 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0