

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
 Data File : BF129627.D
 Acq On : 08 Aug 2022 11:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 08 14:36:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 05 16:27:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	36.856	7.9	72	-0.02
3	Pyridine	40.000	37.184	7.0	71	-0.02
4	n-Nitrosodimethylamine	40.000	38.405	4.0	74	-0.02
5 S	2-Fluorophenol	80.000	77.484	3.1	77	0.00
6	Aniline	40.000	36.217	9.5	71	0.00
7 S	Phenol-d6	80.000	79.031	1.2	79	0.00
8	2-Chlorophenol	40.000	40.469	-1.2	79	0.00
9	Benzaldehyde	40.000	35.766	10.6	74	0.00
10 C	Phenol	40.000	39.787	0.5	79	0.00
11	bis(2-Chloroethyl)ether	40.000	40.013	-0.0	79	-0.01
12	1,3-Dichlorobenzene	40.000	40.287	-0.7	80	0.00
13 C	1,4-Dichlorobenzene	40.000	40.155	-0.4	80	0.00
14	1,2-Dichlorobenzene	40.000	40.275	-0.7	79	0.00
15	Benzyl Alcohol	40.000	42.050	-5.1	81	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	36.236	9.4	71	0.00
17	2-Methylphenol	40.000	39.282	1.8	78	0.00
18	Hexachloroethane	40.000	41.475	-3.7	82	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	41.314	-3.3	83	0.00
20	3+4-Methylphenols	40.000	39.305	1.7	81	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
22	Acetophenone	40.000	40.898	-2.2	84	-0.01
23 S	Nitrobenzene-d5	80.000	82.464	-3.1	82	0.00
24	Nitrobenzene	40.000	39.780	0.5	79	0.00
25	Isophorone	40.000	39.998	0.0	81	0.00
26 C	2-Nitrophenol	40.000	41.614	-4.0	82	-0.01
27	2,4-Dimethylphenol	40.000	40.765	-1.9	82	-0.01
28	bis(2-Chloroethoxy)methane	40.000	40.935	-2.3	83	0.00
29 C	2,4-Dichlorophenol	40.000	40.964	-2.4	81	-0.01
30	1,2,4-Trichlorobenzene	40.000	40.950	-2.4	82	0.00
31	Naphthalene	40.000	40.162	-0.4	81	0.00
32	Benzoic acid	40.000	28.382	29.0#	54	0.00
33	4-Chloroaniline	40.000	39.618	1.0	80	0.00
34 C	Hexachlorobutadiene	40.000	44.019	-10.0	89	-0.01
35	Caprolactam	40.000	39.088	2.3	78	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.465	-3.7	83	0.00
37	2-Methylnaphthalene	40.000	40.736	-1.8	83	-0.01
38	1-Methylnaphthalene	40.000	41.236	-3.1	85	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	80	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.415	-6.0	87	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.781	3.0	72	-0.01
42 S	2,4,6-Tribromophenol	80.000	82.214	-2.8	84	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.032	-0.1	80	0.00
44	2,4,5-Trichlorophenol	40.000	39.660	0.9	80	0.00
45 S	2-Fluorobiphenyl	80.000	78.290	2.1	89	0.00
46	1,1'-Biphenyl	40.000	41.058	-2.6	84	0.00
47	2-Chloronaphthalene	40.000	40.200	-0.5	82	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
 Data File : BF129627.D
 Acq On : 08 Aug 2022 11:23
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 08 14:36:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 05 16:27:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.613	1.0	78	0.00
49	Acenaphthylene	40.000	39.403	1.5	81	0.00
50	Dimethylphthalate	40.000	41.993	-5.0	86	0.00
51	2,6-Dinitrotoluene	40.000	41.098	-2.7	82	0.00
52 C	Acenaphthene	40.000	40.586	-1.5	82	0.00
53	3-Nitroaniline	40.000	38.440	3.9	76	0.00
54 P	2,4-Dinitrophenol	40.000	37.650	5.9	71	0.00
55	Dibenzofuran	40.000	39.964	0.1	82	0.00
56 P	4-Nitrophenol	40.000	34.114	14.7	67	0.00
57	2,4-Dinitrotoluene	40.000	39.326	1.7	77	0.00
58	Fluorene	40.000	41.147	-2.9	86	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.451	3.9	76	0.00
60	Diethylphthalate	40.000	42.308	-5.8	86	-0.01
61	4-Chlorophenyl-phenylether	40.000	42.410	-6.0	88	0.00
62	4-Nitroaniline	40.000	37.224	6.9	75	0.00
63	Azobenzene	40.000	39.978	0.1	81	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.592	-1.5	75	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.161	-2.9	83	0.00
67	4-Bromophenyl-phenylether	40.000	43.310	-8.3	87	0.00
68	Hexachlorobenzene	40.000	42.958	-7.4	86	0.00
69	Atrazine	40.000	40.799	-2.0	81	0.00
70 C	Pentachlorophenol	40.000	36.845	7.9	69	0.00
71	Phenanthrene	40.000	40.004	-0.0	81	0.00
72	Anthracene	40.000	40.361	-0.9	83	0.00
73	Carbazole	40.000	38.570	3.6	78	0.00
74	Di-n-butylphthalate	40.000	43.182	-8.0	88	0.00
75 C	Fluoranthene	40.000	39.127	2.2	79	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	73	0.00
77	Benzidine	40.000	36.501	8.7	64	0.00
78	Pyrene	40.000	42.409	-6.0	78	0.00
79 S	Terphenyl-d14	80.000	90.007	-12.5	85	0.00
80	Butylbenzylphthalate	40.000	44.687	-11.7	80	0.00
81	Benzo(a)anthracene	40.000	40.669	-1.7	76	0.00
82	3,3'-Dichlorobenzidine	40.000	39.157	2.1	72	0.00
83	Chrysene	40.000	39.366	1.6	74	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.978	-12.4	83	0.00
85 c	Di-n-octyl phthalate	40.000	47.414	-18.5	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	81	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.205	-10.5	83	0.00
88	Benzo(b)fluoranthene	40.000	40.078	-0.2	85	0.00
89	Benzo(k)fluoranthene	40.000	38.886	2.8	78	0.00
90 C	Benzo(a)pyrene	40.000	40.130	-0.3	82	0.00
91	Dibenzo(a,h)anthracene	40.000	44.349	-10.9	83	0.00
92	Benzo(g,h,i)perylene	40.000	44.207	-10.5	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
Data File : BF129627.D
Acq On : 08 Aug 2022 11:23
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 08 14:36:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 16:27:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0