


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
 Data File : BF129626.D
 Acq On : 08 Aug 2022 10:17
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Aug 05 16:27:34 2022






AutoFind: Scans 1585, 1586, 1587; Background Corrected with Scan 1580

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	113499	PASS
68	69	0.00	2	1.7	1982	PASS
69	198	0.00	100	39.3	114405	PASS
70	69	0.00	2	0.7	791	PASS
127	198	10	80	52.9	153859	PASS
197	198	0.00	2	0.6	1660	PASS
198	198	100	100	100.0	291051	PASS
199	198	5	9	6.6	19095	PASS
275	198	10	60	28.9	84144	PASS
365	198	1	100	3.9	11224	PASS
441	198	0.01	100	13.8	40241	PASS
442	442	50	100	100.0	259485	PASS
443	442	15	24	18.7	48403	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
Data File : BF129626.D
Acq On : 08 Aug 2022 10:17
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 09 06:46:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 16:27:34 2022
Response via : Initial Calibration



TIC: BF129626.D\data.ms

(70) Pentachlorophenol (C)

11.180min (-0.012) 554323.33 ng




response 200227

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	64.03
264.00	61.80	65.14
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080822\
Data File : BF129626.D
Acq On : 08 Aug 2022 10:17
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 09 06:46:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 16:27:34 2022
Response via : Initial Calibration



TIC: BF129626.D\data.ms

(77) Benzidine

12.716min (-0.006) 65987.92 ng

response 1118971

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.43
183.00	11.60	11.97
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/8/2022	BNA_F	<u>BF129626.D</u>
Compound Name	Response	Retention Time
DDT	662765	13.533
DDD	10885	13.233
DDE	1520	12.892
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12405	675170	1.84

Instrument :
BNA_F
ClientSampleId :
DFTPP