

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081022\
 Data File : BF129690.D
 Acq On : 10 Aug 2022 10:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 10 15:27:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 15:26:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	176136	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	701049	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	381486	20.000	ng	0.00
64) Phenanthrene-d10	11.369	188	652882	20.000	ng	0.00
76) Chrysene-d12	14.021	240	438761	20.000	ng	0.00
86) Perylene-d12	15.492	264	334338	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	824441	76.249	ng	0.00
7) Phenol-d6	6.492	99	1036923	76.296	ng	0.00
23) Nitrobenzene-d5	7.410	82	1009770	78.628	ng	0.00
42) 2,4,6-Tribromophenol	10.675	330	299353	81.619	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	1856679	75.289	ng	0.00
79) Terphenyl-d14	12.957	244	1966784	84.568	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.581	88	173633	35.592	ng	89
3) Pyridine	3.363	79	472583	34.883	ng	90
4) n-Nitrosodimethylamine	3.322	42	213396	35.870	ng	92
6) Aniline	6.510	93	596895	35.329	ng	# 30
8) 2-Chlorophenol	6.634	128	462823	39.179	ng	97
9) Benzaldehyde	6.392	77	313797	33.999	ng	98
10) Phenol	6.510	94	570965	39.451	ng	# 68
11) bis(2-Chloroethyl)ether	6.575	93	446382	38.891	ng	94
12) 1,3-Dichlorobenzene	6.781	146	496399	39.181	ng	97
13) 1,4-Dichlorobenzene	6.857	146	505828	39.258	ng	99
14) 1,2-Dichlorobenzene	7.010	146	468686	39.574	ng	98
15) Benzyl Alcohol	6.992	79	401893	40.773	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.116	45	589003	34.139	ng	# 40
17) 2-Methylphenol	7.110	107	374307	38.195	ng	# 88
18) Hexachloroethane	7.345	117	194579	40.106	ng	90
19) n-Nitroso-di-n-propyla...	7.257	70	332102	40.364	ng	92
20) 3+4-Methylphenols	7.263	107	491336	39.432	ng	# 89
22) Acetophenone	7.251	105	649449	39.790	ng	98
24) Nitrobenzene	7.428	77	509607	38.535	ng	98
25) Isophorone	7.669	82	881649	38.299	ng	97
26) 2-Nitrophenol	7.739	139	256463	39.310	ng	97
27) 2,4-Dimethylphenol	7.786	122	363460	37.140	ng	99
28) bis(2-Chloroethoxy)met...	7.869	93	514712	38.544	ng	99
29) 2,4-Dichlorophenol	7.992	162	396721	39.277	ng	98
30) 1,2,4-Trichlorobenzene	8.063	180	410114	39.226	ng	97
31) Naphthalene	8.145	128	1388113	38.973	ng	100
32) Benzoic acid	7.922	122	161429	22.818	ng	98
33) 4-Chloroaniline	8.198	127	567911	38.837	ng	98
34) Hexachlorobutadiene	8.251	225	249947	42.058	ng	99
35) Caprolactam	8.586	113	88836	27.052	ng	# 76
36) 4-Chloro-3-methylphenol	8.692	107	434419	40.551	ng	97
37) 2-Methylnaphthalene	8.839	142	926101	40.011	ng	100
38) 1-Methylnaphthalene	8.933	142	898377	40.206	ng	99
40) 1,2,4,5-Tetrachloroben...	9.004	216	398576	39.790	ng	99
41) Hexachlorocyclopentadiene	8.980	237	178715	40.066	ng	99
43) 2,4,6-Trichlorophenol	9.122	196	272535	37.460	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081022\
 Data File : BF129690.D
 Acq On : 10 Aug 2022 10:59
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Quant Time: Aug 10 15:27:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 15:26:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.169	196	296554	37.482	ng	# 93
46) 1,1'-Biphenyl	9.298	154	1092384	39.237	ng	100
47) 2-Chloronaphthalene	9.328	162	864160	38.396	ng	100
48) 2-Nitroaniline	9.433	65	285025	38.085	ng	89
49) Acenaphthylene	9.745	152	1323367	38.697	ng	100
50) Dimethylphthalate	9.604	163	1067357	40.530	ng	100
51) 2,6-Dinitrotoluene	9.675	165	231655	39.301	ng	96
52) Acenaphthene	9.916	154	803954	35.993	ng	100
53) 3-Nitroaniline	9.845	138	258740	37.174	ng	# 96
54) 2,4-Dinitrophenol	9.957	184	107874	35.850	ng	91
55) Dibenzofuran	10.086	168	1196489	38.858	ng	95
56) 4-Nitrophenol	10.027	139	191579	37.309	ng	96
57) 2,4-Dinitrotoluene	10.080	165	297890	38.687	ng	96
58) Fluorene	10.433	166	1010311	41.008	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.210	232	227464	37.202	ng	95
60) Diethylphthalate	10.298	149	1048296	41.178	ng	99
61) 4-Chlorophenyl-phenyle...	10.416	204	458492	42.213	ng	# 88
62) 4-Nitroaniline	10.469	138	252503	36.570	ng	94
63) Azobenzene	10.580	77	1007141	38.787	ng	95
65) 4,6-Dinitro-2-methylph...	10.492	198	161443	39.177	ng	93
66) n-Nitrosodiphenylamine	10.545	169	849651	39.078	ng	97
67) 4-Bromophenyl-phenylether	10.910	248	293850	41.102	ng	97
68) Hexachlorobenzene	10.980	284	321054	41.445	ng	98
69) Atrazine	11.069	200	273672	41.439	ng	95
70) Pentachlorophenol	11.180	266	152404	36.190	ng	98
71) Phenanthrene	11.398	178	1391484	39.044	ng	99
72) Anthracene	11.451	178	1383521	39.503	ng	99
73) Carbazole	11.610	167	1235160	37.465	ng	99
74) Di-n-butylphthalate	11.921	149	1673056	42.613	ng	99
75) Fluoranthene	12.586	202	1441724	39.925	ng	98
77) Benzidine	12.710	184	624010	40.258	ng	100
78) Pyrene	12.821	202	1438922	39.218	ng	99
80) Butylbenzylphthalate	13.427	149	691917	43.477	ng	91
81) Benzo(a)anthracene	14.010	228	1160522	38.721	ng	100
82) 3,3'-Dichlorobenzidine	13.968	252	360195	36.399	ng	99
83) Chrysene	14.051	228	1091148	38.629	ng	100
84) Bis(2-ethylhexyl)phtha...	13.980	149	942718	44.995	ng	98
85) Di-n-octyl phthalate	14.592	149	1399071	46.403	ng	96
87) Indeno(1,2,3-cd)pyrene	16.986	276	859190	38.161	ng	98
88) Benzo(b)fluoranthene	15.062	252	892188	40.229	ng	99
89) Benzo(k)fluoranthene	15.092	252	866314	39.861	ng	99
90) Benzo(a)pyrene	15.433	252	702580	39.025	ng	100
91) Dibenzo(a,h)anthracene	16.992	278	713539	38.938	ng	99
92) Benzo(g,h,i)perylene	17.433	276	706611	37.736	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 10 15:27:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 10 15:26:57 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.839 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 176136
Ion Ratio Lower Upper
152 100
150 155.8 123.7 185.5
115 62.7 49.9 74.9



#2
1,4-Dioxane
Concen: 35.592 ng
RT: 2.581 min Scan# 50
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 88 Resp: 173633
Ion Ratio Lower Upper
88 100
58 77.3 71.8 107.8
43 29.7 26.6 39.8









#3
Pyridine
Concen: 34.883 ng
RT: 3.363 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 472583
Ion Ratio Lower Upper
79 100
52 74.7 68.6 102.8
51 37.1 32.9 49.3



#4
n-Nitrosodimethylamine
Concen: 35.870 ng
RT: 3.322 min Scan# 176
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 42 Resp: 213396
Ion Ratio Lower Upper
42 100
74 128.0 94.8 142.2
44 6.7 6.6 10.0









#5
2-Fluorophenol
Concen: 76.249 ng
RT: 5.475 min Scan# 54
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 824441
Ion Ratio Lower Upper
112 100
64 59.1 53.3 79.9
63 31.1 28.0 42.0



#6
Aniline
Concen: 35.329 ng
RT: 6.510 min Scan# 718
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 93 Resp: 596895
Ion Ratio Lower Upper
93 100
66 84.0 32.5 48.7#
65 52.7 16.4 24.6#









#7
Phenol-d6
Concen: 76.296 ng
RT: 6.492 min Scan# 719
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 1036923
Ion Ratio Lower Upper
99 100
42 18.5 17.3 25.9
71 35.6 28.9 43.3



#8
2-Chlorophenol
Concen: 39.179 ng
RT: 6.634 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 128 Resp: 462823
Ion Ratio Lower Upper
128 100
130 32.3 13.1 53.1
64 47.9 30.9 70.9









#9
Benzaldehyde
Concen: 33.999 ng
RT: 6.392 min Scan# 69
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 313797
Ion Ratio Lower Upper
77 100
105 98.9 81.6 121.6
106 94.5 76.1 116.1



#10
Phenol
Concen: 39.451 ng
RT: 6.510 min Scan# 718
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 94 Resp: 570965
Ion Ratio Lower Upper
94 100
65 42.0 11.4 51.4
66 67.0 21.4 61.4#









#11
bis(2-Chloroethyl)ether
Concen: 38.891 ng
RT: 6.575 min Scan# 711
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 446382
Ion Ratio Lower Upper
93 100
63 79.5 67.4 107.4
95 32.8 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 39.181 ng
RT: 6.781 min Scan# 764
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 146 Resp: 496399
Ion Ratio Lower Upper
146 100
148 65.2 50.6 75.8
75 30.8 26.3 39.5









#13
1,4-Dichlorobenzene
Concen: 39.258 ng
RT: 6.857 min Scan# 71
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 505828
Ion Ratio Lower Upper
146 100
148 64.9 51.1 76.7
111 42.8 34.2 51.4



#14
1,2-Dichlorobenzene
Concen: 39.574 ng
RT: 7.010 min Scan# 803
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:146 Resp: 468686
Ion Ratio Lower Upper
146 100
148 64.2 51.8 77.8
111 46.5 35.1 52.7









#15
Benzyl Alcohol
Concen: 40.773 ng
RT: 6.992 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 401893
Ion Ratio Lower Upper
79 100
108 76.3 61.8 92.6
77 63.6 50.2 75.4



#16
2,2'-oxybis(1-Chloropropane)
Concen: 34.139 ng
RT: 7.116 min Scan# 821
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 45 Resp: 589003
Ion Ratio Lower Upper
45 100
77 33.3 0.0 31.3#
79 30.4 0.0 28.3#









#17
2-Methylphenol
Concen: 38.195 ng
RT: 7.110 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 374307
Ion Ratio Lower Upper
107 100
108 110.8 91.9 137.9
77 63.4 37.8 56.8#
79 58.6 37.4 56.0#



#18
Hexachloroethane
Concen: 40.106 ng
RT: 7.345 min Scan# 860
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:117 Resp: 194579
Ion Ratio Lower Upper
117 100
119 95.8 77.4 116.2
201 86.4 84.6 127.0









#19
n-Nitroso-di-n-propylamine
Concen: 40.364 ng
RT: 7.257 min Scan# 845
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 332102




Ion	Ratio	Lower	Upper
70	100		
42	56.6	52.3	78.5
101	10.1	7.5	11.3
130	21.3	16.1	24.1




#20
3+4-Methylphenols
Concen: 39.432 ng
RT: 7.263 min Scan# 846
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 107 Resp: 491336

Ion	Ratio	Lower	Upper
107	100		
108	92.3	71.8	111.8
77	55.9	59.3	99.3#
79	30.8	11.7	51.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.122 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 701049




Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.8	13.2
54	8.1	7.5	11.3
68	5.8	4.7	7.1




#22
Acetophenone
Concen: 39.790 ng
RT: 7.251 min Scan# 844
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:105 Resp: 649449

Ion	Ratio	Lower	Upper
105	100		
71	6.7	5.0	7.4
51	29.8	25.0	37.4
120	21.8	17.8	26.8





#23

Nitrobenzene-d5

Concen: 78.628 ng

RT: 7.410 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :





BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 1009770

Ion	Ratio	Lower	Upper
82	100		
128	44.0	34.0	51.0
54	51.2	46.9	70.3



#24

Nitrobenzene

Concen: 38.535 ng

RT: 7.428 min Scan# 874




Delta R.T. 0.000 min


Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Tgt Ion: 77 Resp: 509607

Ion	Ratio	Lower	Upper
77	100		
123	45.9	35.4	53.2
65	14.0	11.8	17.8









#25
Isophorone
Concen: 38.299 ng
RT: 7.669 min Scan# 915
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 881649
Ion Ratio Lower Upper
82 100
95 7.5 5.7 8.5
138 19.4 14.2 21.2



#26
2-Nitrophenol
Concen: 39.310 ng
RT: 7.739 min Scan# 927
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 139 Resp: 256463
Ion Ratio Lower Upper
139 100
109 30.9 25.0 37.4
65 53.1 45.1 67.7





#27

2,4-Dimethylphenol

Concen: 37.140 ng

RT: 7.786 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:122 Resp: 363460

Ion Ratio Lower Upper

122 100

107 115.1 93.0 139.6

121 59.1 46.6 70.0



#28

bis(2-Chloroethoxy)methane

Concen: 38.544 ng

RT: 7.869 min Scan# 949

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion: 93 Resp: 514712


Ion Ratio Lower Upper

93 100

95 32.7 26.8 40.2

123 11.0 9.2 13.8





#29

2,4-Dichlorophenol

Concen: 39.277 ng

RT: 7.992 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:162 Resp: 396721

Ion Ratio Lower Upper

162 100

164 65.5 44.2 84.2

98 37.5 19.6 59.6



#30

1,2,4-Trichlorobenzene

Concen: 39.226 ng

RT: 8.063 min Scan# 982

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:180 Resp: 410114


Ion Ratio Lower Upper

180 100

182 95.3 79.2 118.8

145 33.3 26.2 39.2









#31
Naphthalene
Concen: 38.973 ng
RT: 8.145 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1388113
Ion Ratio Lower Upper
128 100
129 11.5 9.2 13.8
127 13.4 10.5 15.7



#32
Benzoic acid
Concen: 22.818 ng
RT: 7.922 min Scan# 958
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:122 Resp: 161429
Ion Ratio Lower Upper
122 100
105 124.8 107.1 147.1
77 91.3 74.2 114.2









#33
4-Chloroaniline
Concen: 38.837 ng
RT: 8.198 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	567911
Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.4	39.6
65	30.8	26.4	39.6
92	19.7	16.0	24.0



#34
Hexachlorobutadiene
Concen: 42.058 ng
RT: 8.251 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:	225	Resp:	249947
Ion	Ratio	Lower	Upper
225	100		
223	63.6	51.0	76.6
227	62.3	51.1	76.7









#35
 Caprolactam
 Concen: 27.052 ng
 RT: 8.586 min Scan# 1077
 Delta R.T. 0.000 min
 Lab File: BF129690.D
 Acq: 10 Aug 2022 10:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:113 Resp: 88836
 Ion Ratio Lower Upper
 113 100
 55 168.6 188.5 228.5#
 56 123.0 130.6 170.6#



#36
 4-Chloro-3-methylphenol
 Concen: 40.551 ng
 RT: 8.692 min Scan# 1089
 Delta R.T. 0.000 min
 Lab File: BF129690.D
 Acq: 10 Aug 2022 10:59

Tgt Ion:107 Resp: 434419
 Ion Ratio Lower Upper
 107 100
 144 24.6 19.5 29.3
 142 76.5 58.2 87.4





#37

2-Methylnaphthalene

Concen: 40.011 ng

RT: 8.839 min Scan# 1114

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:142 Resp: 926101

Ion Ratio Lower Upper

142 100

141 88.5 71.0 106.4



#38

1-Methylnaphthalene

Concen: 40.206 ng

RT: 8.933 min Scan# 1130

Delta R.T. 0.000 min

Lab File: BF129690.D




Acq: 10 Aug 2022 10:59


Tgt Ion:142 Resp: 898377

Ion Ratio Lower Upper

142 100

141 92.2 74.2 111.4









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.880 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 381486
Ion Ratio Lower Upper
164 100
162 101.5 79.5 119.3
160 45.1 34.2 51.4



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.790 ng
RT: 9.004 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:216 Resp: 398576
Ion Ratio Lower Upper
216 100
214 79.6 63.6 95.4
179 23.3 17.0 25.6
108 20.4 16.0 24.0









#41
Hexachlorocyclopentadiene
Concen: 40.066 ng
RT: 8.980 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:237 Resp: 178715
Ion Ratio Lower Upper
237 100
235 63.9 43.2 83.2
272 11.9 0.0 31.9



#42
2,4,6-Tribromophenol
Concen: 81.619 ng
RT: 10.675 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:330 Resp: 299353
Ion Ratio Lower Upper
330 100
332 101.7 76.8 115.2
141 34.4 28.0 42.0





#43

2,4,6-Trichlorophenol

Concen: 37.460 ng

RT: 9.122 min Scan# 1162

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:196 Resp: 272535

Ion Ratio Lower Upper

196 100

198 93.7 74.0 111.0

200 29.9 24.1 36.1



#44

2,4,5-Trichlorophenol

Concen: 37.482 ng

RT: 9.169 min Scan# 1170

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Tgt Ion:196 Resp: 296554




Ion Ratio Lower Upper


196 100

198 96.3 75.7 113.5

97 57.3 36.4 54.6#

132 30.9 20.7 31.1





#45

2-Fluorobiphenyl

Concen: 75.289 ng

RT: 9.198 min Scan# 1183

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:172 Resp: 1856679

Ion Ratio Lower Upper

172 100

171 35.5 28.8 43.2

170 23.9 19.8 29.8



#46

1,1'-Biphenyl

Concen: 39.237 ng

RT: 9.298 min Scan# 1192

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:154 Resp: 1092384


Ion Ratio Lower Upper

154 100

153 40.6 20.6 60.6

76 13.5 0.0 33.3









#47
2-Chloronaphthalene
Concen: 38.396 ng
RT: 9.328 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 864160
Ion Ratio Lower Upper
162 100
127 42.9 34.5 51.7
164 33.2 26.5 39.7



#48
2-Nitroaniline
Concen: 38.085 ng
RT: 9.433 min Scan# 1215
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 65 Resp: 285025
Ion Ratio Lower Upper
65 100
92 70.7 51.9 77.9
138 109.4 76.4 114.6





#49

Acenaphthylene

Concen: 38.697 ng

RT: 9.745 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:152 Resp: 1323367

Ion Ratio Lower Upper

152 100

151 20.0 16.2 24.4

153 13.3 10.7 16.1



#50

Dimethylphthalate

Concen: 40.530 ng

RT: 9.604 min Scan# 1244

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:163 Resp: 1067357


Ion Ratio Lower Upper

163 100

194 3.9 3.2 4.8

164 10.1 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 39.301 ng
RT: 9.675 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 231655
Ion Ratio Lower Upper
165 100
63 51.3 44.9 67.3
89 53.5 43.6 65.4



#52
Acenaphthene
Concen: 35.993 ng
RT: 9.916 min Scan# 1297
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:154 Resp: 803954
Ion Ratio Lower Upper
154 100
153 112.0 89.6 134.4
152 52.8 42.4 63.6









#53
3-Nitroaniline
Concen: 37.174 ng
RT: 9.845 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 258740
Ion Ratio Lower Upper
138 100
108 10.7 10.9 16.3#
92 119.2 98.2 147.2



#54
2,4-Dinitrophenol
Concen: 35.850 ng
RT: 9.957 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:184 Resp: 107874
Ion Ratio Lower Upper
184 100
63 68.9 64.5 96.7
154 65.8 50.6 75.8









#55
Dibenzenofuran
Concen: 38.858 ng
RT: 10.086 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1196489
Ion Ratio Lower Upper
168 100
139 42.1 30.6 46.0
169 13.9 11.0 16.6



#56
4-Nitrophenol
Concen: 37.309 ng
RT: 10.027 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:139 Resp: 191579
Ion Ratio Lower Upper
139 100
109 72.9 50.7 90.7
65 99.4 74.5 114.5









#57
2,4-Dinitrotoluene
Concen: 38.687 ng
RT: 10.080 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 297890
Ion Ratio Lower Upper
165 100
63 55.2 39.8 59.6
89 76.2 60.2 90.2
182 2.6 2.2 3.2



#58
Fluorene
Concen: 41.008 ng
RT: 10.433 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:166 Resp: 1010311
Ion Ratio Lower Upper
166 100
165 96.2 78.1 117.1
167 13.4 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 37.202 ng
RT: 10.210 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 227464
Ion Ratio Lower Upper
232 100
131 49.9 36.1 54.1
130 2.7 1.8 2.8
166 31.6 24.2 36.2



#60
Diethylphthalate
Concen: 41.178 ng
RT: 10.298 min Scan# 1362
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:149 Resp: 1048296
Ion Ratio Lower Upper
149 100
177 21.2 17.3 25.9
150 12.8 10.2 15.2









#61
4-Chlorophenyl-phenylether
Concen: 42.213 ng
RT: 10.416 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 458492
Ion Ratio Lower Upper
204 100
206 32.2 25.8 38.6
141 72.3 47.0 70.4#



#62
4-Nitroaniline
Concen: 36.570 ng
RT: 10.469 min Scan# 1391
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:138 Resp: 252503
Ion Ratio Lower Upper
138 100
92 48.6 33.5 73.5
108 76.1 61.0 101.0









#63
Azobenzene
Concen: 38.787 ng
RT: 10.580 min Scan# 1418
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 1007141
Ion Ratio Lower Upper
77 100
182 24.9 5.6 45.6
105 21.9 1.6 41.6
51 29.7 15.1 55.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.369 min Scan# 1544
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion: 188 Resp: 652882
Ion Ratio Lower Upper
188 100
94 10.5 7.0 10.6
80 11.1 7.6 11.4









#65
4,6-Dinitro-2-methylphenol
Concen: 39.177 ng
RT: 10.492 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 161443
Ion Ratio Lower Upper
198 100
51 52.6 37.6 77.6
105 51.7 27.0 67.0



#66
n-Nitrosodiphenylamine
Concen: 39.078 ng
RT: 10.545 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:169 Resp: 849651
Ion Ratio Lower Upper
169 100
168 65.8 50.1 75.1
167 36.2 27.9 41.9









#67
4-Bromophenyl-phenylether
Concen: 41.102 ng
RT: 10.910 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 293850
Ion Ratio Lower Upper
248 100
250 97.7 80.6 120.8
141 76.3 58.9 88.3



#68
Hexachlorobenzene
Concen: 41.445 ng
RT: 10.980 min Scan# 1478
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:284 Resp: 321054
Ion Ratio Lower Upper
284 100
142 35.5 29.0 43.4
249 32.6 25.0 37.6









#69
Atrazine
Concen: 41.439 ng
RT: 11.069 min Scan# 1493
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 273672
Ion Ratio Lower Upper
200 100
173 27.6 6.3 46.3
215 43.8 28.3 68.3



#70
Pentachlorophenol
Concen: 36.190 ng
RT: 11.180 min Scan# 1512
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:266 Resp: 152404
Ion Ratio Lower Upper
266 100
268 64.1 51.0 76.4
264 64.0 49.4 74.2









#71
Phenanthrene
Concen: 39.044 ng
RT: 11.398 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1391484
Ion Ratio Lower Upper
178 100
176 19.2 16.0 24.0
179 15.6 12.6 19.0



#72
Anthracene
Concen: 39.503 ng
RT: 11.451 min Scan# 1558
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:178 Resp: 1383521
Ion Ratio Lower Upper
178 100
176 18.8 15.4 23.2
179 15.7 12.6 19.0









#73
 Carbazole
 Concen: 37.465 ng
 RT: 11.610 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BF129690.D
 Acq: 10 Aug 2022 10:59




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 1235160
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.7 26.5
 139 13.7 11.2 16.8



#74
 Di-n-butylphthalate
 Concen: 42.613 ng
 RT: 11.921 min Scan# 1638
 Delta R.T. 0.000 min
 Lab File: BF129690.D
 Acq: 10 Aug 2022 10:59

Tgt Ion:149 Resp: 1673056
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.5 11.3
 104 5.3 4.4 6.6









#75
Fluoranthene
Concen: 39.925 ng
RT: 12.586 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 1441724
Ion Ratio Lower Upper
202 100
101 12.8 0.0 31.4
203 17.9 0.0 38.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.021 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:240 Resp: 438761
Ion Ratio Lower Upper
240 100
120 10.7 8.9 13.3
236 25.5 20.8 31.2









#77
Benzidine
Concen: 40.258 ng
RT: 12.710 min Scan# 1772
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 624010
Ion Ratio Lower Upper
184 100
185 14.8 11.9 17.9
183 11.5 9.3 13.9



#78
Pyrene
Concen: 39.218 ng
RT: 12.821 min Scan# 1791
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:202 Resp: 1438922
Ion Ratio Lower Upper
202 100
200 21.5 17.7 26.5
203 17.7 14.6 21.8





#79

Terphenyl-d14

Concen: 84.568 ng

RT: 12.957 min Scan# 1814

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:244 Resp: 1966784

Ion Ratio Lower Upper

244 100

212 7.2 6.1 9.1

122 12.4 8.8 13.2



#80

Butylbenzylphthalate

Concen: 43.477 ng

RT: 13.427 min Scan# 1894

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:149 Resp: 691917


Ion Ratio Lower Upper

149 100

91 67.3 61.2 91.8

206 19.8 15.7 23.5









#81
Benzo(a)anthracene
Concen: 38.721 ng
RT: 14.010 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 1160522
Ion Ratio Lower Upper
228 100
226 27.4 22.1 33.1
229 19.9 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 36.399 ng
RT: 13.968 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:252 Resp: 360195
Ion Ratio Lower Upper
252 100
254 63.4 50.2 75.4
126 12.3 8.2 12.4





#83

Chrysene

Concen: 38.629 ng

RT: 14.051 min Scan# 2000

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:228 Resp: 1091148

Ion Ratio Lower Upper

228 100

226 30.1 24.3 36.5

229 20.0 15.9 23.9



#84

Bis(2-ethylhexyl)phthalate

Concen: 44.995 ng

RT: 13.980 min Scan# 1988

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:149 Resp: 942718


Ion Ratio Lower Upper

149 100

167 27.2 21.0 31.4

279 4.5 3.1 4.7





#85

Di-n-octyl phthalate

Concen: 46.403 ng

RT: 14.592 min Scan# 2092

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:149 Resp: 1399071

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9

43 9.8 9.2 13.8



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.492 min Scan# 2245

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:264 Resp: 334338


Ion Ratio Lower Upper

264 100

260 23.0 18.4 27.6

265 21.2 17.2 25.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 38.161 ng
RT: 16.986 min Scan# 2499
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 859190
Ion Ratio Lower Upper
276 100
138 24.9 21.1 31.7
277 25.2 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 40.229 ng
RT: 15.062 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:252 Resp: 892188
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 10.1 8.0 12.0





#89

Benzo(k)fluoranthene

Concen: 39.861 ng

RT: 15.092 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:252 Resp: 866314

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6

125 10.2 8.3 12.5



#90

Benzo(a)pyrene

Concen: 39.025 ng

RT: 15.433 min Scan# 2235

Delta R.T. 0.000 min

Lab File: BF129690.D

Acq: 10 Aug 2022 10:59




Tgt Ion:252 Resp: 702580


Ion Ratio Lower Upper

252 100

253 21.5 17.0 25.6

125 11.0 8.8 13.2









#91
Dibenzo(a,h)anthracene
Concen: 38.938 ng
RT: 16.992 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 713539
Ion Ratio Lower Upper
278 100
139 16.6 12.8 19.2
279 22.9 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 37.736 ng
RT: 17.433 min Scan# 2575
Delta R.T. 0.000 min
Lab File: BF129690.D
Acq: 10 Aug 2022 10:59

Tgt Ion:276 Resp: 706611
Ion Ratio Lower Upper
276 100
277 23.4 19.1 28.7
138 21.8 18.0 27.0

