

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081222\
 Data File : BF129744.D
 Acq On : 12 Aug 2022 15:05
 Operator : CG\JU
 Sample : N4151-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 348

Quant Time: Aug 12 15:55:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 12 15:52:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	141016	20.000	ng	0.00
21) Naphthalene-d8	8.110	136	530242	20.000	ng	0.00
39) Acenaphthene-d10	9.869	164	237320	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	321491	20.000	ng	0.00
76) Chrysene-d12	14.010	240	288068	20.000	ng	0.00
86) Perylene-d12	15.480	264	266987	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	664974	76.817	ng	0.00
7) Phenol-d6	6.487	99	838870	77.096	ng	0.00
23) Nitrobenzene-d5	7.398	82	554717	57.108	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	130490	57.191	ng	0.00
45) 2-Fluorobiphenyl	9.186	172	982764	64.061	ng	0.00
79) Terphenyl-d14	12.945	244	784023	51.346	ng	0.00
Target Compounds						
78) Pyrene	12.804	202	49252	2.045	ng	97
84) Bis(2-ethylhexyl)phtha...	13.968	149	529103	38.464	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081222\
Data File : BF129744.D
Acq On : 12 Aug 2022 15:05
Operator : CG\JU
Sample : N4151-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
348

Quant Time: Aug 12 15:55:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 12 15:52:45 2022
Response via : Initial Calibration

