

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092022\
 Data File : BF130341.D
 Acq On : 20 Sep 2022 13:18
 Operator : CG\JU
 Sample : PB147677BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB147677BL

Quant Time: Sep 21 05:12:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 15:06:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.657	152	108324	20.000	ng	0.00
21) Naphthalene-d8	7.939	136	429050	20.000	ng	0.00
39) Acenaphthene-d10	9.692	164	237434	20.000	ng	0.00
64) Phenanthrene-d10	11.169	188	407813	20.000	ng	0.00
76) Chrysene-d12	13.798	240	293576	20.000	ng	0.00
86) Perylene-d12	15.192	264	224494	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.275	112	905952	145.059	ng	0.02
7) Phenol-d6	6.304	99	1108943	141.910	ng	0.00
23) Nitrobenzene-d5	7.228	82	725623	86.402	ng	0.00
42) 2,4,6-Tribromophenol	10.480	330	366955	161.294	ng	0.00
45) 2-Fluorobiphenyl	9.022	172	1374965	84.327	ng	0.00
79) Terphenyl-d14	12.763	244	1628719	91.900	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092022\
Data File : BF130341.D
Acq On : 20 Sep 2022 13:18
Operator : CG\JU
Sample : PB147677BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB147677BL

Quant Time: Sep 21 05:12:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 15:06:14 2022
Response via : Initial Calibration

