

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092022\
 Data File : BF130343.D
 Acq On : 20 Sep 2022 14:25
 Operator : CG\JU
 Sample : PB147759BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB147759BL

Quant Time: Sep 21 05:13:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 15:06:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.657	152	107520	20.000	ng	0.00
21) Naphthalene-d8	7.939	136	438216	20.000	ng	# 0.00
39) Acenaphthene-d10	9.686	164	235956	20.000	ng	0.00
64) Phenanthrene-d10	11.169	188	418306	20.000	ng	0.00
76) Chrysene-d12	13.798	240	295059	20.000	ng	0.00
86) Perylene-d12	15.186	264	223991	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.275	112	922738	148.852	ng	0.02
7) Phenol-d6	6.304	99	1127410	145.352	ng	0.00
23) Nitrobenzene-d5	7.222	82	728186	84.894	ng	-0.01
42) 2,4,6-Tribromophenol	10.480	330	374470	165.628	ng	0.00
45) 2-Fluorobiphenyl	9.022	172	1396533	86.186	ng	0.00
79) Terphenyl-d14	12.763	244	1648558	92.552	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF092022\
Data File : BF130343.D
Acq On : 20 Sep 2022 14:25
Operator : CG\JU
Sample : PB147759BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB147759BL

Quant Time: Sep 21 05:13:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 15:06:14 2022
Response via : Initial Calibration

