

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102022\
 Data File : BF130734.D
 Acq On : 20 Oct 2022 18:46
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102022

Quant Time: Oct 21 06:19:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 06:12:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	95	0.00
2	1,4-Dioxane	40.000	39.254	1.9	84	0.00
3	Pyridine	40.000	46.013	-15.0	92	0.00
4	n-Nitrosodimethylamine	40.000	44.172	-10.4	86	0.00
5 S	2-Fluorophenol	80.000	87.756	-9.7	96	0.00
6	Aniline	40.000	43.112	-7.8	96	0.00
7 S	Phenol-d6	80.000	82.389	-3.0	97	0.00
8	2-Chlorophenol	40.000	41.057	-2.6	97	0.00
9	Benzaldehyde	40.000	38.483	3.8	96	0.00
10 C	Phenol	40.000	42.731	-6.8	96	0.00
11	bis(2-Chloroethyl)ether	40.000	41.162	-2.9	95	0.00
12	1,3-Dichlorobenzene	40.000	41.191	-3.0	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.714	-1.8	96	0.00
14	1,2-Dichlorobenzene	40.000	40.130	-0.3	97	0.00
15	Benzyl Alcohol	40.000	42.179	-5.4	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.775	3.1	96	0.00
17	2-Methylphenol	40.000	39.563	1.1	97	0.00
18	Hexachloroethane	40.000	39.781	0.5	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.317	1.7	98	0.00
20	3+4-Methylphenols	40.000	41.308	-3.3	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	38.918	2.7	98	0.00
23 S	Nitrobenzene-d5	80.000	76.957	3.8	97	0.00
24	Nitrobenzene	40.000	38.299	4.3	99	0.00
25	Isophorone	40.000	39.571	1.1	99	0.00
26 C	2-Nitrophenol	40.000	39.652	0.9	98	0.00
27	2,4-Dimethylphenol	40.000	40.154	-0.4	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.963	2.6	99	0.00
29 C	2,4-Dichlorophenol	40.000	41.952	-4.9	98	0.00
30	1,2,4-Trichlorobenzene	40.000	42.476	-6.2	99	0.00
31	Naphthalene	40.000	40.975	-2.4	99	0.00
32	Benzoic acid	40.000	38.130	4.7	91	0.00
33	4-Chloroaniline	40.000	41.976	-4.9	99	0.00
34 C	Hexachlorobutadiene	40.000	41.120	-2.8	98	0.00
35	Caprolactam	40.000	42.578	-6.4	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.173	-2.9	99	0.00
37	2-Methylnaphthalene	40.000	40.909	-2.3	99	0.00
38	1-Methylnaphthalene	40.000	40.836	-2.1	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.194	-5.5	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.994	-5.0	97	0.00
42 S	2,4,6-Tribromophenol	80.000	88.800	-11.0	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.805	-4.5	96	0.00
44	2,4,5-Trichlorophenol	40.000	42.850	-7.1	96	0.00
45 S	2-Fluorobiphenyl	80.000	81.762	-2.2	99	0.00
46	1,1'-Biphenyl	40.000	41.813	-4.5	100	0.00
47	2-Chloronaphthalene	40.000	40.818	-2.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102022\
 Data File : BF130734.D
 Acq On : 20 Oct 2022 18:46
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF102022

Quant Time: Oct 21 06:19:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 21 06:12:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.591	-4.0	100	0.00
49	Acenaphthylene	40.000	42.268	-5.7	99	0.00
50	Dimethylphthalate	40.000	42.712	-6.8	100	0.00
51	2,6-Dinitrotoluene	40.000	42.383	-6.0	101	0.00
52 C	Acenaphthene	40.000	41.136	-2.8	99	0.00
53	3-Nitroaniline	40.000	40.915	-2.3	100	0.00
54 P	2,4-Dinitrophenol	40.000	41.616	-4.0	106	0.00
55	Dibenzofuran	40.000	41.405	-3.5	100	0.00
56 P	4-Nitrophenol	40.000	43.114	-7.8	96	0.00
57	2,4-Dinitrotoluene	40.000	42.276	-5.7	100	0.00
58	Fluorene	40.000	41.019	-2.5	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.085	-7.7	94	0.00
60	Diethylphthalate	40.000	40.486	-1.2	101	0.00
61	4-Chlorophenyl-phenylether	40.000	41.829	-4.6	100	0.00
62	4-Nitroaniline	40.000	41.814	-4.5	102	0.00
63	Azobenzene	40.000	42.394	-6.0	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.710	-6.8	99	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.380	4.0	99	0.00
67	4-Bromophenyl-phenylether	40.000	39.090	2.3	98	0.00
68	Hexachlorobenzene	40.000	39.569	1.1	101	0.00
69	Atrazine	40.000	41.235	-3.1	101	0.00
70 C	Pentachlorophenol	40.000	42.665	-6.7	101	0.00
71	Phenanthrene	40.000	41.165	-2.9	101	0.00
72	Anthracene	40.000	40.884	-2.2	102	0.00
73	Carbazole	40.000	40.605	-1.5	104	0.00
74	Di-n-butylphthalate	40.000	39.411	1.5	102	0.00
75 C	Fluoranthene	40.000	49.450	-23.6#	120	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	110	0.00
77	Benzydine	40.000	28.892	27.8#	86	0.00
78	Pyrene	40.000	47.818	-19.5	118	0.00
79 S	Terphenyl-d14	80.000	98.372	-23.0	121	0.00
80	Butylbenzylphthalate	40.000	47.974	-19.9	123	0.00
81	Benzo(a)anthracene	40.000	41.257	-3.1	111	0.00
82	3,3'-Dichlorobenzidine	40.000	37.659	5.9	104	0.00
83	Chrysene	40.000	40.962	-2.4	112	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.296	-5.7	113	0.00
85 c	Di-n-octyl phthalate	40.000	48.438	-21.1#	134	0.00
86 I	Perylene-d12	20.000	20.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.914	7.7	96	0.00
88	Benzo(b)fluoranthene	40.000	36.939	7.7	112	0.00
89	Benzo(k)fluoranthene	40.000	38.788	3.0	118	0.00
90 C	Benzo(a)pyrene	40.000	40.647	-1.6	118	0.00
91	Dibenzo(a,h)anthracene	40.000	36.555	8.6	95	0.00
92	Benzo(g,h,i)perylene	40.000	36.414	9.0	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102022\
Data File : BF130734.D
Acq On : 20 Oct 2022 18:46
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF102022

Quant Time: Oct 21 06:19:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 21 06:12:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 2