

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102521\
 Data File : BF125915.D
 Acq On : 25 Oct 2021 13:52
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 25 17:25:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 17:23:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	140989	20.000	ng	0.00
21) Naphthalene-d8	8.151	136	553650	20.000	ng	# 0.00
39) Acenaphthene-d10	9.904	164	307287	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	550830	20.000	ng	0.00
76) Chrysene-d12	14.027	240	368025	20.000	ng	# 0.00
86) Perylene-d12	15.492	264	263476	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	221776	25.697	ng	0.00
7) Phenol-d6	6.481	99	292020	26.284	ng	-0.01
23) Nitrobenzene-d5	7.428	82	219492	24.643	ng	0.00
42) 2,4,6-Tribromophenol	10.686	330	55374	18.252	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	440698	24.737	ng	0.00
79) Terphenyl-d14	12.974	244	472857	19.572	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.634	88	51855	14.111	ng	# 100
3) Pyridine	3.398	79	129912	13.665	ng	# 85
4) n-Nitrosodimethylamine	3.310	42	83581	24.424	ng	# 77
6) Aniline	6.528	93	168740	13.192	ng	# 92
8) 2-Chlorophenol	6.651	128	112487	11.900	ng	82
9) Benzaldehyde	6.416	77	97634	16.193	ng	93
10) Phenol	6.492	94	148846	13.777	ng	91
11) bis(2-Chloroethyl)ether	6.598	93	116426	13.300	ng	# 81
12) 1,3-Dichlorobenzene	6.810	146	123642	11.990	ng	97
13) 1,4-Dichlorobenzene	6.886	146	123058	11.701	ng	96
14) 1,2-Dichlorobenzene	7.039	146	118335	12.015	ng	97
15) Benzyl Alcohol	7.004	79	109630	14.889	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.145	45	274465	24.516	ng	94
17) 2-Methylphenol	7.110	107	91151	12.113	ng	92
18) Hexachloroethane	7.386	117	45833	12.656	ng	93
19) n-Nitroso-di-n-propyla...	7.275	70	82099	13.998	ng	# 93
20) 3+4-Methylphenols	7.263	107	122671	13.218	ng	88
22) Acetophenone	7.275	105	164264	13.497	ng	# 87
24) Nitrobenzene	7.445	77	118681	13.749	ng	88
25) Isophorone	7.686	82	218139	13.824	ng	# 87
26) 2-Nitrophenol	7.763	139	29647	6.792	ng	# 66
27) 2,4-Dimethylphenol	7.798	122	90227	12.413	ng	90
28) bis(2-Chloroethoxy)met...	7.898	93	145118	13.249	ng	# 96
29) 2,4-Dichlorophenol	7.998	162	84761	10.824	ng	93
30) 1,2,4-Trichlorobenzene	8.092	180	103085	12.005	ng	95
31) Naphthalene	8.175	128	323892	12.099	ng	98
32) Benzoic acid	7.863	122	38201	7.474	ng	94
33) 4-Chloroaniline	8.216	127	133256	11.895	ng	# 89
34) Hexachlorobutadiene	8.298	225	63882	13.318	ng	98
35) Caprolactam	8.557	113	26949	11.911	ng	# 42
36) 4-Chloro-3-methylphenol	8.686	107	97049	12.686	ng	89
37) 2-Methylnaphthalene	8.863	142	208546	12.029	ng	90
38) 1-Methylnaphthalene	8.963	142	202499	12.037	ng	90
40) 1,2,4,5-Tetrachloroben...	9.028	216	96359	11.684	ng	98
41) Hexachlorocyclopentadiene	9.022	237	46877	11.581	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102521\
 Data File : BF125915.D
 Acq On : 25 Oct 2021 13:52
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Oct 25 17:25:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 17:23:45 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	9.133	196	61629	10.310	ng	92
44) 2,4,5-Trichlorophenol	9.169	196	65018	10.230	ng	94
46) 1,1'-Biphenyl	9.328	154	258526	11.443	ng	97
47) 2-Chloronaphthalene	9.351	162	194960	10.628	ng	97
48) 2-Nitroaniline	9.439	65	54930	12.473	ng #	68
49) Acenaphthylene	9.763	152	297159	11.062	ng	98
50) Dimethylphthalate	9.622	163	213197	10.539	ng	97
51) 2,6-Dinitrotoluene	9.680	165	35957	8.794	ng #	67
52) Acenaphthene	9.939	154	187671	11.182	ng	99
53) 3-Nitroaniline	9.845	138	43143	8.924	ng #	63
54) 2,4-Dinitrophenol	9.945	184	8145	4.693	ng #	1
55) Dibenzofuran	10.110	168	288915	11.491	ng	92
56) 4-Nitrophenol	9.992	139	33433	9.277	ng #	74
57) 2,4-Dinitrotoluene	10.080	165	41847	8.094	ng #	94
58) Fluorene	10.451	166	213361	11.587	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.222	232	52197	10.880	ng #	89
60) Diethylphthalate	10.327	149	216952	11.240	ng	98
61) 4-Chlorophenyl-phenyle...	10.445	204	103786	11.701	ng	88
62) 4-Nitroaniline	10.451	138	38762	8.676	ng	91
63) Azobenzene	10.604	77	249606	13.758	ng	96
65) 4,6-Dinitro-2-methylph...	10.486	198	13737	5.000	ng	94
66) n-Nitrosodiphenylamine	10.557	169	187346	9.970	ng	99
67) 4-Bromophenyl-phenylether	10.933	248	63799	10.458	ng	95
68) Hexachlorobenzene	10.998	284	68595	9.922	ng	95
69) Atrazine	11.086	200	56813	10.417	ng	89
70) Pentachlorophenol	11.186	266	34730	9.398	ng	92
71) Phenanthrene	11.410	178	328671	11.152	ng	97
72) Anthracene	11.463	178	326806	11.102	ng	99
73) Carbazole	11.616	167	314598	11.484	ng	97
74) Di-n-butylphthalate	11.957	149	328686	10.868	ng	99
75) Fluoranthene	12.598	202	337455	12.690	ng	98
77) Benzidine	12.721	184	123838	9.195	ng	99
78) Pyrene	12.827	202	347995	9.307	ng	97
80) Butylbenzylphthalate	13.457	149	103610	8.645	ng #	83
81) Benzo(a)anthracene	14.015	228	270768	11.340	ng	100
82) 3,3'-Dichlorobenzidine	13.980	252	81523	10.806	ng #	96
83) Chrysene	14.051	228	269249	11.409	ng	98
84) Bis(2-ethylhexyl)phtha...	14.021	149	120974	9.277	ng #	96
85) Di-n-octyl phthalate	14.633	149	165642	7.624	ng #	90
87) Indeno(1,2,3-cd)pyrene	16.951	276	165918	8.251	ng	94
88) Benzo(b)fluoranthene	15.092	252	192598	12.948	ng	98
89) Benzo(k)fluoranthene	15.092	252	192598	12.611	ng	98
90) Benzo(a)pyrene	15.427	252	146751	11.140	ng	97
91) Dibenzo(a,h)anthracene	16.974	278	137972	8.318	ng	93
92) Benzo(g,h,i)perylene	17.392	276	139319	8.224	ng #	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102521\
 Data File : BF125915.D
 Acq On : 25 Oct 2021 13:52
 Operator : JU/CG
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleID :
 SSTDICC010

Quant Time: Oct 25 17:25:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF102521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 17:23:45 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.869 min Scan# 813
Delta R.T. -0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:152 Resp: 140989
Ion Ratio Lower Upper
152 100
150 158.0 117.6 176.4
115 69.0 47.1 70.7



1,4-Dioxane
Concen: 14.111 ng
RT: 2.634 min Scan# 93
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion: 88 Resp: 51855
Ion Ratio Lower Upper
88 100
58 110.0 0.0 0.0#
43 41.3 0.0 0.0#









Pyridine
 Concen: 13.665 ng
 RT: 3.398 min Scan# 223
 Delta R.T. 0.000 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	79	Resp:	129912
Ion	Ratio	Lower	Upper
79	100		
52	108.7	70.3	105.5#
51	53.7	44.1	66.1



n-Nitrosodimethylamine
 Concen: 24.424 ng
 RT: 3.310 min Scan# 208
 Delta R.T. -0.018 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	42	Resp:	83581
Ion	Ratio	Lower	Upper
42	100		
74	93.2	59.0	88.6#
44	9.6	0.0	0.0#









2-Fluorophenol
 Concen: 25.697 ng
 RT: 5.475 min Scan# 576
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	112	Resp:	221776
Ion	Ratio	Lower	Upper
112	100		
64	70.5	44.2	66.2#
63	36.6	26.2	39.2



Aniline
 Concen: 13.192 ng
 RT: 6.528 min Scan# 755
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	93	Resp:	168740
Ion	Ratio	Lower	Upper
93	100		
66	44.4	32.7	49.1
65	23.2	14.2	21.4#









Phenol-d6
 Concen: 26.284 ng
 RT: 6.481 min Scan# 747
 Delta R.T. -0.012 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	99	Resp:	292020
Ion	Ratio	Lower	Upper
99	100		
42	27.5	23.5	35.3
71	36.6	25.6	38.4



2-Chlorophenol
 Concen: 11.900 ng
 RT: 6.651 min Scan# 776
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	128	Resp:	112487
Ion	Ratio	Lower	Upper
128	100		
130	29.8	10.5	50.5
64	59.3	20.3	60.3









Benzaldehyde
 Concen: 16.193 ng
 RT: 6.416 min Scan# 736
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	77	Resp:	97634
Ion	Ratio	Lower	Upper
77	100		
105	94.0	75.9	115.9
106	91.7	84.0	124.0



Phenol
 Concen: 13.777 ng
 RT: 6.492 min Scan# 749
 Delta R.T. -0.018 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52


Tgt Ion:	94	Resp:	148846
Ion	Ratio	Lower	Upper
94	100		
65	35.1	9.3	49.3
66	45.1	20.3	60.3






Abundance Scan 768 (6.604 min): BF125917.D\data.ms (-764) #1

bis(2-Chloroethyl)ether
Concen: 13.300 ng
RT: 6.598 min Scan# 767
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	93	Resp:	116426
Ion	Ratio	Lower	Upper
93	100		
63	97.1	54.4	94.4#
95	31.6	11.4	51.4






Abundance Scan 803 (6.810 min): BF125917.D\data.ms (-799) #2

1,3-Dichlorobenzene
Concen: 11.990 ng
RT: 6.810 min Scan# 803
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion:	146	Resp:	123642
Ion	Ratio	Lower	Upper
146	100		
148	63.1	49.7	74.5
75	35.1	25.2	37.8






Abundance Scan 817 (6.892 min): BF125917.D\data.ms (-811) #13

1,4-Dichlorobenzene
Concen: 11.701 ng
RT: 6.886 min Scan# 816
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	146	Resp:	123058
Ion	Ratio	Lower	Upper
146	100		
148	66.0	51.9	77.9
111	44.4	31.8	47.6






Abundance Scan 843 (7.045 min): BF125917.D\data.ms (-837) #14

1,2-Dichlorobenzene
Concen: 12.015 ng
RT: 7.039 min Scan# 842
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion:	146	Resp:	118335
Ion	Ratio	Lower	Upper
146	100		
148	62.6	50.6	75.8
111	46.3	33.5	50.3





Abundance Scan 837 (7.010 min): BF125917.D\data.ms (-831) #15

Benzyl Alcohol
Concen: 14.889 ng
RT: 7.004 min Scan# 836
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Instrument :
BNA_F
ClientSampleId :
SSTDIC010



Tgt Ion:	79	Resp:	109630
Ion	Ratio	Lower	Upper
79	100		
108	76.2	60.4	90.6
77	63.2	51.5	77.3






Abundance Scan 836 (7.004 min): BF125915.D\data.ms (-786) (-)




Abundance Scan 861 (7.151 min): BF125917.D\data.ms (-851) #16

2,2'-oxybis(1-Chloropropane)
Concen: 24.516 ng
RT: 7.145 min Scan# 860
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	45	Resp:	274465
Ion	Ratio	Lower	Upper
45	100		
77	8.2	0.0	30.6
79	6.0	0.0	27.5





Abundance Scan 860 (7.145 min): BF125915.D\data.ms (-810) (-)





Abundance Scan 855 (7.116 min): BF125917.D\data.ms (-849) #17

2-Methylphenol
Concen: 12.113 ng
RT: 7.110 min Scan# 854
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010





Tgt Ion:	107	Resp:	91151
Ion	Ratio	Lower	Upper
107	100		
108	118.9	90.6	135.8
77	52.1	35.7	53.5
79	50.0	34.0	51.0






Abundance Scan 901 (7.386 min): BF125917.D\data.ms (-896) #18

Hexachloroethane
Concen: 12.656 ng
RT: 7.386 min Scan# 901
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	117	Resp:	45833
Ion	Ratio	Lower	Upper
117	100		
119	89.3	73.4	110.2
201	91.1	64.9	97.3









n-Nitroso-di-n-propylamine
Concen: 13.998 ng
RT: 7.275 min Scan# 882
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	70	Resp:	82099
Ion	Ratio	Lower	Upper
70	100		
42	93.8	72.2	108.2
101	8.9	7.8	11.8
130	18.1	21.8	32.6#




3+4-Methylphenols
Concen: 13.218 ng
RT: 7.263 min Scan# 880
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	107	Resp:	122671
Ion	Ratio	Lower	Upper
107	100		
108	89.2	76.6	116.6
77	54.9	53.3	93.3
79	29.7	8.7	48.7






Abundance Scan 1032 (8.157 min): BF125917.D\data.ms (-1021)




Naphthalene-d8
Concen: 20.000 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	136	Resp:	553650
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.8	14.6
54	13.3	6.4	9.6#
68	6.4	3.3	4.9#







Abundance Scan 883 (7.281 min): BF125917.D\data.ms (-877) (-)



Acetophenone
Concen: 13.497 ng
RT: 7.275 min Scan# 882
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	105	Resp:	164264
Ion	Ratio	Lower	Upper
105	100		
71	7.8	5.5	8.3
51	43.5	26.0	39.0#
120	20.1	18.5	27.7









Nitrobenzene-d5
 Concen: 24.643 ng
 RT: 7.428 min Scan# 908
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	82	Resp:	219492
Ion	Ratio	Lower	Upper
82	100		
128	38.6	40.6	60.8#
54	73.8	49.6	74.4




Nitrobenzene
 Concen: 13.749 ng
 RT: 7.445 min Scan# 911
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	77	Resp:	118681
Ion	Ratio	Lower	Upper
77	100		
123	42.3	42.3	63.5
65	15.2	12.4	18.6





Abundance Scan 954 (7.698 min): BF125917.D\data.ms (-947) #25




Isophorone
Concen: 13.824 ng
RT: 7.686 min Scan# 952
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	82	Resp:	218139
Ion	Ratio	Lower	Upper
82	100		
95	7.3	6.5	9.7
138	14.9	18.4	27.6#



Abundance Scan 952 (7.686 min): BF125915.D\data.ms (-903) (-)





Abundance Scan 966 (7.769 min): BF125917.D\data.ms (-961) #26




2-Nitrophenol
Concen: 6.792 ng
RT: 7.763 min Scan# 965
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion:	139	Resp:	29647
Ion	Ratio	Lower	Upper
139	100		
109	34.2	26.0	39.0
65	75.2	31.0	46.4#



Abundance Scan 965 (7.763 min): BF125915.D\data.ms (-915) (-)






Abundance Scan 972 (7.804 min): BF125917.D\data.ms (-968) #27




2,4-Dimethylphenol
Concen: 12.413 ng
RT: 7.798 min Scan# 971
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	122	Resp:	90227
Ion	Ratio	Lower	Upper
122	100		
107	118.4	85.0	127.6
121	58.2	43.3	64.9






Abundance Scan 989 (7.904 min): BF125917.D\data.ms (-982) #28




bis(2-Chloroethoxy)methane
Concen: 13.249 ng
RT: 7.898 min Scan# 988
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	93	Resp:	145118
Ion	Ratio	Lower	Upper
93	100		
95	32.2	26.1	39.1
123	10.6	11.7	17.5#






Abundance Scan 1006 (8.004 min): BF125917.D\data.ms (-10629)




2,4-Dichlorophenol
Concen: 10.824 ng
RT: 7.998 min Scan# 1005
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	162	Resp:	84761
Ion	Ratio	Lower	Upper
162	100		
164	65.6	44.5	84.5
98	43.9	13.0	53.0






Abundance Scan 1022 (8.098 min): BF125917.D\data.ms (-10630)



1,2,4-Trichlorobenzene
Concen: 12.005 ng
RT: 8.092 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion:	180	Resp:	103085
Ion	Ratio	Lower	Upper
180	100		
182	100.4	76.8	115.2
145	29.9	26.7	40.1




Abundance Scan 1035 (8.175 min): BF125917.D\data.ms (-10291-)

Naphthalene
Concen: 12.099 ng
RT: 8.175 min Scan# 1035
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010






Tgt Ion:	128	Resp:	323892
Ion	Ratio	Lower	Upper
128	100		
129	10.8	9.4	14.0
127	13.3	10.9	16.3




Abundance Scan 990 (7.910 min): BF125917.D\data.ms (-9751-)


Benzoic acid
Concen: 7.474 ng
RT: 7.863 min Scan# 982
Delta R.T. -0.047 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	122	Resp:	38201
Ion	Ratio	Lower	Upper
122	100		
105	117.9	97.6	137.6
77	91.4	59.6	99.6






Abundance Scan 1043 (8.222 min): BF125917.D\data.ms (-1036)




4-Chloroaniline
Concen: 11.895 ng
RT: 8.216 min Scan# 1042
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.7	38.5
65	40.0	21.2	31.8#
92	20.0	14.0	21.0






Abundance Scan 1056 (8.298 min): BF125917.D\data.ms (-1054)



Hexachlorobutadiene
Concen: 13.318 ng
RT: 8.298 min Scan# 1056
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion	Ratio	Lower	Upper
225	100		
223	60.7	50.6	76.0
227	63.2	50.6	75.8






Abundance Scan 1106 (8.592 min): BF125917.D\data.ms (-11035)

Caprolactam
Concen: 11.911 ng
RT: 8.557 min Scan# 1100
Delta R.T. -0.035 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	113	Resp:	26949
Ion	Ratio	Lower	Upper
113	100		
55	260.6	154.1	194.1#
56	189.2	108.7	148.7#






Abundance Scan 1124 (8.698 min): BF125917.D\data.ms (-11136)


4-Chloro-3-methylphenol
Concen: 12.686 ng
RT: 8.686 min Scan# 1122
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	107	Resp:	97049
Ion	Ratio	Lower	Upper
107	100		
144	24.9	21.9	32.9
142	71.7	67.1	100.7





Abundance Scan 1153 (8.869 min): BF125917.D\data.ms (-11497)




2-Methylnaphthalene
Concen: 12.029 ng
RT: 8.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:142 Resp: 208546
Ion Ratio Lower Upper
142 100
141 79.7 71.2 106.8



Abundance Scan 1152 (8.863 min): BF125915.D\data.ms (-1102) (-)





Abundance Scan 1170 (8.969 min): BF125917.D\data.ms (-11698)




1-Methylnaphthalene
Concen: 12.037 ng
RT: 8.963 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion:142 Resp: 202499
Ion Ratio Lower Upper
142 100
141 85.6 76.0 114.0



Abundance Scan 1169 (8.963 min): BF125915.D\data.ms (-1119) (-)






Abundance Scan 1330 (9.910 min): BF125917.D\data.ms (-132499)




Acenaphthene-d10
Concen: 20.000 ng
RT: 9.904 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	164	Resp:	307287
Ion Ratio	Lower	Upper	
164	100		
162	93.8	81.3	121.9
160	43.9	38.0	57.0






Abundance Scan 1181 (9.033 min): BF125917.D\data.ms (-117510)



1,2,4,5-Tetrachlorobenzene
Concen: 11.684 ng
RT: 9.028 min Scan# 1180
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion:	216	Resp:	96359
Ion Ratio	Lower	Upper	
216	100		
214	80.0	64.9	97.3
179	21.1	17.4	26.2
108	19.2	16.5	24.7






Abundance Scan 1179 (9.022 min): BF125917.D\data.ms (-11741-)

Hexachlorocyclopentadiene
Concen: 11.581 ng
RT: 9.022 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	237	Resp:	46877
Ion Ratio	Lower	Upper	
237	100		
235	62.7	44.7	84.7
272	11.6	0.0	31.6






Abundance Scan 1463 (10.692 min): BF125917.D\data.ms (-1462-)


2,4,6-Tribromophenol
Concen: 18.252 ng
RT: 10.686 min Scan# 1462
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	330	Resp:	55374
Ion Ratio	Lower	Upper	
330	100		
332	103.8	76.2	114.4
141	37.5	44.6	66.8#





Abundance Scan 1199 (9.139 min): BF125917.D\data.ms (-1194)




2,4,6-Trichlorophenol
Concen: 10.310 ng
RT: 9.133 min Scan# 1198
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	196	Resp:	61629
Ion	Ratio	Lower	Upper
196	100		
198	102.8	74.9	112.3
200	32.1	27.0	40.4



Abundance Scan 1198 (9.133 min): BF125915.D\data.ms (-1148) (-)





Abundance Scan 1205 (9.175 min): BF125917.D\data.ms (-1204)




2,4,5-Trichlorophenol
Concen: 10.230 ng
RT: 9.169 min Scan# 1204
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion:	196	Resp:	65018
Ion	Ratio	Lower	Upper
196	100		
198	106.7	80.2	120.2
97	58.9	44.0	66.0
132	30.1	25.4	38.0



Abundance Scan 1204 (9.169 min): BF125915.D\data.ms (-1154) (-)






Abundance Scan 1215 (9.233 min): BF125917.D\data.ms (-12064)




2-Fluorobiphenyl
Concen: 24.737 ng
RT: 9.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	172	Resp:	440698
Ion	Ratio	Lower	Upper
172	100		
171	34.8	26.5	39.7
170	23.3	18.8	28.2






Abundance Scan 1232 (9.333 min): BF125917.D\data.ms (-12264)



1,1'-Biphenyl
Concen: 11.443 ng
RT: 9.328 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	154	Resp:	258526
Ion	Ratio	Lower	Upper
154	100		
153	38.1	20.0	60.0
76	12.1	0.0	30.3



Abundance Scan 1236 (9.357 min): BF125917.D\data.ms (-1228) (-)

2-Chloronaphthalene

Concen: 10.628 ng

RT: 9.351 min Scan# 1235

Delta R.T. -0.006 min

Lab File: BF125915.D

Acq: 25 Oct 2021 13:52

Tgt Ion:162 Resp: 194960

Ion Ratio Lower Upper

162 100

127 40.3 33.0 49.4

164 35.4 25.8 38.8



Instrument :

BNA_F


ClientSampleId :

SSTDIC010

Abundance Scan 1235 (9.351 min): BF125915.D\data.ms



Abundance Scan 1235 (9.351 min): BF125915.D\data.ms (-1185) (-)



Abundance Scan 1251 (9.445 min): BF125917.D\data.ms (-1248) (-)

2-Nitroaniline

Concen: 12.473 ng

RT: 9.439 min Scan# 1250

Delta R.T. -0.006 min

Lab File: BF125915.D

Acq: 25 Oct 2021 13:52

Tgt Ion: 65 Resp: 54930



Ion Ratio Lower Upper

65 100


92 54.4 55.5 83.3#

138 73.8 93.5 140.3#


Abundance Scan 1250 (9.439 min): BF125915.D\data.ms



Abundance Scan 1250 (9.439 min): BF125915.D\data.ms (-1200) (-)




Abundance Scan 1306 (9.769 min): BF125917.D\data.ms (-125917)





Acenaphthylene
Concen: 11.062 ng
RT: 9.763 min Scan# 1305
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	152	Resp:	297159
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.7	23.5
153	13.7	9.8	14.8





Abundance Scan 1283 (9.633 min): BF125917.D\data.ms (-125917)



Dimethylphthalate
Concen: 10.539 ng
RT: 9.622 min Scan# 1281
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion:	163	Resp:	213197
Ion	Ratio	Lower	Upper
163	100		
194	3.5	2.6	3.8
164	10.8	7.4	11.0






Abundance Scan 1292 (9.686 min): BF125917.D\data.ms (-12851-)

2,6-Dinitrotoluene
Concen: 8.794 ng
RT: 9.680 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion:	165	Resp:	35957
Ion	Ratio	Lower	Upper
165	100		
63	83.4	42.6	63.8#
89	66.5	40.6	60.8#






Abundance Scan 1336 (9.945 min): BF125917.D\data.ms (-13252-)

Acenaphthene
Concen: 11.182 ng
RT: 9.939 min Scan# 1335
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	154	Resp:	187671
Ion	Ratio	Lower	Upper
154	100		
153	111.1	89.4	134.2
152	48.4	40.5	60.7



Abundance Scan 1321 (9.857 min): BF125917.D\data.ms (-13143)

3-Nitroaniline

Concen: 8.924 ng

RT: 9.845 min Scan# 1319

Delta R.T. -0.012 min

Lab File: BF125915.D


Acq: 25 Oct 2021 13:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:138 Resp: 43143

Ion Ratio Lower Upper


138 100

108 15.3 9.3 13.9#

92 135.0 76.5 114.7#



Abundance Scan 1319 (9.845 min): BF125915.D\data.ms (-1270) (-)



Abundance Scan 1338 (9.957 min): BF125917.D\data.ms (-13354)

2,4-Dinitrophenol

Concen: 4.693 ng

RT: 9.945 min Scan# 1336

Delta R.T. -0.012 min

Lab File: BF125915.D

Acq: 25 Oct 2021 13:52




Tgt Ion:184 Resp: 8145

Ion Ratio Lower Upper

184 100

63 163.1 59.3 88.9#

154 862.4 68.6 102.8#




Abundance Scan 1336 (9.945 min): BF125915.D\data.ms (-1287) (-)






Abundance Scan 1365 (10.116 min): BF125917.D\data.ms (-1355) (-)

Dibenzofuran
Concen: 11.491 ng
RT: 10.110 min Scan# 1364
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampled :
SSTDIC010




Tgt Ion:	168	Resp:	288915
Ion	Ratio	Lower	Upper
168	100		
139	34.4	32.7	49.1
169	12.9	11.2	16.8







Abundance Scan 1346 (10.004 min): BF125917.D\data.ms (-1355) (-)

4-Nitrophenol
Concen: 9.277 ng
RT: 9.992 min Scan# 1344
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	139	Resp:	33433
Ion	Ratio	Lower	Upper
139	100		
109	78.5	60.4	100.4
65	121.6	58.4	98.4#









Scan 1360 (10.086 min): BF125917.D\data.ms (-1357) (-)
 2,4-Dinitrotoluene
 Concen: 8.094 ng
 RT: 10.080 min Scan# 1359
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	165	Resp:	41847
Ion	Ratio	Lower	Upper
165	100		
63	57.8	40.9	61.3
89	77.8	59.7	89.5
182	2.3	0.0	0.0#



Scan 1423 (10.457 min): BF125917.D\data.ms (-1458) (-)
 Fluorene
 Concen: 11.587 ng
 RT: 10.451 min Scan# 1422
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52


Tgt Ion:	166	Resp:	213361
Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.1	117.1
167	13.1	11.1	16.7






Abundance Scan 1384 (10.227 min): BF125917.D\data.ms (-1359) (-)

2,3,4,6-Tetrachlorophenol
Concen: 10.880 ng
RT: 10.222 min Scan# 1383
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion	Ratio	Lower	Upper
232	100		
131	45.9	45.4	68.0
130	2.5	0.0	0.0#
166	32.3	28.5	42.7






Abundance Scan 1402 (10.333 min): BF125917.D\data.ms (-1359) (-)

Diethylphthalate
Concen: 11.240 ng
RT: 10.327 min Scan# 1401
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion	Ratio	Lower	Upper
149	100		
177	21.8	16.5	24.7
150	11.7	10.1	15.1





Abundance Scan 1422 (10.451 min): BF125917.D\data.ms (-1371) (-)

4-Chlorophenyl-phenylether
Concen: 11.701 ng
RT: 10.445 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Instrument :
BNA_F
ClientSampleId :
SSTDIC010



Tgt Ion:	204	Resp:	103786
Ion Ratio	Lower	Upper	
204	100		
206	31.9	27.6	41.4
141	59.6	58.6	88.0




Abundance Scan 1421 (10.445 min): BF125915.D\data.ms (-1371) (-)





Abundance Scan 1425 (10.469 min): BF125917.D\data.ms (-1372) (-)


4-Nitroaniline
Concen: 8.676 ng
RT: 10.451 min Scan# 1422
Delta R.T. -0.018 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Tgt Ion:	138	Resp:	38762
Ion Ratio	Lower	Upper	
138	100		
92	58.3	29.6	69.6
108	90.9	78.1	118.1



Abundance Scan 1422 (10.451 min): BF125915.D\data.ms (-1374) (-)









Azobenzene
Concen: 13.758 ng
RT: 10.604 min Scan# 1448
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	77	Resp:	249606
Ion	Ratio	Lower	Upper
77	100		
182	21.5	6.3	46.3
105	23.4	2.9	42.9
51	42.4	24.1	64.1



Phenanthrene-d10
Concen: 20.000 ng
RT: 11.392 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	188	Resp:	550830
Ion	Ratio	Lower	Upper
188	100		
94	8.5	6.7	10.1
80	8.8	7.5	11.3









Scan 1430 (10.498 min): BF125917.D\data.ms (-1430)
 4,6-Dinitro-2-methylphenol
 Concen: 5.000 ng
 RT: 10.486 min Scan# 1428
 Delta R.T. -0.012 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion:	198	Resp:	13737
Ion Ratio	Lower	Upper	
198	100		
51	70.3	45.3	85.3
105	46.7	30.6	70.6



Scan 1442 (10.569 min): BF125917.D\data.ms (-1442)
 n-Nitrosodiphenylamine
 Concen: 9.970 ng
 RT: 10.557 min Scan# 1440
 Delta R.T. -0.012 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52



Tgt Ion:	169	Resp:	187346
Ion Ratio	Lower	Upper	
169	100		
168	63.9	50.8	76.2
167	36.6	28.8	43.2





Abundance Scan 1505 (10.939 min): BF125917.D\data.ms (-1507) (-)

4-Bromophenyl-phenylether
Concen: 10.458 ng
RT: 10.933 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010





Tgt Ion	Ratio	Lower	Upper
248	100		
250	98.8	74.2	111.4
141	77.7	60.4	90.6






Abundance Scan 1516 (11.004 min): BF125917.D\data.ms (-1508) (-)

Hexachlorobenzene
Concen: 9.922 ng
RT: 10.998 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion	Ratio	Lower	Upper
284	100		
142	40.5	36.4	54.6
249	32.7	25.6	38.4









Atrazine
Concen: 10.417 ng
RT: 11.086 min Scan# 1530
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	200	Resp:	56813
Ion	Ratio	Lower	Upper
200	100		
173	25.2	11.1	51.1
215	49.3	22.7	62.7



Pentachlorophenol
Concen: 9.398 ng
RT: 11.186 min Scan# 1547
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	266	Resp:	34730
Ion	Ratio	Lower	Upper
266	100		
268	65.8	48.9	73.3
264	61.9	55.9	83.9









Phenanthrene
Concen: 11.152 ng
RT: 11.410 min Scan# 1585
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:	178	Resp:	328671
Ion	Ratio	Lower	Upper
178	100		
176	19.7	14.7	22.1
179	15.8	11.5	17.3



Anthracene
Concen: 11.102 ng
RT: 11.463 min Scan# 1594
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	178	Resp:	326806
Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.3	21.5
179	15.2	12.2	18.4









Carbazole
Concen: 11.484 ng
RT: 11.616 min Scan# 1620
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	167	Resp:	314598
Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.4	26.0
139	12.2	11.0	16.4



Di-n-butylphthalate
Concen: 10.868 ng
RT: 11.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Tgt Ion:	149	Resp:	328686
Ion	Ratio	Lower	Upper
149	100		
150	8.7	7.4	11.2
104	5.0	4.1	6.1






Abundance Scan 1788 (12.604 min): BF125917.D\data.ms (-1785) (-)

Fluoranthene
Concen: 12.690 ng
RT: 12.598 min Scan# 1787
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010







Tgt Ion:	202	Resp:	337455
Ion	Ratio	Lower	Upper
202	100		
101	11.1	0.0	30.7
203	17.2	0.0	36.3




Abundance Scan 2031 (14.033 min): BF125917.D\data.ms (-2025) (-)

Chrysene-d12
Concen: 20.000 ng
RT: 14.027 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	240	Resp:	368025
Ion	Ratio	Lower	Upper
240	100		
120	10.3	10.8	16.2#
236	25.7	21.6	32.4









Scan 1808 (12.721 min): BF125917.D\data.ms (-1807) (-)
 Benzidine
 Concen: 9.195 ng
 RT: 12.721 min Scan# 1808
 Delta R.T. 0.000 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Tgt Ion	Ratio	Lower	Upper
184	100		
185	14.0	10.8	16.2
183	9.8	8.1	12.1



Scan 1827 (12.833 min): BF125917.D\data.ms (-1826) (-)
 Pyrene
 Concen: 9.307 ng
 RT: 12.827 min Scan# 1826
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52


Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.2	15.3	22.9
203	17.1	13.4	20.0






Abundance Scan 1852 (12.980 min): BF125917.D\data.ms (-1879) (-)

Terphenyl-d14
Concen: 19.572 ng
RT: 12.974 min Scan# 1851
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.9	8.9
122	9.9	9.0	13.4







Abundance Scan 1933 (13.457 min): BF125917.D\data.ms (-1880) (-)

Butylbenzylphthalate
Concen: 8.645 ng
RT: 13.457 min Scan# 1933
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion	Ratio	Lower	Upper
149	100		
91	80.3	51.8	77.8#
206	21.0	14.5	21.7









#181 (-)
 Benzo(a)anthracene
 Concen: 11.340 ng
 RT: 14.015 min Scan# 2028
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	228	Resp:	270768
Ion Ratio	Lower	Upper	
228	100		
226	26.9	21.6	32.4
229	19.6	15.6	23.4



#182 (-)
 3,3'-Dichlorobenzidine
 Concen: 10.806 ng
 RT: 13.980 min Scan# 2022
 Delta R.T. 0.000 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	252	Resp:	81523
Ion Ratio	Lower	Upper	
252	100		
254	61.8	51.4	77.0
126	9.4	9.5	14.3#









Chrysene
 Concen: 11.409 ng
 RT: 14.051 min Scan# 2034
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:	228	Resp:	269249
Ion Ratio	Lower	Upper	
228	100		
226	30.5	22.6	34.0
229	19.9	15.9	23.9



Bis(2-ethylhexyl)phthalate
 Concen: 9.277 ng
 RT: 14.021 min Scan# 2029
 Delta R.T. 0.000 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52


Tgt Ion:	149	Resp:	120974
Ion Ratio	Lower	Upper	
149	100		
167	29.0	21.4	32.0
279	4.1	2.6	4.0#








#85 (-)
Di-n-octyl phthalate
Concen: 7.624 ng
RT: 14.633 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52


Instrument :
BNA_F
ClientSampleId :
SSTDIC010





Tgt Ion:	149	Resp:	165642
Ion Ratio		Lower	Upper
149	100		
167	1.5	1.0	1.6
43	16.7	9.9	14.9#




#86 (-)
Perylene-d12
Concen: 20.000 ng
RT: 15.492 min Scan# 2279
Delta R.T. -0.006 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52



Tgt Ion:	264	Resp:	263476
Ion Ratio		Lower	Upper
264	100		
260	24.0	19.6	29.4
265	20.7	17.2	25.8






Abundance Scan 2529 (16.962 min): BF125917.D\data.ms (-2518) (-)




Indeno(1,2,3-cd)pyrene
Concen: 8.251 ng
RT: 16.951 min Scan# 2527
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.3	24.1	36.1
277	24.0	18.8	28.2







Abundance Scan 2207 (15.068 min): BF125917.D\data.ms (-2208) (-)



Benzo(b)fluoranthene
Concen: 12.948 ng
RT: 15.092 min Scan# 2211
Delta R.T. 0.024 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.4	16.8	25.2
125	10.5	9.8	14.8









Benzo(k)fluoranthene
 Concen: 12.611 ng
 RT: 15.092 min Scan# 2211
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52




Instrument :
 BNA_F
 ClientSampleId :
 SSTDIC010

Tgt Ion:	252	Resp:	192598
Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.8	26.8
125	10.5	7.2	10.8




Benzo(a)pyrene
 Concen: 11.140 ng
 RT: 15.427 min Scan# 2268
 Delta R.T. -0.006 min
 Lab File: BF125915.D
 Acq: 25 Oct 2021 13:52

Tgt Ion:	252	Resp:	146751
Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.4	27.6
125	10.6	9.9	14.9






Abundance Scan 2533 (16.986 min): BF125917.D\data.ms (-2522) (-)




Dibenzo(a,h)anthracene
Concen: 8.318 ng
RT: 16.974 min Scan# 2531
Delta R.T. -0.012 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	278	Resp:	137972
Ion Ratio	Lower	Upper	
278	100		
139	15.4	14.9	22.3
279	21.2	19.7	29.5



Abundance Scan 2605 (17.409 min): BF125917.D\data.ms (-2592) (-)



Benzo(g,h,i)perylene
Concen: 8.224 ng
RT: 17.392 min Scan# 2602
Delta R.T. -0.018 min
Lab File: BF125915.D
Acq: 25 Oct 2021 13:52

Tgt Ion:	276	Resp:	139319
Ion Ratio	Lower	Upper	
276	100		
277	22.6	22.0	33.0
138	20.3	22.2	33.4#

