



Benzoic acid



$\text{Response} = 2.301\text{e-}001 * \text{Amt} - 4.797\text{e-}002$
 Coef of Det (r^2) = 0.995507 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102521.M
 Calibration Table Last Updated: Mon Oct 25 18:49:16 2021


2,4-Dinitrophenol



$\text{Response} = 1.125\text{e-}001 * \text{Amt} - 4.007\text{e-}002$
 Coef of Det (r^2) = 0.997426 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102521.M
 Calibration Table Last Updated: Mon Oct 25 18:49:16 2021


Fluorene

Response Ratio



$\text{Response} = 9.180\text{e-}001 * \text{Amt} + 2.300\text{e-}001$
 Coef of Det (r^2) = 0.997604 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102521.M
 Calibration Table Last Updated: Mon Oct 25 18:49:16 2021

4,6-Dinitro-2-methylphenol



Response = $1.005e-001 \cdot \text{Amt} - 3.049e-002$
 Coef of Det (r^2) = 0.998043 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102521.M
 Calibration Table Last Updated: Mon Oct 25 18:49:16 2021