

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF102620\  
 Data File : BF122119.D  
 Acq On : 26 Oct 2020 17:08  
 Operator : JU/CG  
 Sample : L4494-02  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 AB-EN

Quant Time: Oct 26 18:14:13 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF101220.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Oct 26 17:27:32 2020  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc      | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|-----------|--------|----------|
| -----                         |        |      |          |           |        |          |
| Internal Standards            |        |      |          |           |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 6.710  | 152  | 101044   | 20.00     | ng     | 0.00     |
| 21) Naphthalene-d8            | 7.998  | 136  | 392477   | 20.00     | ng     | 0.00     |
| 39) Acenaphthene-d10          | 9.745  | 164  | 205099   | 20.00     | ng     | 0.00     |
| 64) Phenanthrene-d10          | 11.233 | 188  | 397929   | 20.00     | ng     | 0.00     |
| 76) Chrysene-d12              | 13.874 | 240  | 294146   | 20.00     | ng     | 0.00     |
| 86) Perylene-d12              | 15.304 | 264  | 240489   | 20.00     | ng     | 0.00     |
| System Monitoring Compounds   |        |      |          |           |        |          |
| 5) 2-Fluorophenol             | 5.334  | 112  | 636178   | 92.92     | ng     | 0.00     |
| 7) Phenol-d6                  | 6.363  | 99   | 802154   | 91.02     | ng     | 0.00     |
| 23) Nitrobenzene-d5           | 7.281  | 82   | 545111   | 71.23     | ng     | 0.00     |
| 42) 2,4,6-Tribromophenol      | 10.539 | 330  | 196807   | 77.57     | ng     | 0.00     |
| 45) 2-Fluorobiphenyl          | 9.069  | 172  | 949554   | 68.12     | ng     | 0.00     |
| 79) Terphenyl-d14             | 12.816 | 244  | 1059896  | 68.67     | ng     | 0.00     |
| Target Compounds              |        |      |          |           |        |          |
|                               |        |      |          |           | Qvalue |          |
| 9) Benzaldehyde               | 6.175  | 77   | 111      | Below Cal | #      | 3        |
| 19) n-Nitroso-di-n-propyla... | 7.281  | 70   | 73684    | 14.68     | ng     | # 80     |
| 25) Isophorone                | 7.281  | 82   | 545106   | 37.67     | ng     | # 67     |
| 50) Dimethylphthalate         | 9.469  | 163  | 146414   | 9.76      | ng     | 99       |
| 52) Acenaphthene              | 9.781  | 154  | 32184    | 2.72      | ng     | 98       |
| 55) Dibenzofuran              | 9.957  | 168  | 44277    | 2.56      | ng     | 96       |
| 56) 4-Nitrophenol             | 9.951  | 139  | 16814    | 8.26      | ng     | # 12     |
| 57) 2,4-Dinitrotoluene        | 9.745  | 165  | 28230    | 6.97      | ng     | # 21     |
| 58) Fluorene                  | 10.298 | 166  | 39641    | 2.84      | ng     | 98       |
| 65) 4,6-Dinitro-2-methylph... | 10.539 | 198  | 265      | 4.52      | ng     | # 1      |
| 67) 4-Bromophenyl-phenylether | 10.539 | 248  | 11339    | 2.47      | ng     | # 1      |
| 71) Phenanthrene              | 11.257 | 178  | 614258   | 28.15     | ng     | 99       |
| 72) Anthracene                | 11.310 | 178  | 158071   | 6.99      | ng     | 99       |
| 73) Carbazole                 | 11.475 | 167  | 49042    | 2.44      | ng     | 98       |
| 75) Fluoranthene              | 12.445 | 202  | 759816   | 32.48     | ng     | 99       |
| 78) Pyrene                    | 12.675 | 202  | 694655   | 30.99     | ng     | 100      |
| 81) Benzo(a)anthracene        | 13.863 | 228  | 284852   | 14.78     | ng     | 99       |
| 83) Chrysene                  | 13.904 | 228  | 230636   | 12.19     | ng     | 97       |
| 87) Indeno(1,2,3-cd)pyrene    | 16.710 | 276  | 117135   | 7.50      | ng     | 95       |
| 88) Benzo(b)fluoranthene      | 14.898 | 252  | 330164   | 20.31     | ng     | 98       |
| 89) Benzo(k)fluoranthene      | 14.898 | 252  | 330164   | 21.38     | ng     | 98       |
| 90) Benzo(a)pyrene            | 15.245 | 252  | 184431   | 13.14     | ng     | 98       |
| 92) Benzo(g,h,i)perylene      | 17.133 | 276  | 122759   | 9.54      | ng     | 98       |
| -----                         |        |      |          |           |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF102620\  
Data File : BF122119.D  
Acq On : 26 Oct 2020 17:08  
Operator : JU/CG  
Sample : L4494-02  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN

Quant Time: Oct 26 18:14:13 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF101220.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Oct 26 17:27:32 2020  
Response via : Initial Calibration









#1  
1,4-Dichlorobenzene-d4  
Concen: 20.00 ng  
RT: 6.710 min Scan# 786  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion:152 Resp: 101044  
Ion Ratio Lower Upper  
152 100  
150 157.3 125.3 187.9  
115 63.9 51.9 77.9



#5  
2-Fluorophenol  
Concen: 92.92 ng  
RT: 5.334 min Scan# 552  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:112 Resp: 636178  
Ion Ratio Lower Upper  
112 100  
64 66.4 50.4 75.6  
63 33.9 25.8 38.6









#7  
Phenol-d6  
Concen: 91.02 ng  
RT: 6.363 min Scan# 727  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion: 99 Resp: 802154  
Ion Ratio Lower Upper  
99 100  
42 21.4 15.9 23.9  
71 36.7 27.1 40.7



#19  
n-Nitroso-di-n-propylamine  
Concen: 14.68 ng  
RT: 7.281 min Scan# 883  
Delta R.T. 0.141 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion: 70 Resp: 73684  
Ion Ratio Lower Upper  
70 100  
42 52.5 50.2 75.4  
101 0.0 7.5 11.3#  
130 1.9 15.5 23.3#









#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 7.998 min Scan# 1005  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN

Tgt Ion:136 Resp: 392477




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 136 | 100   |       |       |
| 137 | 10.6  | 8.6   | 13.0  |
| 54  | 8.5   | 7.1   | 10.7  |
| 68  | 6.0   | 5.0   | 7.4   |




#23  
Nitrobenzene-d5  
Concen: 71.23 ng  
RT: 7.281 min Scan# 883  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion: 82 Resp: 545111

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 82  | 100   |       |       |
| 128 | 40.7  | 35.0  | 52.4  |
| 54  | 55.6  | 43.9  | 65.9  |









#25  
Isophorone  
Concen: 37.67 ng  
RT: 7.281 min Scan# 883  
Delta R.T. -0.265 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion: 82 Resp: 545106  
Ion Ratio Lower Upper  
82 100  
95 0.0 5.4 8.0#  
138 0.0 13.4 20.0#



#39  
Acenaphthene-d10  
Concen: 20.00 ng  
RT: 9.745 min Scan# 1302  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:164 Resp: 205099  
Ion Ratio Lower Upper  
164 100  
162 103.6 83.2 124.8  
160 46.4 37.5 56.3









#42  
2,4,6-Tribromophenol  
Concen: 77.57 ng  
RT: 10.539 min Scan# 1437  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion: 330 Resp: 196807  
Ion Ratio Lower Upper  
330 100  
332 96.4 77.3 115.9  
141 30.0 22.3 33.5



#45  
2-Fluorobiphenyl  
Concen: 68.12 ng  
RT: 9.069 min Scan# 1187  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion: 172 Resp: 949554  
Ion Ratio Lower Upper  
172 100  
171 36.0 29.1 43.7  
170 23.9 19.8 29.6









#50  
Dimethylphthalate  
Concen: 9.76 ng  
RT: 9.469 min Scan# 1255  
Delta R.T. -0.012 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN

Tgt Ion:163 Resp: 146414  
Ion Ratio Lower Upper  
163 100  
194 3.6 3.3 4.9  
164 9.9 8.2 12.4




#52  
Acenaphthene  
Concen: 2.72 ng  
RT: 9.781 min Scan# 1308  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:154 Resp: 32184  
Ion Ratio Lower Upper  
154 100  
153 114.5 89.8 134.8  
152 54.3 43.2 64.8







#55

Dibenzenofuran

Concen: 2.56 ng

RT: 9.957 min Scan# 1338

Delta R.T. 0.000 min

Lab File: BF122119.D

Acq: 26 Oct 2020 17:08





Tgt Ion:168 Resp: 44277

Ion Ratio Lower Upper

168 100

139 35.9 31.0 46.6

169 14.1 11.3 16.9



#56

4-Nitrophenol

Concen: 8.26 ng

RT: 9.951 min Scan# 1337

Delta R.T. 0.041 min

Lab File: BF122119.D

Acq: 26 Oct 2020 17:08




Tgt Ion:139 Resp: 16814


Ion Ratio Lower Upper

139 100

109 0.0 45.4 85.4#

65 2.2 68.4 108.4#









#57  
2,4-Dinitrotoluene  
Concen: 6.97 ng  
RT: 9.745 min Scan# 1302  
Delta R.T. -0.218 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


|             |       |            |
|-------------|-------|------------|
| Tgt Ion:165 | Resp: | 28230      |
| Ion Ratio   | Lower | Upper      |
| 165         | 100   |            |
| 63          | 1.4   | 41.3 61.9# |
| 89          | 2.1   | 61.8 92.8# |
| 182         | 0.0   | 2.1 3.1#   |



#58  
Fluorene  
Concen: 2.84 ng  
RT: 10.298 min Scan# 1396  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

|             |       |            |
|-------------|-------|------------|
| Tgt Ion:166 | Resp: | 39641      |
| Ion Ratio   | Lower | Upper      |
| 166         | 100   |            |
| 165         | 96.4  | 79.1 118.7 |
| 167         | 14.1  | 11.0 16.6  |









#64  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 11.233 min Scan# 1555  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion:188 Resp: 397929  
Ion Ratio Lower Upper  
188 100  
94 9.9 8.3 12.5  
80 10.8 9.1 13.7



#65  
4,6-Dinitro-2-methylphenol  
Concen: 4.52 ng  
RT: 10.539 min Scan# 1437  
Delta R.T. 0.165 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:198 Resp: 265  
Ion Ratio Lower Upper  
198 100  
51 369.7 29.6 69.6#  
105 238.6 24.4 64.4#









#67  
4-Bromophenyl-phenylether  
Concen: 2.47 ng  
RT: 10.539 min Scan# 1437  
Delta R.T. -0.235 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion:248 Resp: 11339  
Ion Ratio Lower Upper  
248 100  
250 197.8 77.0 115.6#  
141 478.5 53.7 80.5#



#71  
Phenanthrene  
Concen: 28.15 ng  
RT: 11.257 min Scan# 1559  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:178 Resp: 614258  
Ion Ratio Lower Upper  
178 100  
176 19.6 15.9 23.9  
179 15.4 12.2 18.4









#72  
 Anthracene  
 Concen: 6.99 ng  
 RT: 11.310 min Scan# 1568  
 Delta R.T. -0.006 min  
 Lab File: BF122119.D  
 Acq: 26 Oct 2020 17:08




Instrument :  
 BNA\_F  
 ClientSampled :  
 AB-EN


Tgt Ion:178 Resp: 158071  
 Ion Ratio Lower Upper  
 178 100  
 176 18.8 15.7 23.5  
 179 15.6 12.7 19.1



#73  
 Carbazole  
 Concen: 2.44 ng  
 RT: 11.475 min Scan# 1596  
 Delta R.T. 0.000 min  
 Lab File: BF122119.D  
 Acq: 26 Oct 2020 17:08


Tgt Ion:167 Resp: 49042  
 Ion Ratio Lower Upper  
 167 100  
 166 20.9 17.8 26.6  
 139 12.9 10.8 16.2








#75  
Fluoranthene  
Concen: 32.48 ng  
RT: 12.445 min Scan# 1761  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08


Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN





Tgt Ion:202 Resp: 759816  
Ion Ratio Lower Upper  
202 100  
101 12.7 0.0 33.4  
203 17.9 0.0 37.9




#76  
Chrysene-d12  
Concen: 20.00 ng  
RT: 13.874 min Scan# 2004  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08



Tgt Ion:240 Resp: 294146  
Ion Ratio Lower Upper  
240 100  
120 11.1 9.2 13.8  
236 25.9 20.9 31.3









#78  
Pyrene  
Concen: 30.99 ng  
RT: 12.675 min Scan# 1800  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN


Tgt Ion:202 Resp: 694655  
Ion Ratio Lower Upper  
202 100  
200 21.7 17.4 26.2  
203 18.2 14.6 21.8



#79  
Terphenyl-d14  
Concen: 68.67 ng  
RT: 12.816 min Scan# 1824  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:244 Resp: 1059896  
Ion Ratio Lower Upper  
244 100  
212 7.4 5.4 8.2  
122 11.3 8.1 12.1









#81  
Benzo(a)anthracene  
Concen: 14.78 ng  
RT: 13.863 min Scan# 2002  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08




Instrument :  
BNA\_F  
ClientSampled :  
AB-EN

Tgt Ion:228 Resp: 284852  
Ion Ratio Lower Upper  
228 100  
226 27.7 22.3 33.5  
229 19.9 16.2 24.4




#83  
Chrysene  
Concen: 12.19 ng  
RT: 13.904 min Scan# 2009  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:228 Resp: 230636  
Ion Ratio Lower Upper  
228 100  
226 30.4 24.6 37.0  
229 22.8 15.9 23.9











#86  
Perylene-d12  
Concen: 20.00 ng  
RT: 15.304 min Scan# 2247  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN







Tgt Ion:264 Resp: 240489  
Ion Ratio Lower Upper  
264 100  
260 24.1 18.7 28.1  
265 21.6 17.3 25.9



#87  
Indeno(1,2,3-cd)pyrene  
Concen: 7.50 ng  
RT: 16.710 min Scan# 2486  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:276 Resp: 117135  
Ion Ratio Lower Upper  
276 100  
138 20.9 19.9 29.9  
277 24.2 20.4 30.6





#88

Benzo(b)fluoranthene

Concen: 20.31 ng

RT: 14.898 min Scan# 2178

Delta R.T. 0.000 min

Lab File: BF122119.D

Acq: 26 Oct 2020 17:08

Tgt Ion:252 Resp: 330164

Ion Ratio Lower Upper

252 100

253 22.0 18.2 27.2


125 11.3 8.0 12.0

Instrument :




BNA\_F

ClientSampleId :

AB-EN



Abundance



#89

Benzo(k)fluoranthene

Concen: 21.38 ng

RT: 14.898 min Scan# 2178

Delta R.T. -0.029 min

Lab File: BF122119.D

Acq: 26 Oct 2020 17:08


Tgt Ion:252 Resp: 330164

Ion Ratio Lower Upper



252 100


253 22.0 17.8 26.6

125 11.3 7.7 11.5



Abundance









#90  
Benzo(a)pyrene  
Concen: 13.14 ng  
RT: 15.245 min Scan# 2237  
Delta R.T. -0.006 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Instrument :  
BNA\_F  
ClientSampleId :  
AB-EN

Tgt Ion:252 Resp: 184431  
Ion Ratio Lower Upper  
252 100  
253 22.4 17.3 25.9  
125 11.8 8.6 13.0



#92  
Benzo(g,h,i)perylene  
Concen: 9.54 ng  
RT: 17.133 min Scan# 2558  
Delta R.T. 0.000 min  
Lab File: BF122119.D  
Acq: 26 Oct 2020 17:08

Tgt Ion:276 Resp: 122759  
Ion Ratio Lower Upper  
276 100  
277 24.5 19.4 29.2  
138 20.0 17.8 26.6

