

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102822\
 Data File : BF130848.D
 Acq On : 29 Oct 2022 03:57
 Operator : CG\JU
 Sample : N5290-02
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 257248-8

Quant Time: Oct 29 05:49:13 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 29 05:41:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.945	152	97226	20.000	ng	0.00
21) Naphthalene-d8	8.228	136	379788	20.000	ng	0.00
39) Acenaphthene-d10	9.986	164	205899	20.000	ng	-0.01
64) Phenanthrene-d10	11.480	188	369856	20.000	ng	0.00
76) Chrysene-d12	14.127	240	232184	20.000	ng	0.00
86) Perylene-d12	15.639	264	172425	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.557	112	337980	60.264	ng	0.00
7) Phenol-d6	6.551	99	259470	35.603	ng	-0.01
23) Nitrobenzene-d5	7.510	82	663203	105.297	ng	0.00
42) 2,4,6-Tribromophenol	10.780	330	257030	148.293	ng	0.00
45) 2-Fluorobiphenyl	9.310	172	1201715	88.240	ng	0.00
79) Terphenyl-d14	13.074	244	1100409	86.847	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102822\
Data File : BF130848.D
Acq On : 29 Oct 2022 03:57
Operator : CG\JU
Sample : N5290-02
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
257248-8

Quant Time: Oct 29 05:49:13 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 05:41:34 2022
Response via : Initial Calibration

