

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102822\
 Data File : BF130816.D
 Acq On : 28 Oct 2022 09:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 16:47:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 28 04:00:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.01
2	1,4-Dioxane	40.000	41.108	-2.8	98	0.00
3	Pyridine	40.000	42.336	-5.8	102	0.00
4	n-Nitrosodimethylamine	40.000	43.040	-7.6	104	0.00
5 S	2-Fluorophenol	80.000	78.906	1.4	99	0.00
6	Aniline	40.000	41.070	-2.7	100	0.01
7 S	Phenol-d6	80.000	79.548	0.6	100	0.00
8	2-Chlorophenol	40.000	40.467	-1.2	99	0.00
9	Benzaldehyde	40.000	38.256	4.4	105	0.01
10 C	Phenol	40.000	40.554	-1.4	100	0.00
11	bis(2-Chloroethyl)ether	40.000	40.881	-2.2	100	0.01
12	1,3-Dichlorobenzene	40.000	40.256	-0.6	99	0.01
13 C	1,4-Dichlorobenzene	40.000	40.400	-1.0	99	0.00
14	1,2-Dichlorobenzene	40.000	40.715	-1.8	100	0.00
15	Benzyl Alcohol	40.000	40.368	-0.9	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.824	-2.1	100	0.00
17	2-Methylphenol	40.000	40.670	-1.7	99	0.00
18	Hexachloroethane	40.000	41.078	-2.7	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.529	1.2	100	0.00
20	3+4-Methylphenols	40.000	40.522	-1.3	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	41.116	-2.8	101	0.01
23 S	Nitrobenzene-d5	80.000	93.731	-17.2	107	0.00
24	Nitrobenzene	40.000	44.705	-11.8	104	0.00
25	Isophorone	40.000	41.306	-3.3	101	0.01
26 C	2-Nitrophenol	40.000	42.235	-5.6	109	0.01
27	2,4-Dimethylphenol	40.000	41.203	-3.0	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.826	-2.1	100	0.00
29 C	2,4-Dichlorophenol	40.000	41.673	-4.2	100	0.01
30	1,2,4-Trichlorobenzene	40.000	40.951	-2.4	100	0.00
31	Naphthalene	40.000	40.661	-1.7	100	0.00
32	Benzoic acid	40.000	42.040	-5.1	112	0.00
33	4-Chloroaniline	40.000	41.350	-3.4	100	0.00
34 C	Hexachlorobutadiene	40.000	41.241	-3.1	101	0.00
35	Caprolactam	40.000	43.171	-7.9	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.917	-4.8	100	0.00
37	2-Methylnaphthalene	40.000	40.582	-1.5	100	0.01
38	1-Methylnaphthalene	40.000	40.454	-1.1	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.725	-1.8	103	0.01
41 P	Hexachlorocyclopentadiene	40.000	42.501	-6.3	101	0.00
42 S	2,4,6-Tribromophenol	80.000	89.800	-12.2	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.679	-1.7	98	0.00
44	2,4,5-Trichlorophenol	40.000	43.248	-8.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	77.100	3.6	102	0.01
46	1,1'-Biphenyl	40.000	40.299	-0.7	101	0.00
47	2-Chloronaphthalene	40.000	40.231	-0.6	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102822\
 Data File : BF130816.D
 Acq On : 28 Oct 2022 09:10
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 28 16:47:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 28 04:00:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.265	-10.7	115	0.00
49	Acenaphthylene	40.000	40.902	-2.3	101	0.01
50	Dimethylphthalate	40.000	41.129	-2.8	103	0.00
51	2,6-Dinitrotoluene	40.000	44.459	-11.1	112	0.00
52 C	Acenaphthene	40.000	40.985	-2.5	102	0.00
53	3-Nitroaniline	40.000	48.767	-21.9	112	0.01
54 P	2,4-Dinitrophenol	40.000	43.824	-9.6	116	0.00
55	Dibenzofuran	40.000	40.975	-2.4	103	0.00
56 P	4-Nitrophenol	40.000	48.410	-21.0	108	0.00
57	2,4-Dinitrotoluene	40.000	45.230	-13.1	115	0.00
58	Fluorene	40.000	40.233	-0.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.096	-5.2	102	0.01
60	Diethylphthalate	40.000	42.312	-5.8	105	0.00
61	4-Chlorophenyl-phenylether	40.000	40.283	-0.7	103	0.01
62	4-Nitroaniline	40.000	48.623	-21.6	109	0.00
63	Azobenzene	40.000	41.257	-3.1	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.076	-5.2	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.629	-1.6	102	0.00
67	4-Bromophenyl-phenylether	40.000	42.063	-5.2	104	0.01
68	Hexachlorobenzene	40.000	41.044	-2.6	104	0.00
69	Atrazine	40.000	42.817	-7.0	104	0.00
70 C	Pentachlorophenol	40.000	45.861	-14.7	108	0.00
71	Phenanthrene	40.000	40.582	-1.5	102	0.01
72	Anthracene	40.000	40.679	-1.7	102	0.01
73	Carbazole	40.000	40.712	-1.8	103	0.00
74	Di-n-butylphthalate	40.000	41.042	-2.6	103	0.00
75 C	Fluoranthene	40.000	41.618	-4.0	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	113	0.00
77	Benzydine	40.000	42.423	-6.1	116	0.00
78	Pyrene	40.000	40.102	-0.3	108	0.00
79 S	Terphenyl-d14	80.000	79.439	0.7	110	0.01
80	Butylbenzylphthalate	40.000	43.256	-8.1	113	0.00
81	Benzo(a)anthracene	40.000	40.508	-1.3	112	0.01
82	3,3'-Dichlorobenzidine	40.000	41.925	-4.8	116	0.01
83	Chrysene	40.000	40.487	-1.2	114	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	44.566	-11.4	120	0.01
85 c	Di-n-octyl phthalate	40.000	44.741	-11.9	120	0.01
86 I	Perylene-d12	20.000	20.000	0.0	109	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	43.394	-8.5	108	0.02
88	Benzo(b)fluoranthene	40.000	41.138	-2.8	113	0.01
89	Benzo(k)fluoranthene	40.000	42.473	-6.2	113	0.01
90 C	Benzo(a)pyrene	40.000	41.547	-3.9	109	0.01
91	Dibenzo(a,h)anthracene	40.000	43.302	-8.3	106	0.02
92	Benzo(g,h,i)perylene	40.000	43.811	-9.5	106	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102822\
Data File : BF130816.D
Acq On : 28 Oct 2022 09:10
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 28 16:47:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 28 04:00:21 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0