

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102922\
 Data File : BF130852.D
 Acq On : 29 Oct 2022 12:39
 Operator : CG\JU
 Sample : PB148592BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB148592BL

Quant Time: Oct 30 23:22:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 29 05:41:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.939	152	94738	20.000	ng	0.00
21) Naphthalene-d8	8.228	136	364593	20.000	ng	# 0.00
39) Acenaphthene-d10	9.986	164	200862	20.000	ng	-0.01
64) Phenanthrene-d10	11.480	188	365711	20.000	ng	# 0.00
76) Chrysene-d12	14.121	240	268104	20.000	ng	-0.01
86) Perylene-d12	15.639	264	193032	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.563	112	689004	126.081	ng	0.00
7) Phenol-d6	6.557	99	862023	121.387	ng	0.00
23) Nitrobenzene-d5	7.504	82	552229	91.331	ng	-0.01
42) 2,4,6-Tribromophenol	10.780	330	223893	132.414	ng	0.00
45) 2-Fluorobiphenyl	9.304	172	1067519	80.352	ng	0.00
79) Terphenyl-d14	13.068	244	1186135	81.071	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102922\
Data File : BF130852.D
Acq On : 29 Oct 2022 12:39
Operator : CG\JU
Sample : PB148592BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB148592BL

Quant Time: Oct 30 23:22:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 05:41:34 2022
Response via : Initial Calibration

