


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102922\
 Data File : BF130850.D
 Acq On : 29 Oct 2022 11:08
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 02 08:18:35 2022






AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.0	47785	PASS
68	69	0.00	2	1.8	774	PASS
69	198	0.00	100	38.2	42506	PASS
70	69	0.00	2	0.3	148	PASS
127	198	10	80	48.7	54104	PASS
197	198	0.00	2	1.0	1077	PASS
198	198	100	100	100.0	111136	PASS
199	198	5	9	6.5	7212	PASS
275	198	10	60	26.1	29004	PASS
365	198	1	100	3.7	4130	PASS
441	198	0.01	100	14.9	16576	PASS
442	442	50	100	100.0	109533	PASS
443	442	15	24	18.4	20114	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102922\
 Data File : BF130850.D
 Acq On : 29 Oct 2022 11:08
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Oct 31 04:58:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 29 05:41:34 2022
 Response via : Initial Calibration



TIC: BF130850.D\data.ms

(70) Pentachlorophenol (C)

11.274min (-0.006) 17474.76 ng




response 63199

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.70	65.01
264.00	64.50	61.78
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102922\
Data File : BF130850.D
Acq On : 29 Oct 2022 11:08
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Oct 31 04:58:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 29 05:41:34 2022
Response via : Initial Calibration



TIC: BF130850.D\data.ms

(77) Benzidine

12.821min (-0.000) 0.00 ng

response 557825

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.07
183.00	12.10	11.65
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/29/2022	BNA_F	<u>BF130850.D</u>
Compound Name	Response	Retention Time
DDT	161667	13.645
DDD	2342	13.339
DDE	215	13.004
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2557	164224	1.56

Instrument :
BNA_F
ClientSampleId :
DFTPP