

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110522\
 Data File : BF131017.D
 Acq On : 05 Nov 2022 13:42
 Operator : CG\JU
 Sample : N5340-04
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-8-102722

Quant Time: Nov 07 00:48:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 00:46:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	57975	20.000	ng	0.00
21) Naphthalene-d8	8.187	136	230845	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	120838	20.000	ng	0.00
64) Phenanthrene-d10	11.439	188	201781	20.000	ng	0.00
76) Chrysene-d12	14.086	240	134090	20.000	ng	0.00
86) Perylene-d12	15.592	264	147727	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	270743	80.960	ng	0.00
7) Phenol-d6	6.522	99	223791	51.497	ng	0.00
23) Nitrobenzene-d5	7.469	82	516011	134.787	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	178528	175.507	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	850103	106.362	ng	0.00
79) Terphenyl-d14	13.033	244	700874	95.780	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110522\
Data File : BF131017.D
Acq On : 05 Nov 2022 13:42
Operator : CG\JU
Sample : N5340-04
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-8-102722

Quant Time: Nov 07 00:48:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:46:02 2022
Response via : Initial Calibration

