

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110522\
 Data File : BF131033.D
 Acq On : 05 Nov 2022 22:38
 Operator : CG\JU
 Sample : N5335-06DL 10X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-2DL

Quant Time: Nov 07 00:52:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 00:46:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	87684	20.000	ng	0.00
21) Naphthalene-d8	8.186	136	333561	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	180334	20.000	ng	0.00
64) Phenanthrene-d10	11.439	188	307915	20.000	ng	0.00
76) Chrysene-d12	14.086	240	201616	20.000	ng	0.00
86) Perylene-d12	15.586	264	191902	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	41967	8.297	ng	0.00
7) Phenol-d6	6.522	99	32271	4.910	ng	0.00
23) Nitrobenzene-d5	7.463	82	67783	12.253	ng	-0.01
42) 2,4,6-Tribromophenol	10.739	330	23985	15.800	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	128691	10.789	ng	0.00
79) Terphenyl-d14	13.027	244	120768	10.976	ng	0.00
Target Compounds						
35) Caprolactam	8.592	113	29677	18.962	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110522\
Data File : BF131033.D
Acq On : 05 Nov 2022 22:38
Operator : CG\JU
Sample : N5335-06DL 10X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-2DL

Quant Time: Nov 07 00:52:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:46:02 2022
Response via : Initial Calibration

