




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110522\
Data File : BF131014.D
Acq On : 05 Nov 2022 11:27
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 07 03:19:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:46:02 2022
Response via : Initial Calibration



TIC: BF131014.D\data.ms




(70) Pentachlorophenol (C)

11.239min (-0.006) 22760.45 ng

response 81020

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.70	65.34
264.00	64.50	66.34
0.00	0.00	0.00

Quant Time: Nov 07 03:19:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:46:02 2022
Response via : Initial Calibration



TIC: BF131014.D\data.ms

(77) Benzidine

12.786min (+ 0.006) 0.00 ng

response **556682**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	13.57
183.00	12.10	11.79
0.00	0.00	0.00