

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111020\
 Data File : BF122289.D
 Acq On : 10 Nov 2020 10:02
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 10 12:38:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	39.581	1.0	106	0.00
3	Pyridine	40.000	39.224	1.9	103	0.00
4	n-Nitrosodimethylamine	40.000	39.675	0.8	105	0.00
5 S	2-Fluorophenol	80.000	79.706	0.4	105	0.00
6	Aniline	40.000	39.511	1.2	104	0.00
7 S	Phenol-d6	80.000	78.514	1.9	104	0.00
8	2-Chlorophenol	40.000	39.931	0.2	104	0.00
9	Benzaldehyde	40.000	38.471	3.8	103	0.00
10 C	Phenol	40.000	39.667	0.8	104	0.00
11	bis(2-Chloroethyl)ether	40.000	39.505	1.2	105	0.00
12	1,3-Dichlorobenzene	40.000	39.048	2.4	103	0.00
13 C	1,4-Dichlorobenzene	40.000	39.319	1.7	104	0.00
14	1,2-Dichlorobenzene	40.000	39.621	0.9	104	0.00
15	Benzyl Alcohol	40.000	39.964	0.1	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.300	1.8	104	0.00
17	2-Methylphenol	40.000	39.541	1.1	105	0.00
18	Hexachloroethane	40.000	40.492	-1.2	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.133	2.2	104	0.00
20	3+4-Methylphenols	40.000	40.078	-0.2	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	39.108	2.2	104	0.00
23 S	Nitrobenzene-d5	80.000	87.619	-9.5	109	0.00
24	Nitrobenzene	40.000	42.295	-5.7	107	0.00
25	Isophorone	40.000	38.801	3.0	105	0.00
26 C	2-Nitrophenol	40.000	39.177	2.1	113	0.00
27	2,4-Dimethylphenol	40.000	39.931	0.2	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.204	2.0	104	0.00
29 C	2,4-Dichlorophenol	40.000	40.707	-1.8	106	0.00
30	1,2,4-Trichlorobenzene	40.000	39.730	0.7	104	0.00
31	Naphthalene	40.000	39.022	2.4	104	0.00
32	Benzoic acid	40.000	41.754	-4.4	118	0.00
33	4-Chloroaniline	40.000	39.308	1.7	104	0.00
34 C	Hexachlorobutadiene	40.000	40.326	-0.8	107	0.00
35	Caprolactam	40.000	41.016	-2.5	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.197	-0.5	105	0.00
37	2-Methylnaphthalene	40.000	39.519	1.2	106	0.00
38	1-Methylnaphthalene	40.000	39.477	1.3	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.734	0.7	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.122	-5.3	106	0.00
42 S	2,4,6-Tribromophenol	80.000	84.577	-5.7	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.894	-2.2	104	0.00
44	2,4,5-Trichlorophenol	40.000	41.536	-3.8	107	0.00
45 S	2-Fluorobiphenyl	80.000	76.506	4.4	105	0.00
46	1,1'-Biphenyl	40.000	39.050	2.4	105	0.00
47	2-Chloronaphthalene	40.000	39.891	0.3	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111020\
 Data File : BF122289.D
 Acq On : 10 Nov 2020 10:02
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 10 12:38:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.308	-0.8	113	0.00
49	Acenaphthylene	40.000	39.695	0.8	105	0.00
50	Dimethylphthalate	40.000	39.798	0.5	108	0.00
51	2,6-Dinitrotoluene	40.000	40.287	-0.7	111	0.00
52 C	Acenaphthene	40.000	39.690	0.8	106	0.00
53	3-Nitroaniline	40.000	39.817	0.5	110	0.00
54 P	2,4-Dinitrophenol	40.000	42.669	-6.7	124	0.00
55	Dibenzofuran	40.000	39.302	1.7	105	0.00
56 P	4-Nitrophenol	40.000	39.984	0.0	110	0.00
57	2,4-Dinitrotoluene	40.000	40.890	-2.2	113	0.00
58	Fluorene	40.000	38.742	3.1	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.796	-4.5	106	0.00
60	Diethylphthalate	40.000	40.366	-0.9	109	0.00
61	4-Chlorophenyl-phenylether	40.000	39.545	1.1	106	0.00
62	4-Nitroaniline	40.000	39.562	1.1	108	0.00
63	Azobenzene	40.000	39.423	1.4	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.166	-5.4	120	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.271	1.8	107	0.00
67	4-Bromophenyl-phenylether	40.000	39.972	0.1	108	0.00
68	Hexachlorobenzene	40.000	39.124	2.2	106	0.00
69	Atrazine	40.000	40.426	-1.1	108	0.00
70 C	Pentachlorophenol	40.000	41.381	-3.5	109	0.00
71	Phenanthrene	40.000	39.183	2.0	107	0.00
72	Anthracene	40.000	38.624	3.4	105	0.00
73	Carbazole	40.000	38.979	2.6	106	0.00
74	Di-n-butylphthalate	40.000	39.942	0.1	108	0.00
75 C	Fluoranthene	40.000	38.924	2.7	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	39.352	1.6	102	0.00
78	Pyrene	40.000	39.399	1.5	106	0.00
79 S	Terphenyl-d14	80.000	76.517	4.4	106	0.00
80	Butylbenzylphthalate	40.000	39.886	0.3	110	0.00
81	Benzo(a)anthracene	40.000	39.258	1.9	105	0.00
82	3,3'-Dichlorobenzidine	40.000	40.644	-1.6	105	0.00
83	Chrysene	40.000	39.481	1.3	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.497	-3.7	110	0.00
85 c	Di-n-octyl phthalate	40.000	42.870	-7.2	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.131	7.2	98	0.00
88	Benzo(b)fluoranthene	40.000	43.184	-8.0	108	0.00
89	Benzo(k)fluoranthene	40.000	38.445	3.9	99	0.00
90 C	Benzo(a)pyrene	40.000	40.499	-1.2	102	0.00
91	Dibenzo(a,h)anthracene	40.000	37.605	6.0	99	0.00
92	Benzo(g,h,i)perylene	40.000	36.462	8.8	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111020\
Data File : BF122289.D
Acq On : 10 Nov 2020 10:02
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 10 12:38:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 12:25:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0