

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140340.D
 Acq On : 13 Nov 2024 12:48
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Nov 13 14:28:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 13:11:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	147186	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	589077	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	334544	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	628587	20.000	ng	0.00
76) Chrysene-d12	14.063	240	385940	20.000	ng	0.01
86) Perylene-d12	15.580	264	322417	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	677089	78.544	ng	0.00
7) Phenol-d6	6.504	99	945074	80.918	ng	0.00
23) Nitrobenzene-d5	7.440	82	884247	78.210	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	257655	77.449	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	1568502	75.369	ng	0.00
79) Terphenyl-d14	12.986	244	1733989	77.987	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.634	88	151541	39.442	ng	99
3) Pyridine	3.399	79	387454	41.019	ng	99
4) n-Nitrosodimethylamine	3.352	42	193917	40.620	ng	97
6) Aniline	6.540	93	427252	42.052	ng	100
8) 2-Chlorophenol	6.663	128	371904	40.162	ng	100
9) Benzaldehyde	6.428	77	262348	37.479	ng	99
10) Phenol	6.516	94	507443	41.199	ng	99
11) bis(2-Chloroethyl)ether	6.616	93	371673	39.825	ng	99
12) 1,3-Dichlorobenzene	6.816	146	404387	39.579	ng	99
13) 1,4-Dichlorobenzene	6.893	146	409423	39.520	ng	99
14) 1,2-Dichlorobenzene	7.051	146	380948	39.610	ng	100
15) Benzyl Alcohol	7.016	79	356418	42.357	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.151	45	534596	41.471	ng	99
17) 2-Methylphenol	7.128	107	313127	41.105	ng	99
18) Hexachloroethane	7.387	117	151355	39.519	ng	98
19) n-Nitroso-di-n-propyla...	7.287	70	288974	40.966	ng	98
20) 3+4-Methylphenols	7.281	107	397692	41.745	ng	96
22) Acetophenone	7.287	105	529699	38.662	ng	98
24) Nitrobenzene	7.457	77	462701	39.307	ng	99
25) Isophorone	7.698	82	793806	39.616	ng	99
26) 2-Nitrophenol	7.775	139	210859	40.366	ng	99
27) 2,4-Dimethylphenol	7.804	122	262308	39.222	ng	97
28) bis(2-Chloroethoxy)met...	7.904	93	479505	39.027	ng	100
29) 2,4-Dichlorophenol	8.016	162	323936	39.193	ng	99
30) 1,2,4-Trichlorobenzene	8.098	180	355366	38.619	ng	100
31) Naphthalene	8.181	128	1157149	38.723	ng	100
32) Benzoic acid	7.928	122	249383	42.449	ng	96
33) 4-Chloroaniline	8.228	127	400271	38.516	ng	100
34) Hexachlorobutadiene	8.292	225	227805	38.855	ng	99
35) Caprolactam	8.598	113	108216	38.943	ng	96
36) 4-Chloro-3-methylphenol	8.704	107	364230	39.766	ng	99
37) 2-Methylnaphthalene	8.869	142	744201	38.839	ng	99
38) 1-Methylnaphthalene	8.969	142	736798	39.267	ng	99
40) 1,2,4,5-Tetrachloroben...	9.034	216	354978	38.371	ng	99
41) Hexachlorocyclopentadiene	9.022	237	93127	39.207	ng	99
43) 2,4,6-Trichlorophenol	9.145	196	240824	39.666	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111324\
 Data File : BF140340.D
 Acq On : 13 Nov 2024 12:48
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Quant Time: Nov 13 14:28:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 13:11:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.187	196	260135	39.190	ng	99
46) 1,1'-Biphenyl	9.334	154	939571	38.605	ng	99
47) 2-Chloronaphthalene	9.357	162	712904	38.453	ng	99
48) 2-Nitroaniline	9.457	65	243614	39.621	ng	99
49) Acenaphthylene	9.775	152	1094318	39.012	ng	100
50) Dimethylphthalate	9.634	163	843750	38.694	ng	100
51) 2,6-Dinitrotoluene	9.698	165	198967	40.024	ng	99
52) Acenaphthene	9.951	154	733408	38.711	ng	100
53) 3-Nitroaniline	9.869	138	207550	39.755	ng	100
54) 2,4-Dinitrophenol	9.975	184	98545	38.441	ng	98
55) Dibenzofuran	10.122	168	1042453	38.868	ng	100
56) 4-Nitrophenol	10.022	139	157938	41.475	ng	98
57) 2,4-Dinitrotoluene	10.104	165	258886	39.569	ng	98
58) Fluorene	10.463	166	804084	37.951	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.239	232	212533	40.586	ng	99
60) Diethylphthalate	10.334	149	861287	38.627	ng	99
61) 4-Chlorophenyl-phenyle...	10.451	204	403914	38.614	ng	99
62) 4-Nitroaniline	10.481	138	213640	40.022	ng	100
63) Azobenzene	10.616	77	865008	39.263	ng	99
65) 4,6-Dinitro-2-methylph...	10.510	198	136662	40.409	ng	98
66) n-Nitrosodiphenylamine	10.575	169	703286	38.723	ng	100
67) 4-Bromophenyl-phenylether	10.945	248	243728	38.789	ng	100
68) Hexachlorobenzene	11.010	284	273839	38.734	ng	98
69) Atrazine	11.098	200	181993	33.121	ng	99
70) Pentachlorophenol	11.204	266	160120	42.046	ng	98
71) Phenanthrene	11.428	178	1135052	38.440	ng	100
72) Anthracene	11.481	178	1117453	38.613	ng	100
73) Carbazole	11.633	167	1101588	38.551	ng	100
74) Di-n-butylphthalate	11.957	149	1350065	39.365	ng	100
75) Fluoranthene	12.616	202	1279456	38.621	ng	100
77) Benzidine	12.739	184	504166	34.036	ng	99
78) Pyrene	12.845	202	1295809	39.252	ng	100
80) Butylbenzylphthalate	13.463	149	532219	41.968	ng	97
81) Benzo(a)anthracene	14.051	228	975194	38.447	ng	99
82) 3,3'-Dichlorobenzidine	14.010	252	310326	38.491	ng	99
83) Chrysene	14.086	228	901424	38.950	ng	100
84) Bis(2-ethylhexyl)phtha...	14.033	149	715490	42.269	ng	100
85) Di-n-octyl phthalate	14.680	149	995484	41.757	ng	100
87) Indeno(1,2,3-cd)pyrene	17.086	276	811059	40.723	ng	100
88) Benzo(b)fluoranthene	15.139	252	814405	38.614	ng	100
89) Benzo(k)fluoranthene	15.174	252	683609	39.676	ng	99
90) Benzo(a)pyrene	15.516	252	648346	39.585	ng	99
91) Dibenzo(a,h)anthracene	17.104	278	670243	40.712	ng	99
92) Benzo(g,h,i)perylene	17.545	276	685184	40.711	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 13 14:28:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 13:11:08 2024
Response via : Initial Calibration

