

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112822\
 Data File : BF131423.D
 Acq On : 28 Nov 2022 11:22
 Operator : CG\JU
 Sample : N5747-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUP112122


Quant Time: Nov 29 00:20:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 00:18:36 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.951	152	73672	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	258927	20.000	ng	0.00
39) Acenaphthene-d10	10.004	164	148538	20.000	ng	0.00
64) Phenanthrene-d10	11.504	188	250945	20.000	ng	0.00
76) Chrysene-d12	14.162	240	156772	20.000	ng	0.00
86) Perylene-d12	15.698	264	152316	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.563	112	277601	71.714	ng	0.00
7) Phenol-d6	6.569	99	241945	48.988	ng	0.00
23) Nitrobenzene-d5	7.522	82	518731	100.988	ng	0.00
42) 2,4,6-Tribromophenol	10.804	330	228834	183.117	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	978048	95.793	ng	0.00
79) Terphenyl-d14	13.098	244	676522	70.567	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.510	77	107630	34.225	ng	# 1
15) Benzyl Alcohol	7.016	79	102946	25.421	ng	# 11
18) Hexachloroethane	7.486	117	121535	63.972	ng	# 1
19) n-Nitroso-di-n-propyla...	7.522	70	81778	22.619	ng	# 78
22) Acetophenone	7.345	105	141669	21.495	ng	# 54
24) Nitrobenzene	7.528	77	20201	3.775	ng	# 5
25) Isophorone	7.522	82	511369	55.801	ng	# 68
26) 2-Nitrophenol	7.880	139	75	3.626	ng	# 1
27) 2,4-Dimethylphenol	7.880	122	10725	3.019	ng	95
31) Naphthalene	8.269	128	1748725	126.736	ng	99
32) Benzoic acid	8.016	122	6056	8.290	ng	# 47
33) 4-Chloroaniline	8.269	127	234380	43.057	ng	# 29
37) 2-Methylnaphthalene	8.963	142	639344	68.386	ng	99
38) 1-Methylnaphthalene	9.063	142	458307	50.553	ng	100
46) 1,1'-Biphenyl	9.422	154	48421	4.272	ng	99
52) Acenaphthene	10.039	154	21119	2.433	ng	# 85
54) 2,4-Dinitrophenol	10.151	184	229	5.246	ng	# 1
58) Fluorene	10.557	166	25510	2.618	ng	94
63) Azobenzene	10.727	77	22971	2.186	ng	# 1
65) 4,6-Dinitro-2-methylph...	10.627	198	298	4.884	ng	# 1
67) 4-Bromophenyl-phenylether	10.804	248	15605	5.074	ng	# 1
70) Pentachlorophenol	11.292	266	77	4.916	ng	# 19
71) Phenanthrene	11.527	178	27713	2.043	ng	97
72) Anthracene	11.527	178	27713	2.079	ng	96
80) Butylbenzylphthalate	13.574	149	426	4.227	ng	# 29
84) Bis(2-ethylhexyl)phtha...	14.133	149	3596	4.286	ng	# 83
85) Di-n-octyl phthalate	14.762	149	1444	4.090	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 29 00:20:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 00:18:36 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.951 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:152 Resp: 73672
Ion Ratio Lower Upper
152 100
150 149.4 117.8 176.8
115 64.1 47.1 70.7



#5
2-Fluorophenol
Concen: 71.714 ng
RT: 5.563 min Scan# 591
Delta R.T. 0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:112 Resp: 277601
Ion Ratio Lower Upper
112 100
64 74.8 58.6 88.0
63 59.5 31.4 47.0#









#7
Phenol-d6
Concen: 48.988 ng
RT: 6.569 min Scan# 762
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion: 99 Resp: 241945
Ion Ratio Lower Upper
99 100
42 24.6 19.4 29.2
71 42.5 32.3 48.5



#9
Benzaldehyde
Concen: 34.225 ng
RT: 6.510 min Scan# 752
Delta R.T. 0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion: 77 Resp: 107630
Ion Ratio Lower Upper
77 100
105 746.4 74.5 114.5#
106 67.3 68.1 108.1#





#15

Benzyl Alcohol

Concen: 25.421 ng

RT: 7.016 min Scan# 81

Delta R.T. -0.077 min

Lab File: BF131423.D

Acq: 28 Nov 2022 11:22

Instrument :

BNA_F

ClientSampleId :

DUP112122





Tgt Ion: 79 Resp: 102946

Ion Ratio Lower Upper

79 100

108 0.0 55.1 82.7#

77 138.0 51.8 77.8#



#18

Hexachloroethane

Concen: 63.972 ng

RT: 7.486 min Scan# 918

Delta R.T. 0.017 min

Lab File: BF131423.D

Acq: 28 Nov 2022 11:22




Tgt Ion: 117 Resp: 121535


Ion Ratio Lower Upper

117 100

119 350.3 76.6 115.0#

201 0.0 87.0 130.6#









#19
n-Nitroso-di-n-propylamine
Concen: 22.619 ng
RT: 7.522 min Scan# 91
Delta R.T. 0.153 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion: 70 Resp: 81778
Ion Ratio Lower Upper
70 100
42 56.5 57.6 86.4#
101 3.4 8.2 12.2#
130 4.1 16.3 24.5#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1047
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion: 136 Resp: 258927
Ion Ratio Lower Upper
136 100
137 11.1 9.0 13.6
54 9.8 9.4 14.0
68 5.4 4.5 6.7









#22
Acetophenone
Concen: 21.495 ng
RT: 7.345 min Scan# 894
Delta R.T. -0.024 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:	105	Resp:	141669
Ion	Ratio	Lower	Upper
105	100		
71	0.3	4.9	7.3#
51	5.1	29.0	43.4#
120	0.3	16.9	25.3#



#23
Nitrobenzene-d5
Concen: 100.988 ng
RT: 7.522 min Scan# 924
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:	82	Resp:	518731
Ion	Ratio	Lower	Upper
82	100		
128	40.5	31.4	47.0
54	61.8	49.8	74.6









#24
Nitrobenzene
Concen: 3.775 ng
RT: 7.528 min Scan# 91
Delta R.T. -0.018 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion: 77 Resp: 20201
Ion Ratio Lower Upper
77 100
123 0.0 33.3 49.9#
65 87.9 12.7 19.1#



#25
Isophorone
Concen: 55.801 ng
RT: 7.522 min Scan# 924
Delta R.T. -0.259 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion: 82 Resp: 511369
Ion Ratio Lower Upper
82 100
95 0.1 6.2 9.2#
138 0.0 12.7 19.1#









#26
2-Nitrophenol
Concen: 3.626 ng
RT: 7.880 min Scan# 91
Delta R.T. 0.023 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:139 Resp: 75
Ion Ratio Lower Upper
139 100
109 136.6 29.9 44.9#
65 1436.6 55.7 83.5#



#27
2,4-Dimethylphenol
Concen: 3.019 ng
RT: 7.880 min Scan# 985
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:122 Resp: 10725
Ion Ratio Lower Upper
122 100
107 122.8 99.9 149.9
121 50.0 47.4 71.2









#31
Naphthalene
Concen: 126.736 ng
RT: 8.269 min Scan# 1051
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:128 Resp: 1748725
Ion Ratio Lower Upper
128 100
129 11.8 9.2 13.8
127 13.5 10.6 16.0



#32
Benzoic acid
Concen: 8.290 ng
RT: 8.016 min Scan# 1008
Delta R.T. 0.012 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:122 Resp: 6056
Ion Ratio Lower Upper
122 100
105 27.7 103.8 143.8#
77 84.2 74.3 114.3









#33
4-Chloroaniline
Concen: 43.057 ng
RT: 8.269 min Scan# 1169
Delta R.T. -0.047 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:127 Resp: 234380
Ion Ratio Lower Upper
127 100
129 87.3 26.4 39.6#
65 0.0 31.0 46.6#
92 0.2 16.3 24.5#



#37
2-Methylnaphthalene
Concen: 68.386 ng
RT: 8.963 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:142 Resp: 639344
Ion Ratio Lower Upper
142 100
141 88.0 71.0 106.4









#38
1-Methylnaphthalene
Concen: 50.553 ng
RT: 9.063 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:142 Resp: 458307
Ion Ratio Lower Upper
142 100
141 91.9 73.4 110.0



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:164 Resp: 148538
Ion Ratio Lower Upper
164 100
162 101.6 80.6 120.8
160 47.0 34.4 51.6









#42
 2,4,6-Tribromophenol
 Concen: 183.117 ng
 RT: 10.804 min Scan# 1482
 Delta R.T. -0.000 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22




Instrument :
 BNA_F
 ClientSampleId :
 DUP112122


Tgt Ion:330 Resp: 228834
 Ion Ratio Lower Upper
 330 100
 332 97.6 77.5 116.3
 141 36.2 29.3 43.9



#45
 2-Fluorobiphenyl
 Concen: 95.793 ng
 RT: 9.322 min Scan# 1230
 Delta R.T. -0.006 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22

Tgt Ion:172 Resp: 978048
 Ion Ratio Lower Upper
 172 100
 171 35.9 28.2 42.2
 170 22.9 18.7 28.1









#46
1,1'-Biphenyl
Concen: 4.272 ng
RT: 9.422 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:154 Resp: 48421
Ion Ratio Lower Upper
154 100
153 40.0 19.8 59.8
76 11.8 0.0 30.2



#52
Acenaphthene
Concen: 2.433 ng
RT: 10.039 min Scan# 1352
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:154 Resp: 21119
Ion Ratio Lower Upper
154 100
153 126.5 88.4 132.6
152 64.9 42.8 64.2#









#54
2,4-Dinitrophenol
Concen: 5.246 ng
RT: 10.151 min Scan# 11
Delta R.T. 0.088 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:184 Resp: 229
Ion Ratio Lower Upper
184 100
63 274.9 72.1 108.1#
154 224.1 56.3 84.5#



#58
Fluorene
Concen: 2.618 ng
RT: 10.557 min Scan# 1440
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:166 Resp: 25510
Ion Ratio Lower Upper
166 100
165 105.0 79.0 118.4
167 14.3 10.9 16.3





#63

Azobenzene

Concen: 2.186 ng

RT: 10.727 min Scan# 1469

Delta R.T. 0.012 min

Lab File: BF131423.D

Acq: 28 Nov 2022 11:22

Instrument :

BNA_F

ClientSampleId :

DUP112122

Tgt Ion: 77 Resp: 22971





Ion Ratio Lower Upper

77 100

182 87.1 3.1 43.1#

105 141.6 0.0 39.2#

51 48.8 17.8 57.8



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.504 min Scan# 1601

Delta R.T. -0.000 min

Lab File: BF131423.D

Acq: 28 Nov 2022 11:22




Tgt Ion: 188 Resp: 250945


Ion Ratio Lower Upper

188 100

94 7.3 6.3 9.5

80 7.7 6.8 10.2









#65
4,6-Dinitro-2-methylphenol
Concen: 4.884 ng
RT: 10.627 min Scan# 1450
Delta R.T. 0.023 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:198 Resp: 298
Ion Ratio Lower Upper
198 100
51 419.4 50.1 90.1#
105 315.8 32.5 72.5#



#67
4-Bromophenyl-phenylether
Concen: 5.074 ng
RT: 10.804 min Scan# 1482
Delta R.T. -0.241 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Tgt Ion:248 Resp: 15605
Ion Ratio Lower Upper
248 100
250 208.9 78.3 117.5#
141 490.4 58.0 87.0#









#70
 Pentachlorophenol
 Concen: 4.916 ng
 RT: 11.292 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22




Instrument :
 BNA_F
 ClientSampleId :
 DUP112122


Tgt Ion:266 Resp: 77
 Ion Ratio Lower Upper
 266 100
 268 0.0 50.6 75.8#
 264 0.0 50.4 75.6#



#71
 Phenanthrene
 Concen: 2.043 ng
 RT: 11.527 min Scan# 1605
 Delta R.T. -0.006 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22

Tgt Ion:178 Resp: 27713
 Ion Ratio Lower Upper
 178 100
 176 21.2 15.6 23.4
 179 16.1 12.4 18.6









#72
 Anthracene
 Concen: 2.079 ng
 RT: 11.527 min Scan# 1
 Delta R.T. -0.059 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22




Instrument :
 BNA_F
 ClientSampleId :
 DUP112122


Tgt Ion:178 Resp: 27713
 Ion Ratio Lower Upper
 178 100
 176 21.2 14.7 22.1
 179 16.1 12.2 18.4



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.162 min Scan# 2053
 Delta R.T. -0.006 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22

Tgt Ion:240 Resp: 156772
 Ion Ratio Lower Upper
 240 100
 120 7.2 5.8 8.8
 236 26.4 20.3 30.5









#79
 Terphenyl-d14
 Concen: 70.567 ng
 RT: 13.098 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22




Instrument :
 BNA_F
 ClientSampleId :
 DUP112122


Tgt Ion:244 Resp: 676522
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.9 8.9
 122 9.5 7.5 11.3



#80
 Butylbenzylphthalate
 Concen: 4.227 ng
 RT: 13.574 min Scan# 1953
 Delta R.T. -0.006 min
 Lab File: BF131423.D
 Acq: 28 Nov 2022 11:22

Tgt Ion:149 Resp: 426
 Ion Ratio Lower Upper
 149 100
 91 144.8 61.9 92.9#
 206 0.0 16.5 24.7#









#84
Bis(2-ethylhexyl)phthalate
Concen: 4.286 ng
RT: 14.133 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22




Instrument :
BNA_F
ClientSampleId :
DUP112122


Tgt Ion:149 Resp: 3596
Ion Ratio Lower Upper
149 100
167 37.2 21.9 32.9#
279 5.4 3.0 4.4#



#85
Di-n-octyl phthalate
Concen: 4.090 ng
RT: 14.762 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22


Tgt Ion:149 Resp: 1444
Ion Ratio Lower Upper
149 100
167 3.7 1.2 1.8#
43 59.0 10.7 16.1#





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.698 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF131423.D
Acq: 28 Nov 2022 11:22

Instrument :
BNA_F
ClientSampleId :
DUP112122



Tgt Ion:264 Resp: 152316

Ion	Ratio	Lower	Upper
264	100		
260	23.3	19.0	28.4
265	20.5	16.4	24.6

