

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112822\
 Data File : BF131430.D
 Acq On : 28 Nov 2022 15:27
 Operator : CG\JU
 Sample : PB149225TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB149225TB

Quant Time: Nov 29 00:55:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 29 00:18:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.951	152	83024	20.000	ng	0.00
21) Naphthalene-d8	8.239	136	310197	20.000	ng	0.00
39) Acenaphthene-d10	10.004	164	185722	20.000	ng	0.00
64) Phenanthrene-d10	11.504	188	362750	20.000	ng	0.00
76) Chrysene-d12	14.157	240	266799	20.000	ng	-0.01
86) Perylene-d12	15.692	264	201772	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.569	112	654826	150.110	ng	0.01
7) Phenol-d6	6.569	99	812517	145.984	ng	0.00
23) Nitrobenzene-d5	7.516	82	546349	88.785	ng	-0.01
42) 2,4,6-Tribromophenol	10.804	330	266870	170.798	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	1035905	81.146	ng	0.00
79) Terphenyl-d14	13.104	244	1406240	86.191	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112822\
Data File : BF131430.D
Acq On : 28 Nov 2022 15:27
Operator : CG\JU
Sample : PB149225TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB149225TB

Quant Time: Nov 29 00:55:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 29 00:18:36 2022
Response via : Initial Calibration

