

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/15/2021	BNA_F	<u>BF126194.D</u>
Compound Name	Response	Retention Time
DDT	592528	13.49
DDD	4289	13.19
DDE	351	12.86
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4640	597168	0.78