

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
 Data File : BF122783.D
 Acq On : 18 Dec 2020 13:37
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.594	0.577	2.9	91	0.02
3	Pyridine	1.569	1.526	2.7	88	0.01
4	n-Nitrosodimethylamine	0.629	0.571	9.2	83	0.02
5 S	2-Fluorophenol	1.263	1.179	6.7	87	0.00
6	Aniline	1.972	1.906	3.3	89	0.00
7 S	Phenol-d6	1.675	1.554	7.2	86	0.00
8	2-Chlorophenol	1.392	1.337	4.0	89	0.00
9	Benzaldehyde	1.014	0.978	3.6	102	0.00
10 C	Phenol	1.736	1.626	6.3	88	0.00
11	bis(2-Chloroethyl)ether	1.326	1.287	2.9	90	0.00
12	1,3-Dichlorobenzene	1.648	1.553	5.8	88	0.00
13 C	1,4-Dichlorobenzene	1.661	1.562	6.0	88	0.00
14	1,2-Dichlorobenzene	1.604	1.402	12.6	86	0.00
15	Benzyl Alcohol	1.154	1.064	7.8	83	0.00
16	2,2'-oxybis(1-Chloropropane	1.891	1.703	9.9	83	0.00
17	2-Methylphenol	1.107	1.110	-0.3	91	0.00
18	Hexachloroethane	0.592	0.560	5.4	88	0.00
19 P	n-Nitroso-di-n-propylamine	0.960	0.891	7.2	86	0.00
20	3+4-Methylphenols	1.398	1.260	9.9	86	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
22	Acetophenone	0.497	0.452	9.1	87	0.00
23 S	Nitrobenzene-d5	0.399	0.440	-10.3	104	0.00
24	Nitrobenzene	0.397	0.371	6.5	88	0.00
25	Isophorone	0.688	0.652	5.2	89	0.00
26 C	2-Nitrophenol	0.194	0.197	-1.5	93	0.00
27	2,4-Dimethylphenol	0.284	0.274	3.5	91	0.00
28	bis(2-Chloroethoxy)methane	0.471	0.441	6.4	88	0.00
29 C	2,4-Dichlorophenol	0.332	0.314	5.4	89	0.00
30	1,2,4-Trichlorobenzene	0.376	0.346	8.0	88	0.00
31	Naphthalene	1.104	1.028	6.9	89	0.00
32	Benzoic acid	0.226	0.209	7.5	81	0.01
33	4-Chloroaniline	0.461	0.443	3.9	90	0.00
34 C	Hexachlorobutadiene	0.231	0.211	8.7	86	0.00
35	Caprolactam	0.100	0.101	-1.0	94	0.01
36 C	4-Chloro-3-methylphenol	0.327	0.319	2.4	91	0.00
37	2-Methylnaphthalene	0.730	0.679	7.0	89	0.00
38	1-Methylnaphthalene	0.691	0.644	6.8	89	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.603	8.9	87	0.00
41 P	Hexachlorocyclopentadiene	0.293	0.332	-13.3	102	0.00
42 S	2,4,6-Tribromophenol	0.262	0.253	3.4	91	0.00
43 C	2,4,6-Trichlorophenol	0.458	0.414	9.6	85	0.00
44	2,4,5-Trichlorophenol	0.473	0.426	9.9	86	0.00
45 S	2-Fluorobiphenyl	1.479	1.420	4.0	99	0.00
46	1,1'-Biphenyl	1.660	1.512	8.9	88	0.00
47	2-Chloronaphthalene	1.375	1.263	8.1	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
 Data File : BF122783.D
 Acq On : 18 Dec 2020 13:37
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.401	0.388	3.2	90	0.00
49	Acenaphthylene	2.031	1.881	7.4	90	0.00
50	Dimethylphthalate	1.597	1.540	3.6	93	0.00
51	2,6-Dinitrotoluene	0.337	0.336	0.3	93	0.00
52 C	Acenaphthene	1.314	1.229	6.5	89	0.00
53	3-Nitroaniline	0.390	0.390	0.0	93	0.00
54 P	2,4-Dinitrophenol	0.165	0.166	-0.6	90	0.00
55	Dibenzofuran	1.849	1.676	9.4	88	0.00
56 P	4-Nitrophenol	0.278	0.258	7.2	84	0.00
57	2,4-Dinitrotoluene	0.449	0.453	-0.9	93	0.00
58	Fluorene	1.470	1.275	13.3	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.415	0.385	7.2	87	0.00
60	Diethylphthalate	1.554	1.451	6.6	90	0.00
61	4-Chlorophenyl-phenylether	0.724	0.655	9.5	89	0.00
62	4-Nitroaniline	0.396	0.400	-1.0	95	0.00
63	Azobenzene	1.428	1.312	8.1	88	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.125	-2.5	95	0.00
66 c	n-Nitrosodiphenylamine	0.707	0.647	8.5	90	0.00
67	4-Bromophenyl-phenylether	0.271	0.252	7.0	91	0.00
68	Hexachlorobenzene	0.300	0.277	7.7	91	0.00
69	Atrazine	0.229	0.187	18.3	81	0.00
70 C	Pentachlorophenol	0.159	0.142	10.7	80	0.00
71	Phenanthrene	1.189	1.075	9.6	91	0.00
72	Anthracene	1.179	1.069	9.3	91	0.00
73	Carbazole	1.069	0.990	7.4	92	0.00
74	Di-n-butylphthalate	1.343	1.201	10.6	89	0.00
75 C	Fluoranthene	1.246	1.121	10.0	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.433	0.484	-11.8	98	0.00
78	Pyrene	1.546	1.408	8.9	89	0.00
79 S	Terphenyl-d14	1.143	1.145	-0.2	104	0.00
80	Butylbenzylphthalate	0.697	0.637	8.6	89	0.00
81	Benzo(a)anthracene	1.337	1.258	5.9	94	0.00
82	3,3'-Dichlorobenzidine	0.479	0.467	2.5	96	0.00
83	Chrysene	1.330	1.231	7.4	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.954	0.850	10.9	90	0.00
85 c	Di-n-octyl phthalate	1.582	1.397	11.7	89	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.313	1.202	8.5	99	0.00
88	Benzo(b)fluoranthene	1.403	1.360	3.1	98	0.00
89	Benzo(k)fluoranthene	1.283	1.156	9.9	97	0.00
90 C	Benzo(a)pyrene	1.228	1.154	6.0	99	0.00
91	Dibenzo(a,h)anthracene	1.074	0.993	7.5	100	0.00
92	Benzo(g,h,i)perylene	1.040	0.939	9.7	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
Data File : BF122783.D
Acq On : 18 Dec 2020 13:37
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0