

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
 Data File : BF122783.D
 Acq On : 18 Dec 2020 13:37
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	38.882	2.8	91	0.02
3	Pyridine	40.000	38.880	2.8	88	0.01
4	n-Nitrosodimethylamine	40.000	36.301	9.2	83	0.02
5 S	2-Fluorophenol	80.000	74.688	6.6	87	0.00
6	Aniline	40.000	38.665	3.3	89	0.00
7 S	Phenol-d6	80.000	74.194	7.3	86	0.00
8	2-Chlorophenol	40.000	38.406	4.0	89	0.00
9	Benzaldehyde	40.000	38.572	3.6	102	0.00
10 C	Phenol	40.000	37.453	6.4	88	0.00
11	bis(2-Chloroethyl)ether	40.000	38.802	3.0	90	0.00
12	1,3-Dichlorobenzene	40.000	37.704	5.7	88	0.00
13 C	1,4-Dichlorobenzene	40.000	37.621	5.9	88	0.00
14	1,2-Dichlorobenzene	40.000	34.962	12.6	86	0.00
15	Benzyl Alcohol	40.000	36.898	7.8	83	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.013	10.0	83	0.00
17	2-Methylphenol	40.000	40.103	-0.3	91	0.00
18	Hexachloroethane	40.000	37.863	5.3	88	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.149	7.1	86	0.00
20	3+4-Methylphenols	40.000	36.034	9.9	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	36.347	9.1	87	0.00
23 S	Nitrobenzene-d5	80.000	88.267	-10.3	104	0.00
24	Nitrobenzene	40.000	37.347	6.6	88	0.00
25	Isophorone	40.000	37.914	5.2	89	0.00
26 C	2-Nitrophenol	40.000	40.610	-1.5	93	0.00
27	2,4-Dimethylphenol	40.000	38.586	3.5	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.456	6.4	88	0.00
29 C	2,4-Dichlorophenol	40.000	37.849	5.4	89	0.00
30	1,2,4-Trichlorobenzene	40.000	36.799	8.0	88	0.00
31	Naphthalene	40.000	37.257	6.9	89	0.00
32	Benzoic acid	40.000	36.986	7.5	81	0.01
33	4-Chloroaniline	40.000	38.369	4.1	90	0.00
34 C	Hexachlorobutadiene	40.000	36.500	8.8	86	0.00
35	Caprolactam	40.000	40.491	-1.2	94	0.01
36 C	4-Chloro-3-methylphenol	40.000	39.095	2.3	91	0.00
37	2-Methylnaphthalene	40.000	37.220	7.0	89	0.00
38	1-Methylnaphthalene	40.000	37.272	6.8	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.415	9.0	87	0.00
41 P	Hexachlorocyclopentadiene	40.000	45.363	-13.4	102	0.00
42 S	2,4,6-Tribromophenol	80.000	77.219	3.5	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.154	9.6	85	0.00
44	2,4,5-Trichlorophenol	40.000	36.022	9.9	86	0.00
45 S	2-Fluorobiphenyl	80.000	76.807	4.0	99	0.00
46	1,1'-Biphenyl	40.000	36.428	8.9	88	0.00
47	2-Chloronaphthalene	40.000	36.738	8.2	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
 Data File : BF122783.D
 Acq On : 18 Dec 2020 13:37
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.720	3.2	90	0.00
49	Acenaphthylene	40.000	37.037	7.4	90	0.00
50	Dimethylphthalate	40.000	38.572	3.6	93	0.00
51	2,6-Dinitrotoluene	40.000	39.805	0.5	93	0.00
52 C	Acenaphthene	40.000	37.414	6.5	89	0.00
53	3-Nitroaniline	40.000	40.041	-0.1	93	0.00
54 P	2,4-Dinitrophenol	40.000	40.258	-0.6	90	0.00
55	Dibenzofuran	40.000	36.271	9.3	88	0.00
56 P	4-Nitrophenol	40.000	37.183	7.0	84	0.00
57	2,4-Dinitrotoluene	40.000	40.357	-0.9	93	0.00
58	Fluorene	40.000	34.698	13.3	89	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.024	7.4	87	0.00
60	Diethylphthalate	40.000	37.357	6.6	90	0.00
61	4-Chlorophenyl-phenylether	40.000	36.182	9.5	89	0.00
62	4-Nitroaniline	40.000	40.434	-1.1	95	0.00
63	Azobenzene	40.000	36.761	8.1	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.020	-2.6	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.649	8.4	90	0.00
67	4-Bromophenyl-phenylether	40.000	37.207	7.0	91	0.00
68	Hexachlorobenzene	40.000	37.021	7.4	91	0.00
69	Atrazine	40.000	32.560	18.6	81	0.00
70 C	Pentachlorophenol	40.000	35.687	10.8	80	0.00
71	Phenanthrene	40.000	36.191	9.5	91	0.00
72	Anthracene	40.000	36.272	9.3	91	0.00
73	Carbazole	40.000	37.066	7.3	92	0.00
74	Di-n-butylphthalate	40.000	35.773	10.6	89	0.00
75 C	Fluoranthene	40.000	35.983	10.0	89	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	44.774	-11.9	98	0.00
78	Pyrene	40.000	36.410	9.0	89	0.00
79 S	Terphenyl-d14	80.000	80.156	-0.2	104	0.00
80	Butylbenzylphthalate	40.000	36.610	8.5	89	0.00
81	Benzo(a)anthracene	40.000	37.651	5.9	94	0.00
82	3,3'-Dichlorobenzidine	40.000	39.055	2.4	96	0.00
83	Chrysene	40.000	37.023	7.4	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	35.628	10.9	90	0.00
85 c	Di-n-octyl phthalate	40.000	35.308	11.7	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.636	8.4	99	0.00
88	Benzo(b)fluoranthene	40.000	38.764	3.1	98	0.00
89	Benzo(k)fluoranthene	40.000	36.026	9.9	97	0.00
90 C	Benzo(a)pyrene	40.000	37.607	6.0	99	0.00
91	Dibenzo(a,h)anthracene	40.000	36.962	7.6	100	0.00
92	Benzo(g,h,i)perylene	40.000	36.120	9.7	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121820\
Data File : BF122783.D
Acq On : 18 Dec 2020 13:37
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 18 14:24:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0