

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062719\
 Data File : BF115253.D
 Acq On : 28 Jun 2019 00:01
 Operator : HP/JU
 Sample : K3507-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NB-07-062619-B

Manual Integrations
 APPROVED

mohammad
 6/28/2019 2:58:46 PM

Quant Time: Jun 28 07:19:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:26:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	217815	20.00	ng	0.00
21) Naphthalene-d8	7.88	136	816041	20.00	ng	0.00
39) Acenaphthene-d10	9.62	164	408308	20.00	ng	0.00
64) Phenanthrene-d10	11.10	188	815647	20.00	ng	0.00
76) Chrysene-d12	13.71	240	650221	20.00	ng	0.00
87) Perylene-d12	15.08	264	581604	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	1056820	74.87	ng	0.00
7) Phenol-d6	6.23	99	1348728	78.73	ng	0.00
23) Nitrobenzene-d5	7.15	82	782747	59.01	ng	-0.01
42) 2,4,6-Tribromophenol	10.40	330	369603	85.84	ng	0.00
45) 2-Fluorobiphenyl	8.95	172	1564518	65.09	ng	0.00
79) Terphenyl-d14	12.68	244	1888782	61.76	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.35	163	297999	9.920	ng	99
71) Phenanthrene	11.11	178	173543	3.952	ng	98
75) Fluoranthene	12.29	202	340101	7.762	ng	96
78) Pyrene	12.52	202	452792	9.061	ng	99
81) Benzo(a)anthracene	13.70	228	288665	6.187	ng	99
83) Chrysene	13.74	228	230296	5.389	ng	95
88) Benzo(b)fluoranthene	14.70	252	236215m	6.509	ng	
89) Benzo(k)fluoranthene	14.72	252	66380m	2.040	ng	
90) Benzo(a)pyrene	15.02	252	140396	4.292	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF062719\
Data File : BF115253.D
Acq On : 28 Jun 2019 00:01
Operator : HP/JU
Sample : K3507-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NB-07-062619-B

Manual Integrations
APPROVED

mohammad
6/28/2019 2:58:46 PM

Quant Time: Jun 28 07:19:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF062119.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:26:43 2019
Response via : Initial Calibration

