


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF071219\
 Data File : BF115526.D
 Acq On : 12 Jul 2019 15:12
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF071219.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 12 14:03:52 2019



AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.3	38077	PASS
68	69	0.00	2	1.8	734	PASS
69	198	0.00	100	35.4	41802	PASS
70	69	0.00	2	0.3	116	PASS
127	198	10	80	46.5	54906	PASS
197	198	0.00	2	0.5	569	PASS
198	198	100	100	100.0	117984	PASS
199	198	5	9	6.7	7896	PASS
275	198	10	60	27.3	32161	PASS
365	198	1	100	3.1	3696	PASS
441	198	0.01	100	13.1	15402	PASS
442	198	50	100	82.3	97042	PASS
443	442	15	24	19.5	18935	PASS