

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF073119\
 Data File : BF115866.D
 Acq On : 31 Jul 2019 15:19
 Operator : HP/JU
 Sample : K3591-09
 Misc : MDL-WATER-8PPM
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Aug 01 09:12:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 01 09:04:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	139515	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	548906	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	289704	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	498722	20.00	ng	0.00
76) Chrysene-d12	14.02	240	382209	20.00	ng	-0.01
87) Perylene-d12	15.49	264	352151	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.47	112	845509	101.45	ng	0.00
7) Phenol-d6	6.48	99	1060114	100.82	ng	0.00
23) Nitrobenzene-d5	7.41	82	781768	82.21	ng	0.00
42) 2,4,6-Tribromophenol	10.68	330	277953	98.96	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	1293416	83.63	ng	0.00
79) Terphenyl-d14	12.96	244	1371397	75.61	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.61	88	34900	8.289	ng	99
3) Pyridine	3.42	79	78987	6.716	ng	99
4) n-Nitrosodimethylamine	3.30	42	33785	7.279	ng	# 99
6) Aniline	6.52	93	123217	8.183	ng	98
8) 2-Chlorophenol	6.64	128	77986	8.462	ng	95
9) Benzaldehyde	6.40	77	67932	9.363	ng	98
10) Phenol	6.49	94	105186	8.495	ng	100
11) bis(2-Chloroethyl)ether	6.59	93	73485	8.072	ng	96
12) 1,3-Dichlorobenzene	6.79	146	86532	8.125	ng	99
13) 1,4-Dichlorobenzene	6.87	146	88247	8.275	ng	98
14) 1,2-Dichlorobenzene	7.02	146	81623	8.313	ng	98
15) Benzyl Alcohol	6.99	79	71013	8.899	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.13	45	97409	7.845	ng	98
17) 2-Methylphenol	7.10	107	63164	8.094	ng	98
18) Hexachloroethane	7.36	117	29343	7.474	ng	95
19) n-Nitroso-di-n-propylamine	7.26	70	53785	8.255	ng	98
20) 3+4-Methylphenols	7.26	107	81166	8.790	ng	# 49
22) Acetophenone	7.26	105	111362	9.251	ng	98
24) Nitrobenzene	7.43	77	91001	8.863	ng	97
25) Isophorone	7.66	82	149572	8.226	ng	97
26) 2-Nitrophenol	7.75	139	36739	7.591	ng	99
27) 2,4-Dimethylphenol	7.78	122	64997	8.926	ng	98
28) bis(2-Chloroethoxy)methane	7.87	93	90256	8.008	ng	99
29) 2,4-Dichlorophenol	7.99	162	62028	8.068	ng	99
30) 1,2,4-Trichlorobenzene	8.07	180	70329	8.024	ng	97
31) Naphthalene	8.15	128	230192	8.912	ng	99
32) Benzoic acid	7.87	122	6043	1.177	ng	90
33) 4-Chloroaniline	8.20	127	92766	8.038	ng	97
34) Hexachlorobutadiene	8.27	225	39564	7.837	ng	97
35) Caprolactam	8.54	113	18719	7.528	ng	95
36) 4-Chloro-3-methylphenol	8.69	107	65124	8.142	ng	98
37) 2-Methylnaphthalene	8.85	142	153478	9.022	ng	98
38) 1-Methylnaphthalene	8.95	142	144326	8.968	ng	99
40) 1,2,4,5-Tetrachlorobenzene	9.02	216	66601	8.599	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF073119\
 Data File : BF115866.D
 Acq On : 31 Jul 2019 15:19
 Operator : HP/JU
 Sample : K3591-09
 Misc : MDL-WATER-8PPM
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Aug 01 09:12:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 01 09:04:21 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	9.00	237	12110	4.000	ng	97
43) 2,4,6-Trichlorophenol	9.13	196	38438	7.210	ng	97
44) 2,4,5-Trichlorophenol	9.17	196	40678	7.382	ng	98
46) 1,1'-Biphenyl	9.31	154	187395	9.162	ng	98
47) 2-Chloronaphthalene	9.34	162	140457	8.251	ng	99
48) 2-Nitroaniline	9.43	65	42243	7.508	ng	# 83
49) Acenaphthylene	9.75	152	227175	9.147	ng	98
50) Dimethylphthalate	9.60	163	155742	7.895	ng	99
51) 2,6-Dinitrotoluene	9.67	165	33800	7.885	ng	93
52) Acenaphthene	9.92	154	135304	8.881	ng	99
53) 3-Nitroaniline	9.84	138	37620	7.102	ng	98
54) 2,4-Dinitrophenol	9.96	184	4376	2.404	ng	90
55) Dibenzofuran	10.10	168	201174	9.080	ng	99
56) 4-Nitrophenol	10.02	139	16418	4.839	ng	89
57) 2,4-Dinitrotoluene	10.07	165	42413	7.431	ng	# 83
58) Fluorene	10.44	166	156632	9.188	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.22	232	35835	7.729	ng	98
60) Diethylphthalate	10.30	149	147490	8.009	ng	98
61) 4-Chlorophenyl-phenylether	10.43	204	74406	8.618	ng	97
62) 4-Nitroaniline	10.45	138	39623	7.239	ng	98
63) Azobenzene	10.59	77	160967	8.498	ng	98
65) 4,6-Dinitro-2-methylphenol	10.49	198	13116	4.963	ng	97
66) n-Nitrosodiphenylamine	10.55	169	132217	8.808	ng	98
67) 4-Bromophenyl-phenylether	10.92	248	43448	8.138	ng	93
68) Hexachlorobenzene	10.99	284	48202	7.808	ng	99
69) Atrazine	11.07	200	38667	7.830	ng	97
70) Pentachlorophenol	11.19	266	13924	4.646	ng	98
71) Phenanthrene	11.40	178	227614	8.928	ng	98
72) Anthracene	11.45	178	232695	9.018	ng	99
73) Carbazole	11.61	167	203359	8.687	ng	97
74) Di-n-butylphthalate	11.93	149	209605	7.676	ng	99
75) Fluoranthene	12.59	202	234666	8.792	ng	97
77) Benzidine	12.71	184	103945	8.050	ng	99
78) Pyrene	12.82	202	242041	8.401	ng	98
80) Butylbenzylphthalate	13.43	149	79624	6.533	ng	100
81) Benzo(a)anthracene	14.01	228	197168	8.071	ng	98
82) 3,3'-Dichlorobenzidine	13.97	252	65789	7.752	ng	99
83) Chrysene	14.04	228	194792	8.213	ng	98
84) Bis(2-ethylhexyl)phthalate	13.99	149	106848	6.747	ng	97
85) Di-n-octyl phthalate	14.60	149	158730	6.310	ng	99
86) Indeno(1,2,3-cd)pyrene	16.96	276	137356	6.895	ng	100
88) Benzo(b)fluoranthene	15.06	252	153080	7.518	ng	99
89) Benzo(k)fluoranthene	15.09	252	168498	8.496	ng	97
90) Benzo(a)pyrene	15.42	252	144903	7.961	ng	99
91) Dibenzo(a,h)anthracene	16.97	278	115016	7.300	ng	99
92) Benzo(g,h,i)perylene	17.40	276	112174	7.249	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF073119\
Data File : BF115866.D
Acq On : 31 Jul 2019 15:19
Operator : HP/JU
Sample : K3591-09
Misc : MDL-WATER-8PPM
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Aug 01 09:12:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 01 09:04:21 2019
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 6.85 min Scan# 810

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:152 Resp: 139515

Ion Ratio Lower Upper

152 100

150 154.4 127.4 191.0




115 62.0 51.8 77.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 8.289 ng

RT: 2.61 min Scan# 89

Delta R.T. 0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion: 88 Resp: 34900

Ion Ratio Lower Upper

88 100

58 68.7 54.1 81.1



43 26.2 20.7 31.1




Abundance Ion 88.00 (87.70 to 88.70): BF1

Ion 58.00 (57.70 to 58.70): BF1

Ion 43.00 (42.70 to 43.70): BF1









#3
 Pvridine
 Concen: 6.716 ng
 RT: 3.42 min Scan# 226
 Delta R.T. 0.06 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion: 79 Resp: 78987
 Ion Ratio Lower Upper
 79 100
 52 63.1 50.2 75.4
 51 30.8 25.8 38.6



#4
 n-Nitrosodimethylamine
 Concen: 7.279 ng
 RT: 3.30 min Scan# 207
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion: 42 Resp: 33785
 Ion Ratio Lower Upper
 42 100
 74 140.3 111.0 166.4
 44 7.6 4.7 7.1#





#5

2-Fluorophenol

Concen: 101.454 ng

RT: 5.47 min Scan# 575

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:112 Resp: 845509

Ion Ratio Lower Upper

112 100

64 58.1 46.8 70.2




63 30.8 24.2 36.2



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 8.183 ng

RT: 6.52 min Scan# 753

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion: 93 Resp: 123217

Ion Ratio Lower Upper

93 100

66 38.0 31.4 47.0



65 20.2 16.5 24.7



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 100.822 ng

RT: 6.48 min Scan# 747

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampled :


Tot Ion: 99 Resp: 1060114

Ion Ratio Lower Upper

99 100

42 17.2 13.8 20.6




71 33.3 27.1 40.7



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 8.462 ng

RT: 6.64 min Scan# 774

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:128 Resp: 77986

Ion Ratio Lower Upper

128 100

130 32.1 12.9 52.9



64 41.5 26.1 66.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1









#9
Benzaldehyde
Concen: 9.363 ng
RT: 6.40 min Scan# 734
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 77 Resp: 67932
Ion Ratio Lower Upper
77 100
105 97.8 79.0 119.0
106 92.5 76.2 116.2



#10
Phenol
Concen: 8.495 ng
RT: 6.49 min Scan# 749
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion: 94 Resp: 105186
Ion Ratio Lower Upper
94 100
65 28.9 8.9 48.9
66 38.5 18.0 58.0









#11
bis(2-Chloroethyl)ether
Concen: 8.072 ng
RT: 6.59 min Scan# 765
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 93 Resp: 73485
Ion Ratio Lower Upper
93 100
63 70.8 54.9 94.9
95 32.7 13.3 53.3



#12
1,3-Dichlorobenzene
Concen: 8.125 ng
RT: 6.79 min Scan# 800
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion: 146 Resp: 86532
Ion Ratio Lower Upper
146 100
148 65.1 52.1 78.1
75 32.7 25.0 37.6









#13
1,4-Dichlorobenzene
Concen: 8.275 ng
RT: 6.87 min Scan# 813
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampled :

Tot Ion:146 Resp: 88247
Ion Ratio Lower Upper
146 100
148 63.2 51.2 76.8
111 40.0 33.2 49.8




Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E





#14
1,2-Dichlorobenzene
Concen: 8.313 ng
RT: 7.02 min Scan# 839
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:146 Resp: 81623
Ion Ratio Lower Upper
146 100
148 63.1 51.6 77.4
111 42.8 33.6 50.4



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 8.899 ng

RT: 6.99 min Scan# 834





Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampled :

Tot Ion:	79	Resp:	71013
Ion	Ratio	Lower	Upper
79	100		
108	79.0	62.8	94.2
77	64.4	52.0	78.0



#16

2,2'-oxybis(1-Chloropropane)

Concen: 7.845 ng




RT: 7.13 min Scan# 857


Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Tgt Ion:	45	Resp:	97409
Ion	Ratio	Lower	Upper
45	100		
77	16.5	0.0	36.3
79	13.9	0.0	32.5









#17
2-Methylphenol
Concen: 8.094 ng
RT: 7.10 min Scan# 852
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :

Tot Ion:107 Resp: 63164
Ion Ratio Lower Upper
107 100
108 115.5 91.4 137.0
77 48.1 37.6 56.4
79 47.0 36.2 54.4



#18
Hexachloroethane
Concen: 7.474 ng
RT: 7.36 min Scan# 897
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:117 Resp: 29343
Ion Ratio Lower Upper
117 100
119 98.1 75.8 113.6
201 96.5 82.0 123.0









#19
n-Nitroso-di-n-propylamine
Concen: 8.255 ng
RT: 7.26 min Scan# 879
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 70 Resp: 53785
Ion Ratio Lower Upper
70 100
42 54.1 41.8 62.8
101 9.4 7.7 11.5
130 21.0 17.8 26.6



#20
3+4-Methylphenols
Concen: 8.790 ng
RT: 7.26 min Scan# 879
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:107 Resp: 81166
Ion Ratio Lower Upper
107 100
108 90.7 67.9 107.9
77 202.9 77.3 117.3#
79 28.5 6.6 46.6





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.13 min Scan# 1028

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 548906



Ion Ratio Lower Upper

136 100

137 11.3 9.1 13.7

54 6.8 6.1 9.1

68 5.8 5.4 8.2





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 9.251 ng

RT: 7.26 min Scan# 879

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Tgt Ion:105 Resp: 111362



Ion Ratio Lower Upper

105 100

71 7.5 5.4 8.0

51 29.2 24.6 36.8

120 21.8 17.0 25.6





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E








#23
Nitrobenzene-d5
Concen: 82.211 ng
RT: 7.41 min Scan# 905
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampled :

Tot Ion	82	Resp	781768
Ion Ratio	100		
Lower	36.6		
Upper	55.0		
128	43.1		
54	46.0		



#24
Nitrobenzene
Concen: 8.863 ng
RT: 7.43 min Scan# 908
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion	77	Resp	91001
Ion Ratio	100		
Lower	35.5		
Upper	53.3		
123	42.3		
65	13.6		









#25
Isophorone
Concen: 8.226 ng
RT: 7.66 min Scan# 948
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :


Tot Ion: 82 Resp: 149572
Ion Ratio Lower Upper
82 100
95 6.5 5.5 8.3
138 15.7 13.9 20.9



#26
2-Nitrophenol
Concen: 7.591 ng
RT: 7.75 min Scan# 963
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion: 139 Resp: 36739
Ion Ratio Lower Upper
139 100
109 28.3 23.6 35.4
65 51.3 40.6 60.8









#27
2,4-Dimethylphenol
Concen: 8.926 ng
RT: 7.78 min Scan# 968
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion:122 Resp: 64997
Ion Ratio Lower Upper
122 100
107 117.5 91.7 137.5
121 57.3 46.6 70.0




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E





#28
bis(2-Chloroethoxy)methane
Concen: 8.008 ng
RT: 7.87 min Scan# 984
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion: 93 Resp: 90256
Ion Ratio Lower Upper
93 100
95 32.4 26.0 39.0
123 11.3 9.5 14.3



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 123.00 (122.70 to 123.70): E









#29
 2,4-Dichlorophenol
 Concen: 8.068 ng
 RT: 7.99 min Scan# 1004
 Delta R.T. 0.00 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:162 Resp: 62028
 Ion Ratio Lower Upper
 162 100
 164 63.0 43.2 83.2
 98 36.7 17.3 57.3




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 8.024 ng
 RT: 8.07 min Scan# 1018
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:180 Resp: 70329
 Ion Ratio Lower Upper
 180 100
 182 92.7 76.6 114.8
 145 29.6 23.2 34.8



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 8.912 ng

RT: 8.15 min Scan# 1031

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:128 Resp: 230192

Ion Ratio Lower Upper

128 100

129 11.3 9.2 13.8




127 13.2 11.0 16.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 1.177 ng

RT: 7.87 min Scan# 983

Delta R.T. -0.05 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:122 Resp: 6043

Ion Ratio Lower Upper

122 100

105 114.0 103.1 143.1



77 84.1 76.8 116.8




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 8.038 ng
RT: 8.20 min Scan# 1039
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampled :

Tot Ion:127 Resp: 92766
Ion Ratio Lower Upper
127 100
129 31.7 26.0 39.0
65 32.1 23.9 35.9
92 20.2 14.9 22.3




Abundance Ion 127.00 (126.70 to 127.70): E
120000
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1





#34
Hexachlorobutadiene
Concen: 7.837 ng
RT: 8.27 min Scan# 1052
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:225 Resp: 39564
Ion Ratio Lower Upper
225 100
223 60.1 50.4 75.6
227 62.4 51.6 77.4



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 7.528 ng
 RT: 8.54 min Scan# 1097
 Delta R.T. -0.05 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:113 Resp: 18719
 Ion Ratio Lower Upper
 113 100
 55 163.1 138.1 178.1
 56 127.2 100.6 140.6




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF1
 Ion 56.00 (55.70 to 56.70): BF1






#36
 4-Chloro-3-methylphenol
 Concen: 8.142 ng
 RT: 8.69 min Scan# 1122
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:107 Resp: 65124
 Ion Ratio Lower Upper
 107 100
 144 25.5 21.4 32.0
 142 79.7 65.5 98.3



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 9.022 ng

RT: 8.85 min Scan# 1149

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:142 Resp: 153478

Ion Ratio Lower Upper

142 100

141 87.4 71.4 107.2




115 36.4 28.9 43.3



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

1-Methylnaphthalene

Concen: 8.968 ng

RT: 8.95 min Scan# 1166

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:142 Resp: 144326

Ion Ratio Lower Upper

142 100

141 91.5 73.8 110.6



116 3.9 3.1 4.7




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.89 min Scan# 1327

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampled :


Tot Ion:164 Resp: 289704

Ion Ratio Lower Upper

164 100

162 100.1 80.8 121.2

160 46.5 36.2 54.4






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 8.599 ng

RT: 9.02 min Scan# 1178

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Tgt Ion:216 Resp: 66601


Ion Ratio Lower Upper

216 100

214 78.8 62.8 94.2

179 21.6 17.6 26.4

108 20.5 16.1 24.1





Abundance

Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 4.000 ng

RT: 9.00 min Scan# 1176

Delta R.T. -0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:237 Resp: 12110

Ion Ratio Lower Upper




237 100

235 60.3 42.5 82.5

272 12.2 0.0 33.3



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 98.959 ng

RT: 10.68 min Scan# 1461

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:330 Resp: 277953

Ion Ratio Lower Upper



330 100


332 94.0 77.7 116.5

141 34.0 28.1 42.1



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E








#43
 2,4,6-Trichlorophenol
 Concen: 7.210 ng
 RT: 9.13 min Scan# 1197
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:196 Resp: 38438
 Ion Ratio Lower Upper
 196 100
 198 95.1 79.1 118.7
 200 31.6 26.0 39.0




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#44
 2,4,5-Trichlorophenol
 Concen: 7.382 ng
 RT: 9.17 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:196 Resp: 40678
 Ion Ratio Lower Upper
 196 100
 198 100.0 80.1 120.1
 97 51.9 38.9 58.3
 132 29.0 22.6 33.8



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E









#45
2-Fluorobiphenyl
Concen: 83.625 ng
RT: 9.22 min Scan# 1212
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion:172 Resp: 1293416
Ion Ratio Lower Upper
172 100
171 36.4 29.1 43.7
170 25.3 20.1 30.1




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 9.162 ng
RT: 9.31 min Scan# 1228
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:154 Resp: 187395
Ion Ratio Lower Upper
154 100
153 39.8 20.9 60.9
76 13.2 0.0 34.3



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1









#47
2-Chloronaphthalene
Concen: 8.251 ng
RT: 9.34 min Scan# 1233
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
162	100		
127	39.9	31.3	46.9
164	33.8	26.6	40.0




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 127.00 (126.70 to 127.70): E
Ion 164.00 (163.70 to 164.70): E






#48
2-Nitroaniline
Concen: 7.508 ng
RT: 9.43 min Scan# 1248
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion	Ratio	Lower	Upper
65	100		
92	64.1	57.1	85.7
138	96.5	97.1	145.7#



Abundance Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1
Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 9.147 ng

RT: 9.75 min Scan# 1303

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:152 Resp: 227175

Ion Ratio Lower Upper

152 100

151 20.3 16.8 25.2




153 13.7 11.6 17.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 7.895 ng

RT: 9.60 min Scan# 1278

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:163 Resp: 155742

Ion Ratio Lower Upper

163 100

194 3.7 3.0 4.6



164 9.9 8.4 12.6




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 7.885 ng
 RT: 9.67 min Scan# 1289
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:165 Resp: 33800
 Ion Ratio Lower Upper
 165 100
 63 54.8 39.4 59.2
 89 57.9 43.0 64.6




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1






#52
 Acenaphthene
 Concen: 8.881 ng
 RT: 9.92 min Scan# 1332
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:154 Resp: 135304
 Ion Ratio Lower Upper
 154 100
 153 113.1 90.1 135.1
 152 51.7 42.6 63.8



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 7.102 ng
RT: 9.84 min Scan# 1318
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampled :

Tot Ion:138 Resp: 37620
Ion Ratio Lower Upper
138 100
108 10.3 7.4 11.2
92 107.0 83.9 125.9




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 2.404 ng
RT: 9.96 min Scan# 1339
Delta R.T. 0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:184 Resp: 4376
Ion Ratio Lower Upper
184 100
63 84.0 56.7 85.1
154 68.3 51.9 77.9



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E









#55
Dibenzenofuran
Concen: 9.080 ng
RT: 10.10 min Scan# 1362
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion:168 Resp: 201174
Ion Ratio Lower Upper
168 100
139 37.4 30.2 45.2
169 12.7 10.8 16.2




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#56
4-Nitrophenol
Concen: 4.839 ng
RT: 10.02 min Scan# 1348
Delta R.T. 0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:139 Resp: 16418
Ion Ratio Lower Upper
139 100
109 75.7 52.0 92.0
65 108.3 72.6 112.6



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1





#57

2,4-Dinitrotoluene

Concen: 7.431 ng

RT: 10.07 min Scan# 1358

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :

Tot Ion:165 Resp: 42413



Ion Ratio Lower Upper

165 100

63 52.2 33.2 49.8#

89 85.8 57.3 85.9

182 2.3 2.1 3.1





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#58

Fluorene

Concen: 9.188 ng

RT: 10.44 min Scan# 1420

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19



Tgt Ion:166 Resp: 156632

Ion Ratio Lower Upper

166 100

165 96.7 78.6 118.0


167 13.3 10.9 16.3



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
2,3,4,6-Tetrachlorophenol
Concen: 7.729 ng
RT: 10.22 min Scan# 1382
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampled :

Tot Ion:232 Resp: 35835
Ion Ratio Lower Upper
232 100
131 46.3 37.6 56.4
130 2.2 1.8 2.8
166 29.5 24.6 37.0




Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E






#60
Diethylphthalate
Concen: 8.009 ng
RT: 10.30 min Scan# 1397
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:149 Resp: 147490
Ion Ratio Lower Upper
149 100
177 21.6 18.4 27.6
150 12.9 10.1 15.1



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E





#61

4-Chlorophenyl-phenylether

Concen: 8.618 ng

RT: 10.43 min Scan# 1418

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampled :


Tot Ion:204 Resp: 74406

Ion Ratio Lower Upper




204 100

206 32.3 26.4 39.6

141 61.1 47.0 70.4



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E



#62

4-Nitroaniline

Concen: 7.239 ng

RT: 10.45 min Scan# 1422

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:138 Resp: 39623

Ion Ratio Lower Upper



138 100


92 48.9 31.0 71.0

108 75.4 56.7 96.7



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF1
Ion 108.00 (107.70 to 108.70): E









#63
Azobenzene
Concen: 8.498 ng
RT: 10.59 min Scan# 1446
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :


Tot Ion:	77	Resp:	160967
Ion	Ratio	Lower	Upper
77	100		
182	25.6	2.6	42.6
105	21.7	1.6	41.6
51	27.2	7.4	47.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.38 min Scan# 1580
Delta R.T. 0.00 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:	188	Resp:	498722
Ion	Ratio	Lower	Upper
188	100		
94	8.6	8.0	12.0
80	8.9	8.6	12.8









#65
 4,6-Dinitro-2-methylphenol
 Concen: 4.963 ng
 RT: 10.49 min Scan# 1429
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:198 Resp: 13116
 Ion Ratio Lower Upper
 198 100
 51 41.8 21.1 61.1
 105 47.9 24.5 64.5



#66
 n-Nitrosodiphenylamine
 Concen: 8.808 ng
 RT: 10.55 min Scan# 1438
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:169 Resp: 132217
 Ion Ratio Lower Upper
 169 100
 168 67.8 53.0 79.6
 167 36.7 29.9 44.9









#67
 4-Bromophenyl-phenylether
 Concen: 8.138 ng
 RT: 10.92 min Scan# 1502
 Delta R.T. 0.00 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:248 Resp: 43448
 Ion Ratio Lower Upper
 248 100
 250 97.2 77.8 116.6
 141 67.2 64.3 96.5




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 7.808 ng
 RT: 10.99 min Scan# 1514
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:284 Resp: 48202
 Ion Ratio Lower Upper
 284 100
 142 36.0 28.2 42.2
 249 30.8 24.0 36.0



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E









#69
 Atrazine
 Concen: 7.830 ng
 RT: 11.07 min Scan# 1527
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
200	100		
173	25.7	7.3	47.3
215	45.4	26.9	66.9



#70
 Pentachlorophenol
 Concen: 4.646 ng
 RT: 11.19 min Scan# 1548
 Delta R.T. 0.00 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.7	51.6	77.4
264	63.2	49.4	74.0





#71

Phenanthrene

Concen: 8.928 ng

RT: 11.40 min Scan# 1584

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:178 Resp: 227614

Ion Ratio Lower Upper

178 100

176 19.6 16.5 24.7




179 14.7 12.6 18.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 9.018 ng

RT: 11.45 min Scan# 1592

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:178 Resp: 232695

Ion Ratio Lower Upper

178 100

176 19.4 15.7 23.5



179 15.9 12.4 18.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 8.687 ng
 RT: 11.61 min Scan# 1619
 Delta R.T. 0.00 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:167 Resp: 203359
 Ion Ratio Lower Upper
 167 100
 166 21.0 18.0 27.0
 139 13.0 11.4 17.2



#74
 Di-n-butylphthalate
 Concen: 7.676 ng
 RT: 11.93 min Scan# 1674
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion:149 Resp: 209605
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.7 11.5
 104 5.3 4.3 6.5









#75
 Fluoranthene
 Concen: 8.792 ng
 RT: 12.59 min Scan# 1786
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19




Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Resp	Lower	Upper
202	100	234666		
101	10.0		0.0	31.5
203	17.2		0.0	37.9



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.02 min Scan# 2028
 Delta R.T. -0.01 min
 Lab File: BF115866.D
 Acq: 31 Jul 2019 15:19

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	382209		
120	12.5		8.5	12.7
236	26.6		22.2	33.2





#77

Benzidine

Concen: 8.050 ng

RT: 12.71 min Scan# 1806

Delta R.T. 0.00 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampled :


Tot Ion:184 Resp: 103945

Ion Ratio Lower Upper

184 100

185 17.1 13.4 20.2




183 10.9 9.1 13.7



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 8.401 ng

RT: 12.82 min Scan# 1825

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:202 Resp: 242041

Ion Ratio Lower Upper

202 100

200 21.8 18.2 27.2



203 17.0 14.3 21.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 75.607 ng

RT: 12.96 min Scan# 1849

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Instrument :

BNA_F

ClientSampleId :

Tot Ion:244 Resp: 1371397




Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.6	8.4
122	11.8	8.6	13.0



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 6.533 ng

RT: 13.43 min Scan# 1929


Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Tgt Ion:149 Resp: 79624



Ion	Ratio	Lower	Upper
149	100		
91	71.7	57.0	85.6
206	21.0	16.9	25.3




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 8.071 ng

RT: 14.01 min Scan# 2027

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :





Tot Ion:228 Resp: 197168

Ion Ratio Lower Upper

228 100

226 26.6 22.5 33.7

229 19.5 16.1 24.1



#82

3,3'-Dichlorobenzidine

Concen: 7.752 ng

RT: 13.97 min Scan# 2020

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19




Tgt Ion:252 Resp: 65789


Ion Ratio Lower Upper

252 100

254 64.7 51.2 76.8

126 10.9 8.1 12.1





#83

Chrysene

Concen: 8.213 ng

RT: 14.04 min Scan# 2033

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampled :


Tot Ion:228 Resp: 194792

Ion Ratio Lower Upper

228 100

226 30.1 24.8 37.2

229 19.4 16.4 24.6






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 6.747 ng

RT: 13.99 min Scan# 2024

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:149 Resp: 106848

Ion Ratio Lower Upper

149 100

167 26.6 22.4 33.6

279 3.4 3.3 4.9





Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 6.310 ng

RT: 14.60 min Scan# 2128

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19

Instrument :

BNA_F

ClientSampleId :


Tot Ion:149 Resp: 158730

Ion Ratio Lower Upper

149 100

167 1.4 1.2 1.8

43 9.6 7.4 11.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1

250000

200000

150000



100000

50000

0

14.55 14.60 14.65 14.70

Time-->



#86

Indeno(1,2,3-cd)pyrene

Concen: 6.895 ng

RT: 16.96 min Scan# 2528

Delta R.T. -0.01 min

Lab File: BF115866.D

Acq: 31 Jul 2019 15:19


Tgt Ion:276 Resp: 137356

Ion Ratio Lower Upper

276 100

138 24.4 19.6 29.4

277 24.8 20.1 30.1



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

60000


40000


20000

0

16.90 17.00 17.10

Time-->









#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.49 min Scan# 2278
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion:264 Resp: 352151




Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.4	29.2
265	21.7	17.5	26.3



#88
Benzo(b)fluoranthene
Concen: 7.518 ng
RT: 15.06 min Scan# 2205
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion:252 Resp: 153080

Ion	Ratio	Lower	Upper
252	100		
253	22.4	17.2	25.8
125	9.4	7.4	11.0









#89
Benzo(k)fluoranthene
Concen: 8.496 ng
RT: 15.09 min Scan# 2210
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19




Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.0	18.2	27.4
125	9.3	8.9	13.3



#90
Benzo(a)pyrene
Concen: 7.961 ng
RT: 15.42 min Scan# 2267
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.8	26.6
125	11.2	8.8	13.2









#91
Dibenzo(a,h)anthracene
Concen: 7.300 ng
RT: 16.97 min Scan# 2530
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
278	100		
139	15.4	12.6	19.0
279	23.8	18.8	28.2



#92
Benzo(g,h,i)perylene
Concen: 7.249 ng
RT: 17.40 min Scan# 2603
Delta R.T. -0.01 min
Lab File: BF115866.D
Acq: 31 Jul 2019 15:19

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.5	29.3
138	22.4	17.8	26.8

