

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080219\
 Data File : BF115946.D
 Acq On : 2 Aug 2019 15:05
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 15:30:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.604	0.598	1.0	115	0.00
3	Pyridine	1.686	1.563	7.3	106	0.00
4	n-Nitrosodimethylamine	0.665	0.664	0.2	114	0.00
5 S	2-Fluorophenol	1.195	1.206	-0.9	113	0.00
6	Aniline	2.159	2.132	1.3	110	0.00
7 S	Phenol-d6	1.507	1.549	-2.8	116	0.00
8	2-Chlorophenol	1.321	1.399	-5.9	119	0.00
9	Benzaldehyde	1.040	0.954	8.3	103	0.00
10 C	Phenol	1.775	1.810	-2.0	115	0.00
11	bis(2-Chloroethyl)ether	1.305	1.363	-4.4	118	0.00
12	1,3-Dichlorobenzene	1.527	1.536	-0.6	113	0.01
13 C	1,4-Dichlorobenzene	1.529	1.538	-0.6	113	0.01
14	1,2-Dichlorobenzene	1.408	1.415	-0.5	110	0.00
15	Benzyl Alcohol	1.144	1.206	-5.4	116	0.00
16	2,2'-oxybis(1-Chloropropane	1.780	1.832	-2.9	114	0.00
17	2-Methylphenol	1.119	1.162	-3.8	119	0.00
18	Hexachloroethane	0.563	0.576	-2.3	114	0.01
19 P	n-Nitroso-di-n-propylamine	0.934	0.955	-2.2	117	0.00
20	3+4-Methylphenols	1.324	1.343	-1.4	112	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	117	0.01
22	Acetophenone	0.439	0.428	2.5	113	0.00
23 S	Nitrobenzene-d5	0.346	0.338	2.3	114	0.00
24	Nitrobenzene	0.374	0.373	0.3	116	0.00
25	Isophorone	0.663	0.668	-0.8	118	0.00
26 C	2-Nitrophenol	0.176	0.185	-5.1	121	0.00
27	2,4-Dimethylphenol	0.265	0.262	1.1	115	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.415	-1.0	118	0.00
29 C	2,4-Dichlorophenol	0.280	0.284	-1.4	116	0.00
30	1,2,4-Trichlorobenzene	0.319	0.315	1.3	114	0.01
31	Naphthalene	0.941	0.935	0.6	114	0.00
32	Benzoic acid	0.187	0.158	15.5	108	0.00
33	4-Chloroaniline	0.421	0.418	0.7	115	0.00
34 C	Hexachlorobutadiene	0.184	0.183	0.5	115	0.00
35	Caprolactam	0.091	0.094	-3.3	120	0.00
36 C	4-Chloro-3-methylphenol	0.291	0.300	-3.1	120	0.01
37	2-Methylnaphthalene	0.620	0.623	-0.5	115	0.01
38	1-Methylnaphthalene	0.586	0.592	-1.0	116	0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.01
40	1,2,4,5-Tetrachlorobenzene	0.535	0.557	-4.1	115	0.01
41 P	Hexachlorocyclopentadiene	0.209	0.186	11.0	97	0.02
42 S	2,4,6-Tribromophenol	0.194	0.208	-7.2	119	0.02
43 C	2,4,6-Trichlorophenol	0.368	0.378	-2.7	114	0.01
44	2,4,5-Trichlorophenol	0.380	0.401	-5.5	117	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080219\
 Data File : BF115946.D
 Acq On : 2 Aug 2019 15:05
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 02 15:30:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.068	1.094	-2.4	112	0.01
46	1,1'-Biphenyl	1.412	1.463	-3.6	114	0.01
47	2-Chloronaphthalene	1.175	1.201	-2.2	113	0.01
48	2-Nitroaniline	0.388	0.410	-5.7	118	0.00
49	Acenaphthylene	1.715	1.824	-6.4	117	0.02
50	Dimethylphthalate	1.362	1.470	-7.9	120	0.02
51	2,6-Dinitrotoluene	0.296	0.326	-10.1	122	0.01
52 C	Acenaphthene	1.052	1.095	-4.1	114	0.01
53	3-Nitroaniline	0.366	0.381	-4.1	116	0.01
54 P	2,4-Dinitrophenol	0.126	0.129	-2.4	112	0.01
55	Dibenzofuran	1.530	1.604	-4.8	114	0.01
56 P	4-Nitrophenol	0.234	0.235	-0.4	111	0.01
57	2,4-Dinitrotoluene	0.394	0.424	-7.6	117	0.01
58	Fluorene	1.177	1.260	-7.1	117	0.02
59	2,3,4,6-Tetrachlorophenol	0.320	0.337	-5.3	117	0.02
60	Diethylphthalate	1.271	1.377	-8.3	119	0.01
61	4-Chlorophenyl-phenylether	0.596	0.632	-6.0	115	0.01
62	4-Nitroaniline	0.378	0.394	-4.2	115	0.01
63	Azobenzene	1.308	1.399	-7.0	118	0.02
64 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.02
65	4,6-Dinitro-2-methylphenol	0.106	0.106	0.0	120	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.587	2.5	119	0.01
67	4-Bromophenyl-phenylether	0.214	0.211	1.4	119	0.02
68	Hexachlorobenzene	0.248	0.240	3.2	117	0.02
69	Atrazine	0.198	0.165	16.7	102	0.01
70 C	Pentachlorophenol	0.120	0.109	9.2	108	0.02
71	Phenanthrene	1.022	0.970	5.1	115	0.01
72	Anthracene	1.035	1.001	3.3	118	0.02
73	Carbazole	0.939	0.926	1.4	119	0.02
74	Di-n-butylphthalate	1.095	1.139	-4.0	123	0.02
75 C	Fluoranthene	1.070	1.054	1.5	120	0.01
76 I	Chrysene-d12	1.000	1.000	0.0	116	0.02
77	Benzidine	0.676	0.675	0.1	108	0.02
78	Pyrene	1.508	1.514	-0.4	119	0.01
79 S	Terphenyl-d14	0.949	0.945	0.4	118	0.02
80	Butylbenzylphthalate	0.638	0.702	-10.0	126	0.02
81	Benzo(a)anthracene	1.278	1.280	-0.2	116	0.02
82	3,3'-Dichlorobenzidine	0.444	0.453	-2.0	115	0.02
83	Chrysene	1.241	1.238	0.2	115	0.02
84	Bis(2-ethylhexyl)phthalate	0.829	0.971	-17.1	133	0.02
85 c	Di-n-octyl phthalate	1.316	1.501	-14.1	128	0.02
86	Indeno(1,2,3-cd)pyrene	1.042	1.039	0.3	121	0.04
87 I	Perylene-d12	1.000	1.000	0.0	114	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080219\
Data File : BF115946.D
Acq On : 2 Aug 2019 15:05
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 02 15:30:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:31:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.156	1.147	0.8	108	0.02
89	Benzo(k)fluoranthene	1.126	1.124	0.2	118	0.02
90 C	Benzo(a)pyrene	1.034	1.031	0.3	113	0.02
91	Dibenzo(a,h)anthracene	0.895	0.924	-3.2	121	0.04
92	Benzo(q,h,i)perylene	0.879	0.917	-4.3	126	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0