



2,4-Dinitrophenol

Response Ratio



Resp Ratio = $1.85 \times 10^{-1} \times \text{Amt} - 6.92 \times 10^{-2}$
Coef of Det (r^2) = 0.991 Curve Fit: Linear


Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081820.M
Calibration Table Last Updated: Wed Aug 19 12:14:25 2020



Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081820.M
Calibration Table Last Updated: Wed Aug 19 12:14:25 2020

Hexachlorocyclopentadiene

Response Ratio



Resp Ratio = $2.48 \times 10^{-1} \times \text{Amt} - 1.15 \times 10^{-1}$
Coef of Det (r^2) = 0.994 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081820.M
Calibration Table Last Updated: Wed Aug 19 12:14:25 2020