


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082020\
 Data File : BF121334.D
 Acq On : 20 Aug 2020 7:52
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF081920.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 20 04:56:10 2020



AutoFind: Scans 1594, 1595, 1596; Background Corrected with Scan 1589

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.6	41768	PASS
68	69	0.00	2	1.6	613	PASS
69	198	0.00	100	34.3	39114	PASS
70	69	0.00	2	0.3	120	PASS
127	198	10	80	47.9	54645	PASS
197	198	0.00	2	0.7	745	PASS
198	198	100	100	100.0	114088	PASS
199	198	5	9	6.6	7579	PASS
275	198	10	60	27.6	31522	PASS
365	198	1	100	3.5	4046	PASS
441	198	0.01	100	13.0	14801	PASS
442	198	50	100	82.4	94002	PASS
443	442	15	24	18.8	17643	PASS