

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110920\
 Data File : BF122286.D
 Acq On : 9 Nov 2020 15:58
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF110920

Quant Time: Nov 09 16:45:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 09 15:51:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.597	0.577	3.4	106	0.00
3	Pyridine	1.643	1.614	1.8	106	0.00
4	n-Nitrosodimethylamine	0.774	0.754	2.6	105	0.00
5 S	2-Fluorophenol	1.303	1.288	1.2	106	0.00
6	Aniline	2.236	2.193	1.9	106	0.00
7 S	Phenol-d6	1.704	1.665	2.3	106	0.00
8	2-Chlorophenol	1.351	1.354	-0.2	106	0.00
9	Benzaldehyde	0.961	0.875	8.9	105	0.00
10 C	Phenol	1.678	1.656	1.3	106	0.00
11	bis(2-Chloroethyl)ether	1.334	1.329	0.4	108	0.00
12	1,3-Dichlorobenzene	1.533	1.503	2.0	105	0.00
13 C	1,4-Dichlorobenzene	1.540	1.515	1.6	107	0.00
14	1,2-Dichlorobenzene	1.438	1.423	1.0	106	0.00
15	Benzyl Alcohol	1.211	1.215	-0.3	106	0.00
16	2,2'-oxybis(1-Chloropropane	1.944	1.955	-0.6	109	0.00
17	2-Methylphenol	1.127	1.121	0.5	108	0.00
18	Hexachloroethane	0.536	0.548	-2.2	109	0.00
19 P	n-Nitroso-di-n-propylamine	1.018	1.018	0.0	109	0.00
20	3+4-Methylphenols	1.387	1.391	-0.3	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
22	Acetophenone	0.521	0.506	2.9	107	0.00
23 S	Nitrobenzene-d5	0.327	0.365	-11.6	114	0.00
24	Nitrobenzene	0.374	0.394	-5.3	110	0.00
25	Isophorone	0.728	0.717	1.5	109	0.00
26 C	2-Nitrophenol	0.126	0.162	-28.6#	123	0.00
27	2,4-Dimethylphenol	0.344	0.347	-0.9	110	0.00
28	bis(2-Chloroethoxy)methane	0.445	0.442	0.7	109	0.00
29 C	2,4-Dichlorophenol	0.304	0.310	-2.0	109	0.00
30	1,2,4-Trichlorobenzene	0.336	0.332	1.2	106	0.00
31	Naphthalene	1.063	1.037	2.4	107	0.00
32	Benzoic acid	0.155	0.209	-34.8#	132	0.00
33	4-Chloroaniline	0.463	0.455	1.7	107	0.00
34 C	Hexachlorobutadiene	0.193	0.193	0.0	109	0.00
35	Caprolactam	0.096	0.097	-1.0	108	0.00
36 C	4-Chloro-3-methylphenol	0.317	0.318	-0.3	108	0.00
37	2-Methylnaphthalene	0.702	0.683	2.7	107	0.00
38	1-Methylnaphthalene	0.658	0.644	2.1	107	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.616	0.611	0.8	109	0.00
41 P	Hexachlorocyclopentadiene	0.360	0.392	-8.9	113	0.00
42 S	2,4,6-Tribromophenol	0.215	0.227	-5.6	111	0.00
43 C	2,4,6-Trichlorophenol	0.400	0.418	-4.5	111	0.00
44	2,4,5-Trichlorophenol	0.420	0.433	-3.1	110	0.00
45 S	2-Fluorobiphenyl	1.280	1.198	6.4	106	0.00
46	1,1'-Biphenyl	1.592	1.546	2.9	108	0.00
47	2-Chloronaphthalene	1.264	1.229	2.8	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110920\
 Data File : BF122286.D
 Acq On : 9 Nov 2020 15:58
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICSVBF110920

Quant Time: Nov 09 16:45:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 09 15:51:40 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.310	0.370	-19.4	117	0.00
49	Acenaphthylene	1.942	1.889	2.7	107	0.00
50	Dimethylphthalate	1.422	1.389	2.3	110	0.00
51	2,6-Dinitrotoluene	0.275	0.315	-14.5	116	0.00
52 C	Acenaphthene	1.202	1.183	1.6	109	0.00
53	3-Nitroaniline	0.317	0.351	-10.7	113	0.00
54 P	2,4-Dinitrophenol	0.084	0.102	-21.4	140	0.00
55	Dibenzofuran	1.761	1.718	2.4	108	0.00
56 P	4-Nitrophenol	0.271	0.287	-5.9	114	0.00
57	2,4-Dinitrotoluene	0.332	0.403	-21.4	118	0.00
58	Fluorene	1.349	1.279	5.2	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.349	0.372	-6.6	112	0.00
60	Diethylphthalate	1.395	1.408	-0.9	112	0.00
61	4-Chlorophenyl-phenylether	0.637	0.629	1.3	110	0.00
62	4-Nitroaniline	0.313	0.348	-11.2	114	0.00
63	Azobenzene	1.463	1.431	2.2	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.076	0.091	-19.7	133	0.00
66 c	n-Nitrosodiphenylamine	0.681	0.660	3.1	108	0.00
67	4-Bromophenyl-phenylether	0.235	0.236	-0.4	111	0.00
68	Hexachlorobenzene	0.254	0.251	1.2	109	0.00
69	Atrazine	0.170	0.171	-0.6	110	0.00
70 C	Pentachlorophenol	0.146	0.158	-8.2	117	0.00
71	Phenanthrene	1.135	1.099	3.2	108	0.00
72	Anthracene	1.170	1.142	2.4	108	0.00
73	Carbazole	1.064	1.037	2.5	109	0.00
74	Di-n-butylphthalate	1.134	1.164	-2.6	113	0.00
75 C	Fluoranthene	1.185	1.159	2.2	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.555	0.552	0.5	107	0.00
78	Pyrene	1.405	1.374	2.2	108	0.00
79 S	Terphenyl-d14	0.944	0.904	4.2	110	0.00
80	Butylbenzylphthalate	0.505	0.563	-11.5	116	0.00
81	Benzo(a)anthracene	1.244	1.212	2.6	108	0.00
82	3,3'-Dichlorobenzidine	0.464	0.486	-4.7	112	0.00
83	Chrysene	1.253	1.218	2.8	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.677	0.719	-6.2	116	0.00
85 c	Di-n-octyl phthalate	1.148	1.313	-14.4	118	0.00
86 I	Perylene-d12	1.000	1.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	1.200	1.226	-2.2	123	0.00
88	Benzo(b)fluoranthene	1.295	1.274	1.6	111	0.00
89	Benzo(k)fluoranthene	1.263	1.233	2.4	115	0.00
90 C	Benzo(a)pyrene	1.129	1.121	0.7	114	0.00
91	Dibenzo(a,h)anthracene	1.005	1.019	-1.4	122	0.00
92	Benzo(g,h,i)perylene	0.978	0.971	0.7	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110920\
Data File : BF122286.D
Acq On : 9 Nov 2020 15:58
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF110920

Quant Time: Nov 09 16:45:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 09 15:51:40 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1