

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF112618\
 Data File : BF110961.D
 Acq On : 26 Nov 2018 14:00
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 15:29:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF112618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 26 15:26:09 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.578	0.582	-0.7	99	0.00
3	Pyridine	1.265	1.238	2.1	94	0.00
4	n-Nitrosodimethylamine	0.579	0.585	-1.0	101	0.00
5 S	2-Fluorophenol	1.206	1.249	-3.6	98	0.00
6	Aniline	1.943	1.989	-2.4	97	0.00
7 S	Phenol-d6	1.572	1.583	-0.7	99	0.00
8	2-Chlorophenol	1.429	1.447	-1.3	99	0.00
9	Benzaldehyde	0.839	0.864	-3.0	101	0.00
10 C	Phenol	1.770	1.826	-3.2	98	0.00
11	bis(2-Chloroethyl)ether	1.333	1.343	-0.8	101	0.00
12	1,3-Dichlorobenzene	1.570	1.581	-0.7	99	0.00
13 C	1,4-Dichlorobenzene	1.615	1.612	0.2	98	0.00
14	1,2-Dichlorobenzene	1.520	1.510	0.7	98	0.00
15	Benzyl Alcohol	1.208	1.248	-3.3	99	0.00
16	2,2'-oxybis(1-Chloropropane	2.122	2.144	-1.0	100	0.00
17	2-Methylphenol	1.136	1.147	-1.0	99	0.00
18	Hexachloroethane	0.594	0.605	-1.9	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.024	1.052	-2.7	101	0.00
20	3+4-Methylphenols	1.511	1.522	-0.7	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.469	0.472	-0.6	99	0.00
23 S	Nitrobenzene-d5	0.349	0.352	-0.9	99	0.00
24	Nitrobenzene	0.363	0.359	1.1	99	0.00
25	Isophorone	0.573	0.576	-0.5	99	0.00
26 C	2-Nitrophenol	0.177	0.185	-4.5	100	0.00
27	2,4-Dimethylphenol	0.267	0.273	-2.2	99	0.00
28	bis(2-Chloroethoxy)methane	0.384	0.389	-1.3	100	0.00
29 C	2,4-Dichlorophenol	0.279	0.287	-2.9	98	0.00
30	1,2,4-Trichlorobenzene	0.313	0.317	-1.3	99	0.00
31	Naphthalene	0.938	0.930	0.9	99	0.00
32	Benzoic acid	0.198	0.201	-1.5	98	0.00
33	4-Chloroaniline	0.399	0.409	-2.5	100	0.00
34 C	Hexachlorobutadiene	0.178	0.179	-0.6	99	0.00
35	Caprolactam	0.095	0.097	-2.1	100	0.00
36 C	4-Chloro-3-methylphenol	0.310	0.314	-1.3	98	0.00
37	2-Methylnaphthalene	0.619	0.618	0.2	99	0.00
38	1-Methylnaphthalene	0.602	0.602	0.0	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.629	0.631	-0.3	100	0.00
41 P	Hexachlorocyclopentadiene	0.063	0.062	1.6	100	0.00
42 S	2,4,6-Tribromophenol	0.199	0.204	-2.5	102	0.00
43 C	2,4,6-Trichlorophenol	0.376	0.389	-3.5	100	0.00
44	2,4,5-Trichlorophenol	0.396	0.409	-3.3	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF112618\
 Data File : BF110961.D
 Acq On : 26 Nov 2018 14:00
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 26 15:29:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF112618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 26 15:26:09 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.380	1.357	1.7	101	0.00
46	1,1'-Biphenyl	1.844	1.834	0.5	99	0.00
47	2-Chloronaphthalene	1.337	1.320	1.3	99	0.00
48	2-Nitroaniline	0.418	0.424	-1.4	100	0.00
49	Acenaphthylene	1.901	1.893	0.4	100	0.00
50	Dimethylphthalate	1.467	1.462	0.3	101	0.00
51	2,6-Dinitrotoluene	0.327	0.331	-1.2	100	0.00
52 C	Acenaphthene	1.219	1.207	1.0	100	0.00
53	3-Nitroaniline	0.381	0.387	-1.6	100	0.00
54 P	2,4-Dinitrophenol	0.104	0.114	-9.6	98	0.00
55	Dibenzofuran	1.778	1.737	2.3	100	0.00
56 P	4-Nitrophenol	0.190	0.198	-4.2	97	0.00
57	2,4-Dinitrotoluene	0.424	0.428	-0.9	100	0.00
58	Fluorene	1.383	1.358	1.8	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.325	0.328	-0.9	101	0.00
60	Diethylphthalate	1.513	1.474	2.6	100	0.00
61	4-Chlorophenyl-phenylether	0.724	0.724	0.0	101	0.00
62	4-Nitroaniline	0.354	0.361	-2.0	100	0.00
63	Azobenzene	1.474	1.450	1.6	99	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.095	0.103	-8.4	105	0.00
66 c	n-Nitrosodiphenylamine	0.666	0.685	-2.9	101	0.00
67	4-Bromophenyl-phenylether	0.218	0.225	-3.2	100	0.00
68	Hexachlorobenzene	0.232	0.234	-0.9	99	0.00
69	Atrazine	0.171	0.196	-14.6	100	0.00
70 C	Pentachlorophenol	0.092	0.101	-9.8	101	0.00
71	Phenanthrene	1.059	1.063	-0.4	100	0.00
72	Anthracene	1.079	1.082	-0.3	99	0.00
73	Carbazole	1.047	1.061	-1.3	100	0.00
74	Di-n-butylphthalate	1.224	1.231	-0.6	99	0.00
75 C	Fluoranthene	1.089	1.088	0.1	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
77	Benzidine	0.512	0.545	-6.4	116	0.00
78	Pyrene	1.461	1.486	-1.7	100	0.00
79 S	Terphenyl-d14	1.027	1.018	0.9	99	0.00
80	Butylbenzylphthalate	0.659	0.670	-1.7	98	0.00
81	Benzo(a)anthracene	1.181	1.197	-1.4	98	0.00
82	3,3'-Dichlorobenzidine	0.405	0.412	-1.7	98	0.00
83	Chrysene	1.162	1.147	1.3	97	0.00
84	Bis(2-ethylhexyl)phthalate	0.879	0.909	-3.4	100	0.00
85 c	Di-n-octyl phthalate	1.387	1.420	-2.4	99	0.00
86	Indeno(1,2,3-cd)pyrene	0.834	0.809	3.0	100	0.00
87 I	Perylene-d12	1.000	1.000	0.0	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF112618\
Data File : BF110961.D
Acq On : 26 Nov 2018 14:00
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 26 15:29:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF112618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 26 15:26:09 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.175	1.190	-1.3	96	0.00
89	Benzo(k)fluoranthene	1.209	1.246	-3.1	98	0.00
90 C	Benzo(a)pyrene	1.094	1.099	-0.5	98	0.00
91	Dibenzo(a,h)anthracene	0.902	0.917	-1.7	98	0.00
92	Benzo(q,h,i)perylene	0.870	0.910	-4.6	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0