

Data File : BF101846.D
 Acq On : 2 Jan 2018 13:02
 Operator : SJ/JU
 Sample : I7100-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EPE-105-1(0-5)

Manual Integrations
 APPROVED

Sohil
 1/3/2018 10:26:15 AM

Quant Time: Jan 02 13:43:22 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 29 13:21:45 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	144025	20.00	ng	0.00
21) Naphthalene-d8	8.06	136	561488	20.00	ng	0.00
38) Acenaphthene-d10	9.82	164	224411	20.00	ng	0.00
63) Phenanthrene-d10	11.31	188	344372	20.00	ng	0.00
75) Chrysene-d12	13.96	240	269289	20.00	ng	-0.01
86) Perylene-d12	15.43	264	278163	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	977364	101.08	ng	0.00
7) Phenol-d6	6.43	99	1226831	101.95	ng	0.00
23) Nitrobenzene-d5	7.35	82	765217	75.86	ng	-0.01
41) 2,4,6-Tribromophenol	10.62	330	173824	80.39	ng	0.00
44) 2-Fluorobiphenyl	9.13	172	1194279	82.55	ng	0.00
78) Terphenyl-d14	12.89	244	815747	73.03	ng	0.00
Target Compounds						
10) Phenol	6.45	94	52890	3.86	ng	Qvalue # 73
49) Dimethylphthalate	9.53	163	210879	11.68	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF101846.D

Acq On : 2 Jan 2018 13:02

Operator : SJ/JU

Sample : I7100-01

Misc :

ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

EPE-105-1(0-5)

Manual Integrations
APPROVED

Sohil

1/3/2018 10:26:15 AM

Quant Time: Jan 02 13:43:22 2018

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Dec 29 13:21:45 2017

Response via : Initial Calibration

