

Data File : BF101849.D
Acq On : 2 Jan 2018 14:23
Operator : SJ/JU
Sample : I7100-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EPE-106-2(5-10)

Manual Integrations
APPROVED

Sohil
1/3/2018 10:26:17 AM

Quant Time: Jan 03 01:24:06 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 29 13:21:45 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	135380	20.00	ng	0.00
21) Naphthalene-d8	8.06	136	531140	20.00	ng	0.00
38) Acenaphthene-d10	9.82	164	214288	20.00	ng	0.00
63) Phenanthrene-d10	11.31	188	337912	20.00	ng	0.00
75) Chrysene-d12	13.96	240	259999	20.00	ng	-0.01
86) Perylene-d12	15.43	264	271024	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	973371	107.10	ng	0.00
7) Phenol-d6	6.43	99	1178020	104.15	ng	0.00
23) Nitrobenzene-d5	7.35	82	747636	78.36	ng	-0.01
41) 2,4,6-Tribromophenol	10.62	330	194054	93.98	ng	0.00
44) 2-Fluorobiphenyl	9.13	172	979770	70.93	ng	0.00
78) Terphenyl-d14	12.89	244	715711	66.36	ng	-0.01

Target Compounds						Qvalue
49) Dimethylphthalate	9.53	163	186422	10.82	ng	99
70) Phenanthrene	11.33	178	82411	4.39	ng	97
74) Fluoranthene	12.53	202	122086	6.19	ng	99
77) Pyrene	12.76	202	108243	5.55	ng	96
80) Benzo(a)anthracene	13.95	228	59766m	3.48	ng	
82) Chrysene	13.99	228	53167	3.28	ng	98
87) Benzo(b)fluoranthene	15.01	252	59327m	3.59	ng	
89) Benzo(a)pyrene	15.38	252	47156	3.10	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF101849.D

Acq On : 2 Jan 2018 14:23

Operator : SJ/JU

Sample : I7100-04

Misc :

ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

EPE-106-2(5-10)

Manual Integrations
APPROVED

Sohil

1/3/2018 10:26:17 AM

Quant Time: Jan 03 01:24:06 2018

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Dec 29 13:21:45 2017

Response via : Initial Calibration

