

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010220\
 Data File : BF118431.D
 Acq On : 2 Jan 2020 12:58
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 02 14:27:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 15:58:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2	1,4-Dioxane	0.466	0.382	18.0	51	0.01
3	Pyridine	1.241	0.965	22.2	48#	0.02
4	n-Nitrosodimethylamine	0.513	0.506	1.4	59	0.02
5 S	2-Fluorophenol	1.172	0.955	18.5	50#	0.00
6	Aniline	1.846	1.764	4.4	58	0.00
7 S	Phenol-d6	1.453	1.415	2.6	61	0.00
8	2-Chlorophenol	1.327	1.318	0.7	61	0.00
9	Benzaldehyde	0.852	0.734	13.8	55	0.00
10 C	Phenol	1.634	1.595	2.4	60	0.00
11	bis(2-Chloroethyl)ether	1.168	1.128	3.4	59	0.00
12	1,3-Dichlorobenzene	1.528	1.524	0.3	62	0.00
13 C	1,4-Dichlorobenzene	1.552	1.570	-1.2	62	0.00
14	1,2-Dichlorobenzene	1.466	1.473	-0.5	62	0.00
15	Benzyl Alcohol	1.107	1.164	-5.1	64	0.00
16	2,2'-oxybis(1-Chloropropane	1.247	1.240	0.6	61	0.00
17	2-Methylphenol	1.064	1.053	1.0	61	0.00
18	Hexachloroethane	0.570	0.586	-2.8	63	0.00
19 P	n-Nitroso-di-n-propylamine	0.895	0.915	-2.2	64	0.00
20	3+4-Methylphenols	1.362	1.398	-2.6	62	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
22	Acetophenone	0.490	0.479	2.2	63	0.00
23 S	Nitrobenzene-d5	0.350	0.340	2.9	63	0.00
24	Nitrobenzene	0.350	0.334	4.6	62	0.00
25	Isophorone	0.609	0.579	4.9	62	0.00
26 C	2-Nitrophenol	0.186	0.182	2.2	63	0.00
27	2,4-Dimethylphenol	0.253	0.240	5.1	62	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.384	3.8	63	0.00
29 C	2,4-Dichlorophenol	0.291	0.282	3.1	63	0.00
30	1,2,4-Trichlorobenzene	0.339	0.333	1.8	64	0.00
31	Naphthalene	1.017	1.043	-2.6	66	0.00
32	Benzoic acid	0.199	0.168	15.6	54	0.02
33	4-Chloroaniline	0.436	0.422	3.2	62	0.00
34 C	Hexachlorobutadiene	0.202	0.196	3.0	63	0.00
35	Caprolactam	0.091	0.082	9.9	57	0.00
36 C	4-Chloro-3-methylphenol	0.313	0.298	4.8	62	0.00
37	2-Methylnaphthalene	0.696	0.658	5.5	61	0.00
38	1-Methylnaphthalene	0.656	0.620	5.5	62	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	56	0.00
40	1,2,4,5-Tetrachlorobenzene	0.601	0.620	-3.2	61	0.00
41 P	Hexachlorocyclopentadiene	0.206	0.186	9.7	55	0.00
42 S	2,4,6-Tribromophenol	0.247	0.250	-1.2	59	0.00
43 C	2,4,6-Trichlorophenol	0.401	0.411	-2.5	60	0.00
44	2,4,5-Trichlorophenol	0.413	0.413	0.0	58	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010220\
 Data File : BF118431.D
 Acq On : 2 Jan 2020 12:58
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 02 14:27:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 24 15:58:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.310	1.448	-10.5	64	0.00
46	1,1'-Biphenyl	1.582	1.763	-11.4	65	0.00
47	2-Chloronaphthalene	1.273	1.416	-11.2	64	0.00
48	2-Nitroaniline	0.360	0.402	-11.7	65	0.00
49	Acenaphthylene	1.888	1.996	-5.7	61	0.00
50	Dimethylphthalate	1.501	1.583	-5.5	62	0.00
51	2,6-Dinitrotoluene	0.324	0.340	-4.9	62	0.00
52 C	Acenaphthene	1.150	1.159	-0.8	59	0.00
53	3-Nitroaniline	0.380	0.374	1.6	57	0.00
54 P	2,4-Dinitrophenol	0.146	0.171	-17.1	69	0.01
55	Dibenzofuran	1.699	1.749	-2.9	59	0.00
56 P	4-Nitrophenol	0.235	0.288	-22.6	71	0.02
57	2,4-Dinitrotoluene	0.432	0.439	-1.6	59	0.00
58	Fluorene	1.372	1.396	-1.7	59	0.00
59	2,3,4,6-Tetrachlorophenol	0.351	0.353	-0.6	58	0.00
60	Diethylphthalate	1.485	1.517	-2.2	59	0.00
61	4-Chlorophenyl-phenylether	0.685	0.687	-0.3	58	0.00
62	4-Nitroaniline	0.395	0.384	2.8	56	0.00
63	Azobenzene	1.272	1.348	-6.0	62	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	54	0.00
65	4,6-Dinitro-2-methylphenol	0.108	0.108	0.0	56	0.00
66 c	n-Nitrosodiphenylamine	0.641	0.668	-4.2	59	0.00
67	4-Bromophenyl-phenylether	0.234	0.229	2.1	56	0.00
68	Hexachlorobenzene	0.276	0.266	3.6	56	0.00
69	Atrazine	0.198	0.191	3.5	53	0.00
70 C	Pentachlorophenol	0.119	0.123	-3.4	57	0.00
71	Phenanthrene	1.090	1.149	-5.4	60	0.00
72	Anthracene	1.091	1.175	-7.7	61	0.00
73	Carbazole	0.968	0.994	-2.7	59	0.00
74	Di-n-butylphthalate	1.228	1.375	-12.0	63	0.00
75 C	Fluoranthene	1.100	1.196	-8.7	61	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
77	Benzidine	0.548	0.487	11.1	52	0.00
78	Pyrene	1.633	1.541	5.6	63	0.00
79 S	Terphenyl-d14	1.206	1.183	1.9	64	0.00
80	Butylbenzylphthalate	0.735	0.699	4.9	62	0.00
81	Benzo(a)anthracene	1.400	1.388	0.9	64	0.00
82	3,3'-Dichlorobenzidine	0.475	0.431	9.3	56	0.00
83	Chrysene	1.340	1.327	1.0	64	0.00
84	Bis(2-ethylhexyl)phthalate	0.967	0.938	3.0	63	0.00
85 c	Di-n-octyl phthalate	1.422	1.399	1.6	64	0.00
86	Indeno(1,2,3-cd)pyrene	1.142	0.957	16.2	58	0.02
87 I	Perylene-d12	1.000	1.000	0.0	61	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010220\
Data File : BF118431.D
Acq On : 2 Jan 2020 12:58
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 02 14:27:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF122419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 24 15:58:52 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.328	1.315	1.0	65	0.00
89	Benzo(k)fluoranthene	1.275	1.217	4.5	59	0.00
90 C	Benzo(a)pyrene	1.170	1.137	2.8	63	0.00
91	Dibenzo(a,h)anthracene	1.015	0.919	9.5	59	0.00
92	Benzo(q,h,i)perylene	0.980	0.827	15.6	55	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0