


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010224\
Data File : BF136863.D
Acq On : 02 Jan 2024 10:50
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Dec 21 01:54:25 2023






AutoFind: Scans 1611, 1612, 1613; Background Corrected with Scan 1605

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.4	47249	PASS
68	69	0.00	2	1.5	663	PASS
69	198	0.00	100	35.4	43558	PASS
70	69	0.00	2	0.4	192	PASS
127	198	10	80	46.0	56636	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	123037	PASS
199	198	5	9	6.9	8475	PASS
275	198	10	60	30.7	37775	PASS
365	198	1	100	3.8	4691	PASS
441	198	0.01	100	18.1	22285	PASS
442	442	50	100	100.0	138592	PASS
443	442	15	24	18.3	25371	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010224\
Data File : BF136863.D
Acq On : 02 Jan 2024 10:50
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 02 12:21:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration



TIC: BF136863.D\data.ms

(70) Pentachlorophenol (C)

11.116min (-0.006) 40234.39 ng




response 119533

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.70	64.70
264.00	61.20	61.62
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010224\
Data File : BF136863.D
Acq On : 02 Jan 2024 10:50
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 02 12:21:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration



TIC: BF136863.D\data.ms

(77) Benzidine

12.657min (-0.006) 0.00 ng m

response 677873

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	13.86
183.00	11.60	11.34
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/2/2024	BNA_F	<u>BF136863.D</u>
Compound Name	Response	Retention Time
DDT	315011	13.48
DDD	13129	13.174
DDE	392	12.833
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13521	328532	4.12

Instrument :
BNA_F
ClientSampleId :
DFTPP