

Data Path : Z:\HPCHEM1\BNA F\DATA\BF010318\
 Data File : BF101868.D
 Acq On : 3 Jan 2018 18:58
 Operator : SJ/JU
 Sample : I7105-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT1500

Manual Integrations
 APPROVED

Sohil
 1/4/2018 1:56:34 PM

Quant Time: Jan 04 02:58:45 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 29 13:21:45 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	124201	20.00	ng	-0.01
21) Naphthalene-d8	8.05	136	440512	20.00	ng	-0.02
38) Acenaphthene-d10	9.80	164	169661	20.00	ng	-0.02
63) Phenanthrene-d10	11.30	188	319048	20.00	ng	-0.02
75) Chrysene-d12	13.95	240	279196	20.00	ng	-0.02
86) Perylene-d12	15.42	264	204865	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	800282	95.98	ng	0.00
7) Phenol-d6	6.42	99	924924	89.13	ng	-0.01
23) Nitrobenzene-d5	7.35	82	541372	68.41	ng	-0.02
41) 2,4,6-Tribromophenol	10.60	330	157455	96.31	ng	-0.02
44) 2-Fluorobiphenyl	9.12	172	836487	76.48	ng	-0.02
78) Terphenyl-d14	12.88	244	771845	66.65	ng	-0.02
Target Compounds						
10) Phenol	6.44	94	38814	3.282	ng	Qvalue # 67
49) Dimethylphthalate	9.52	163	70451	5.162	ng	99
70) Phenanthrene	11.33	178	44672	2.518	ng	97
74) Fluoranthene	12.52	202	71699	3.850	ng	99
77) Pyrene	12.75	202	65191	3.111	ng	99
87) Benzo(b)fluoranthene	15.00	252	25304m	2.026	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF010318\
Data File : BF101868.D
Acq On : 3 Jan 2018 18:58
Operator : SJ/JU
Sample : I7105-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RT1500

Manual Integrations
APPROVED

Sohil
1/4/2018 1:56:34 PM

Quant Time: Jan 04 02:58:45 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 29 13:21:45 2017
Response via : Initial Calibration

