

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010324\
 Data File : BF136899.D
 Acq On : 03 Jan 2024 10:15
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 03 11:11:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	-0.01
2	1,4-Dioxane	40.000	39.775	0.6	106	0.00
3	Pyridine	40.000	40.291	-0.7	104	0.00
4	n-Nitrosodimethylamine	40.000	37.857	5.4	96	0.00
5 S	2-Fluorophenol	80.000	82.429	-3.0	107	0.00
6	Aniline	40.000	38.142	4.6	98	-0.01
7 S	Phenol-d6	80.000	81.608	-2.0	106	0.00
8	2-Chlorophenol	40.000	41.065	-2.7	106	0.00
9	Benzaldehyde	40.000	37.036	7.4	98	-0.01
10 C	Phenol	40.000	40.510	-1.3	106	0.00
11	bis(2-Chloroethyl)ether	40.000	40.469	-1.2	104	0.00
12	1,3-Dichlorobenzene	40.000	41.079	-2.7	106	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.748	-1.9	106	-0.01
14	1,2-Dichlorobenzene	40.000	40.914	-2.3	106	-0.01
15	Benzyl Alcohol	40.000	41.303	-3.3	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.001	2.5	100	-0.01
17	2-Methylphenol	40.000	39.290	1.8	103	0.00
18	Hexachloroethane	40.000	39.942	0.1	104	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	39.312	1.7	103	0.00
20	3+4-Methylphenols	40.000	40.427	-1.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	40.442	-1.1	105	-0.01
23 S	Nitrobenzene-d5	80.000	80.615	-0.8	103	0.00
24	Nitrobenzene	40.000	40.147	-0.4	101	0.00
25	Isophorone	40.000	39.638	0.9	102	-0.01
26 C	2-Nitrophenol	40.000	42.551	-6.4	106	-0.01
27	2,4-Dimethylphenol	40.000	41.106	-2.8	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.990	0.0	101	0.00
29 C	2,4-Dichlorophenol	40.000	41.534	-3.8	106	0.00
30	1,2,4-Trichlorobenzene	40.000	40.989	-2.5	105	-0.01
31	Naphthalene	40.000	40.988	-2.5	105	-0.01
32	Benzoic acid	40.000	39.671	0.8	94	0.00
33	4-Chloroaniline	40.000	38.759	3.1	99	0.00
34 C	Hexachlorobutadiene	40.000	40.677	-1.7	105	-0.01
35	Caprolactam	40.000	46.931	-17.3	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.148	-0.4	103	0.00
37	2-Methylnaphthalene	40.000	40.412	-1.0	105	-0.01
38	1-Methylnaphthalene	40.000	40.592	-1.5	106	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	41.548	-3.9	104	-0.01
41 P	Hexachlorocyclopentadiene	40.000	51.945	-29.9#	139	-0.01
42 S	2,4,6-Tribromophenol	80.000	79.926	0.1	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.111	-0.3	100	0.00
44	2,4,5-Trichlorophenol	40.000	40.359	-0.9	101	0.00
45 S	2-Fluorobiphenyl	80.000	80.347	-0.4	104	-0.01
46	1,1'-Biphenyl	40.000	40.720	-1.8	104	-0.01
47	2-Chloronaphthalene	40.000	40.944	-2.4	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010324\
 Data File : BF136899.D
 Acq On : 03 Jan 2024 10:15
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 03 11:11:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.517	-1.3	101	0.00
49	Acenaphthylene	40.000	40.438	-1.1	102	-0.01
50	Dimethylphthalate	40.000	40.376	-0.9	103	-0.01
51	2,6-Dinitrotoluene	40.000	42.054	-5.1	105	0.00
52 C	Acenaphthene	40.000	40.578	-1.4	103	-0.01
53	3-Nitroaniline	40.000	39.376	1.6	97	0.00
54 P	2,4-Dinitrophenol	40.000	43.960	-9.9	108	0.00
55	Dibenzofuran	40.000	39.791	0.5	102	0.00
56 P	4-Nitrophenol	40.000	45.362	-13.4	109	0.00
57	2,4-Dinitrotoluene	40.000	40.730	-1.8	101	0.00
58	Fluorene	40.000	40.507	-1.3	104	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	37.773	5.6	93	0.00
60	Diethylphthalate	40.000	40.059	-0.1	101	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.353	1.6	100	0.00
62	4-Nitroaniline	40.000	38.449	3.9	100	0.00
63	Azobenzene	40.000	39.525	1.2	101	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.148	-2.9	95	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.416	-1.0	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.722	-1.8	104	-0.01
68	Hexachlorobenzene	40.000	41.462	-3.7	106	-0.01
69	Atrazine	40.000	43.356	-8.4	131	0.00
70 C	Pentachlorophenol	40.000	32.648	18.4	76	0.00
71	Phenanthrene	40.000	41.220	-3.0	105	-0.01
72	Anthracene	40.000	40.723	-1.8	105	-0.01
73	Carbazole	40.000	40.671	-1.7	104	0.00
74	Di-n-butylphthalate	40.000	41.659	-4.1	107	-0.01
75 C	Fluoranthene	40.000	40.965	-2.4	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzidine	40.000	39.672	0.8	110	0.00
78	Pyrene	40.000	41.768	-4.4	106	0.00
79 S	Terphenyl-d14	80.000	84.014	-5.0	107	-0.01
80	Butylbenzylphthalate	40.000	44.080	-10.2	110	-0.01
81	Benzo(a)anthracene	40.000	41.154	-2.9	106	0.00
82	3,3'-Dichlorobenzidine	40.000	41.391	-3.5	109	0.00
83	Chrysene	40.000	40.691	-1.7	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.894	-14.7	115	0.00
85 c	Di-n-octyl phthalate	40.000	41.129	-2.8	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.232	-3.1	100	0.00
88	Benzo(b)fluoranthene	40.000	41.464	-3.7	107	0.00
89	Benzo(k)fluoranthene	40.000	39.046	2.4	96	0.00
90 C	Benzo(a)pyrene	40.000	40.569	-1.4	102	0.00
91	Dibenzo(a,h)anthracene	40.000	41.222	-3.1	99	0.00
92	Benzo(g,h,i)perylene	40.000	41.738	-4.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010324\
Data File : BF136899.D
Acq On : 03 Jan 2024 10:15
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 03 11:11:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0