




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010417\
Data File : BF092072.D
Acq On : 4 Jan 2017 16:59
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 05 04:53:42 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 05 00:16:49 2017
Response via : Initial Calibration



TIC: BF092072.D

(69) Pentachlorophenol (C)




10.985min (-0.012) 0.00ng

response 225750

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	64.07
264.00	62.90	64.77
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010417\
Data File : BF092072.D
Acq On : 4 Jan 2017 16:59
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 05 04:53:42 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 05 00:16:49 2017
Response via : Initial Calibration



TIC: BF092072.D

(76) Benzidine

12.516min (-0.000) -20.00ng

response 1421030

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.54
183.00	11.50	11.53
0.00	0.00	0.00