

Data Path : U:\HPCHEM1\BNA F\DATA\BF010518\
 Data File : BF101934.D
 Acq On : 5 Jan 2018 17:15
 Operator : SJ/JU
 Sample : J1012-12
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-010318-F

Quant Time: Jan 05 18:12:16 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 29 13:21:45 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	119699	20.00	ng	0.00
21) Naphthalene-d8	8.06	136	453490	20.00	ng	-0.01
38) Acenaphthene-d10	9.81	164	165629	20.00	ng	-0.01
63) Phenanthrene-d10	11.30	188	262191	20.00	ng	-0.01
75) Chrysene-d12	13.96	240	246558	20.00	ng	-0.02
86) Perylene-d12	15.43	264	198527	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	1019346	126.85	ng	0.00
7) Phenol-d6	6.43	99	1095750	109.57	ng	0.00
23) Nitrobenzene-d5	7.35	82	820642	100.73	ng	-0.02
41) 2,4,6-Tribromophenol	10.61	330	175192	109.77	ng	-0.01
44) 2-Fluorobiphenyl	9.13	172	1109749	103.94	ng	-0.01
78) Terphenyl-d14	12.88	244	859264	84.02	ng	-0.02


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF010518\
Data File : BF101934.D
Acq On : 5 Jan 2018 17:15
Operator : SJ/JU
Sample : J1012-12
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-010318-F

Quant Time: Jan 05 18:12:16 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 29 13:21:45 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.77 min Scan# 797
Delta R.T. -0.01 min
Lab File: BF101934.D
Acq: 5 Jan 2018 17:15

Instrument :
BNA_F
ClientSampleId :
JC-03-010318-F

Tot Ion:152 Resp: 119699
Ion Ratio Lower Upper
152 100
150 154.1 124.6 186.8
115 61.1 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 126.851 ng
RT: 5.41 min Scan# 565
Delta R.T. 0.01 min
Lab File: BF101934.D
Acq: 5 Jan 2018 17:15

Tgt Ion:112 Resp: 1019346
Ion Ratio Lower Upper
112 100
64 60.9 47.9 71.9
63 32.5 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 109.566 ng

RT: 6.43 min Scan# 738

Delta R.T. -0.01 min

Lab File: BF101934.D

Acq: 5 Jan 2018 17:15

Instrument :


BNA_F

ClientSampleId :




JC-03-010318-F

Tot Ion: 99 Resp: 1095750

Ion	Ratio	Lower	Upper
99	100		
42	21.7	14.5	21.7#
71	33.2	25.4	38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.06 min Scan# 1015


Delta R.T. -0.01 min

Lab File: BF101934.D



Acq: 5 Jan 2018 17:15


Tgt Ion: 136 Resp: 453490

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	8.5	6.6	9.8
68	6.3	5.0	7.4



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 100.734 ng
 RT: 7.35 min Scan# 894
 Delta R.T. -0.02 min
 Lab File: BF101934.D
 Acq: 5 Jan 2018 17:15

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-010318-F

Tot Ion: 82 Resp: 820642
 Ion Ratio Lower Upper
 82 100
 128 41.1 36.1 54.1
 54 57.1 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.81 min Scan# 1313
 Delta R.T. -0.01 min
 Lab File: BF101934.D
 Acq: 5 Jan 2018 17:15

Tgt Ion:164 Resp: 165629
 Ion Ratio Lower Upper
 164 100
 162 104.0 86.1 129.1
 160 48.6 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 109.772 ng
 RT: 10.61 min Scan# 1449
 Delta R.T. -0.01 min
 Lab File: BF101934.D
 Acq: 5 Jan 2018 17:15

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-010318-F

Tot Ion	Ratio	Lower	Upper
330	100		
332	98.0	77.0	115.6
141	40.0	29.9	44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 103.936 ng
 RT: 9.13 min Scan# 1197
 Delta R.T. -0.01 min
 Lab File: BF101934.D
 Acq: 5 Jan 2018 17:15

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.7	44.5
170	24.7	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.30 min Scan# 1567

Delta R.T. -0.01 min

Lab File: BF101934.D

Acq: 5 Jan 2018 17:15

Instrument :

BNA_F

ClientSampleId :

JC-03-010318-F


Tot Ion:188 Resp: 262191

Ion Ratio Lower Upper

188 100

94 9.8 8.5 12.7




80 11.3 9.2 13.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#75

Chrysene-d12

Concen: 20.000 ng

RT: 13.96 min Scan# 2018

Delta R.T. -0.02 min

Lab File: BF101934.D

Acq: 5 Jan 2018 17:15


Tgt Ion:240 Resp: 246558

Ion Ratio Lower Upper

240 100

120 13.0 9.5 14.3



236 25.7 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 84.016 ng

RT: 12.88 min Scan# 1835

Delta R.T. -0.02 min

Lab File: BF101934.D

Acq: 5 Jan 2018 17:15

Instrument :

BNA_F

ClientSampleId :

JC-03-010318-F


Tot Ion:244 Resp: 859264

Ion Ratio Lower Upper

244 100

212 7.3 5.4 8.0




122 12.9 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.43 min Scan# 2269

Delta R.T. -0.01 min

Lab File: BF101934.D

Acq: 5 Jan 2018 17:15


Tgt Ion:264 Resp: 198527

Ion Ratio Lower Upper

264 100

260 23.5 19.2 28.8

265 21.2 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

